

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG NO.: Q2073

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 05/28/2025

Lab File ID: BN037112.D Time Analyzed: 17:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2022	7.611	5050	10.38	2480	14.26
UPPER LIMIT	4044	8.111	10100	10.883	4960	14.756
LOWER LIMIT	1011	7.111	2525	9.883	1240	13.756
EPA SAMPLE NO.						
01 PB168100BL	1765	7.61	4396	10.39	2346	14.26
02 PB168100BS	2482	7.61	6249	10.38	2970	14.26
03 PB168100BSD	2246	7.61	5562	10.38	2593	14.26

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.