

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: GENV01

Lab Code: CHEM

SAS No.: Q2073

SDG NO.: Q2073

Lab File ID: BP024752.D

DFTPP Injection Date: 05/22/2025

Instrument ID: BNA_P

DFTPP Injection Time: 09:35

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	30.1
68	Less than 2.0% of mass 69	0.3 (1.3) 1
69	Mass 69 relative abundance	32.0
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	46.0
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.0
275	10.0 - 60.0% of mass 198	33.7
365	Greater than 1% of mass 198	5.4
441	Present, but less than mass 443	22.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.6 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024753.D	05/22/2025	10:56
PB168098BL	PB168098BL	BP024754.D	05/22/2025	11:37
PB168098BS	PB168098BS	BP024755.D	05/22/2025	12:18
PB168098BSD	PB168098BSD	BP024756.D	05/22/2025	12:58