

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG NO.: Q2073

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/23/2025

Lab File ID: BP024788.D Time Analyzed: 11:40

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	85825	7.646	348651	10.42	223501	14.28
UPPER LIMIT	171650	8.146	697302	10.916	447002	14.781
LOWER LIMIT	42912.5	7.146	174326	9.916	111751	13.781
EPA SAMPLE NO.						
01 GDW1	100868	7.65	399490	10.42	287857	14.28
02 GDW2	87294	7.65	334159	10.42	218481	14.29

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.