



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046254.D
Acq On : 19 May 2025 09:33
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.1	8149	PASS
75	95	30	60	58.5	21593	PASS
95	95	100	100	100.0	36885	PASS
96	95	5	9	6.6	2433	PASS
173	174	0.00	2	1.1	273	PASS
174	95	50	100	67.9	25035	PASS
175	174	5	9	8.1	2023	PASS
176	174	95	101	99.1	24816	PASS
177	176	5	9	6.9	1720	PASS