

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following "Results Qualifiers" are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. "10 U". This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as "12 B".
E	Indicates the analyte 's concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a "P".
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

LAB CHRONICLE

OrderID:	Q2073	OrderDate:	5/16/2025 2:51:00 PM
Client:	G Environmental	Project:	Nelson
Contact:	Gary Landis	Location:	L41,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2073-01	GDW1	Water	VOCMS Group1	8260-Low	05/16/25		05/19/25	05/16/25
Q2073-02	GDW2	Water	VOCMS Group1	8260-Low	05/16/25		05/19/25	05/16/25

Hit Summary Sheet
SW-846

SDG No.: Q2073
Client: G Environmental

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	GDW1							
Q2073-01	GDW1	Water	Acetone	6.80		1.50	5.00	ug/L
Q2073-01	GDW1	Water	Methyl Acetate	3.50		0.27	1.00	ug/L
			Total Voc :	10.3				
			Total Concentration:	10.3				
Client ID:	GDW2							
Q2073-02	GDW2	Water	Chloromethane	1.10		0.32	1.00	ug/L
Q2073-02	GDW2	Water	Acetone	5.20		1.50	5.00	ug/L
Q2073-02	GDW2	Water	Methyl Acetate	1.40		0.27	1.00	ug/L
			Total Voc :	7.70				
			Total Concentration:	7.70				



QC SUMMARY

Surrogate Summary

SDG No.: Q2073

Client: G Environmental

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2073-01	GDW1	1,2-Dichloroethane-d4	50	53.5	107	70 (74)	130 (125)
		Dibromofluoromethane	50	50.8	102	70 (75)	130 (124)
		Toluene-d8	50	51.2	102	70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.5	101	70 (77)	130 (121)
Q2073-02	GDW2	1,2-Dichloroethane-d4	50	54.3	109	70 (74)	130 (125)
		Dibromofluoromethane	50	50.1	100	70 (75)	130 (124)
		Toluene-d8	50	49.8	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.8	97	70 (77)	130 (121)
VX0519WBL01	VX0519WBL01	1,2-Dichloroethane-d4	50	54.1	108	70 (74)	130 (125)
		Dibromofluoromethane	50	51.4	103	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.6	99	70 (77)	130 (121)
VX0519WBS01	VX0519WBS01	1,2-Dichloroethane-d4	50	51.6	103	70 (74)	130 (125)
		Dibromofluoromethane	50	53.1	106	70 (75)	130 (124)
		Toluene-d8	50	51.6	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.4	103	70 (77)	130 (121)
VX0519WBSD0	VX0519WBSD01	1,2-Dichloroethane-d4	50	52.0	104	70 (74)	130 (125)
		Dibromofluoromethane	50	52.9	106	70 (75)	130 (124)
		Toluene-d8	50	51.4	103	70 (86)	130 (113)
		4-Bromofluorobenzene	50	51.5	103	70 (77)	130 (121)

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2073

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VX046258.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0519WBS01	Dichlorodifluoromethane	20	18.7	ug/L	94			40 (69)	160 (116)	
	Chloromethane	20	17.7	ug/L	89			40 (65)	160 (116)	
	Vinyl chloride	20	17.2	ug/L	86			70 (65)	130 (117)	
	Bromomethane	20	18.6	ug/L	93			40 (58)	160 (125)	
	Chloroethane	20	19.3	ug/L	97			40 (56)	160 (128)	
	Trichlorofluoromethane	20	18.8	ug/L	94			40 (73)	160 (115)	
	1,1,2-Trichlorotrifluoroethane	20	19.0	ug/L	95			70 (80)	130 (112)	
	1,1-Dichloroethene	20	18.4	ug/L	92			70 (74)	130 (110)	
	Acetone	100	95.5	ug/L	96			40 (60)	160 (125)	
	Carbon disulfide	20	15.7	ug/L	79			40 (64)	160 (112)	
	Methyl tert-butyl Ether	20	19.0	ug/L	95			70 (78)	130 (114)	
	Methyl Acetate	20	22.6	ug/L	113			70 (67)	130 (125)	
	Methylene Chloride	20	18.1	ug/L	91			70 (72)	130 (114)	
	trans-1,2-Dichloroethene	20	18.2	ug/L	91			70 (75)	130 (108)	
	1,1-Dichloroethane	20	19.3	ug/L	97			70 (78)	130 (112)	
	Cyclohexane	20	17.8	ug/L	89			70 (75)	130 (110)	
	2-Butanone	100	98.6	ug/L	99			40 (65)	160 (122)	
	Carbon Tetrachloride	20	19.1	ug/L	96			70 (77)	130 (113)	
	cis-1,2-Dichloroethene	20	18.4	ug/L	92			70 (77)	130 (110)	
	Bromochloromethane	20	21.4	ug/L	107			70 (70)	130 (124)	
	Chloroform	20	19.7	ug/L	99			70 (79)	130 (113)	
	1,1,1-Trichloroethane	20	19.1	ug/L	96			70 (80)	130 (108)	
	Methylcyclohexane	20	17.6	ug/L	88			70 (72)	130 (115)	
	Benzene	20	19.0	ug/L	95			70 (82)	130 (109)	
	1,2-Dichloroethane	20	19.8	ug/L	99			70 (80)	130 (115)	
	Trichloroethene	20	18.2	ug/L	91			70 (77)	130 (113)	
	1,2-Dichloropropane	20	19.9	ug/L	100			70 (83)	130 (111)	
	Bromodichloromethane	20	20.0	ug/L	100			70 (83)	130 (110)	
	4-Methyl-2-Pentanone	100	99.9	ug/L	100			40 (74)	160 (118)	
	Toluene	20	19.5	ug/L	98			70 (82)	130 (110)	
	t-1,3-Dichloropropene	20	18.5	ug/L	93			70 (79)	130 (110)	
	cis-1,3-Dichloropropene	20	19.0	ug/L	95			70 (82)	130 (110)	
	1,1,2-Trichloroethane	20	20.4	ug/L	102			70 (83)	130 (112)	
	2-Hexanone	100	100	ug/L	100			40 (73)	160 (117)	
	Dibromochloromethane	20	19.9	ug/L	100			70 (82)	130 (110)	
	1,2-Dibromoethane	20	19.6	ug/L	98			70 (81)	130 (110)	
	Tetrachloroethene	20	17.9	ug/L	90			70 (67)	130 (123)	
	Chlorobenzene	20	18.5	ug/L	93			70 (82)	130 (109)	
	Ethyl Benzene	20	18.9	ug/L	95			70 (83)	130 (109)	
	m/p-Xylenes	40	37.1	ug/L	93			70 (82)	130 (110)	
	o-Xylene	20	19.0	ug/L	95			70 (83)	130 (109)	
	Styrene	20	19.7	ug/L	99			70 (80)	130 (111)	
	Bromoform	20	19.0	ug/L	95			70 (79)	130 (109)	
	Isopropylbenzene	20	19.5	ug/L	98			70 (83)	130 (112)	
	1,1,2,2-Tetrachloroethane	20	19.3	ug/L	97			70 (76)	130 (118)	
	1,3-Dichlorobenzene	20	18.8	ug/L	94			70 (82)	130 (108)	
	1,4-Dichlorobenzene	20	18.2	ug/L	91			70 (82)	130 (107)	
	1,2-Dichlorobenzene	20	19.3	ug/L	97			70 (82)	130 (109)	

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2073

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VX046258.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0519WBS01	1,2-Dibromo-3-Chloropropane	20	19.0	ug/L	95			40 (68)	160 (112)	
	1,2,4-Trichlorobenzene	20	17.8	ug/L	89			70 (75)	130 (113)	
	1,2,3-Trichlorobenzene	20	18.1	ug/L	91			70 (76)	130 (114)	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2073

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VX046259.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0519WBSD01	Dichlorodifluoromethane	20	18.0	ug/L	90	4		40 (69)	160 (116)	20 (20)
	Chloromethane	20	17.3	ug/L	86	3		40 (65)	160 (116)	20 (20)
	Vinyl chloride	20	16.7	ug/L	84	2		70 (65)	130 (117)	20 (20)
	Bromomethane	20	17.2	ug/L	86	8		40 (58)	160 (125)	20 (20)
	Chloroethane	20	17.8	ug/L	89	9		40 (56)	160 (128)	20 (20)
	Trichlorofluoromethane	20	18.4	ug/L	92	2		40 (73)	160 (115)	20 (20)
	1,1,2-Trichlorotrifluoroethane	20	18.5	ug/L	93	2		70 (80)	130 (112)	20 (20)
	1,1-Dichloroethene	20	17.9	ug/L	90	2		70 (74)	130 (110)	20 (20)
	Acetone	100	98.7	ug/L	99	3		40 (60)	160 (125)	20 (20)
	Carbon disulfide	20	15.1	ug/L	76	4		40 (64)	160 (112)	20 (20)
	Methyl tert-butyl Ether	20	19.5	ug/L	98	3		70 (78)	130 (114)	20 (20)
	Methyl Acetate	20	23.4	ug/L	117	3		70 (67)	130 (125)	20 (20)
	Methylene Chloride	20	17.8	ug/L	89	2		70 (72)	130 (114)	20 (20)
	trans-1,2-Dichloroethene	20	18.0	ug/L	90	1		70 (75)	130 (108)	20 (20)
	1,1-Dichloroethane	20	19.0	ug/L	95	2		70 (78)	130 (112)	20 (20)
	Cyclohexane	20	17.6	ug/L	88	1		70 (75)	130 (110)	20 (20)
	2-Butanone	100	100	ug/L	100	1		40 (65)	160 (122)	20 (20)
	Carbon Tetrachloride	20	18.8	ug/L	94	2		70 (77)	130 (113)	20 (20)
	cis-1,2-Dichloroethene	20	19.0	ug/L	95	3		70 (77)	130 (110)	20 (20)
	Bromochloromethane	20	20.9	ug/L	104	3		70 (70)	130 (124)	20 (20)
	Chloroform	20	19.9	ug/L	100	1		70 (79)	130 (113)	20 (20)
	1,1,1-Trichloroethane	20	19.2	ug/L	96	0		70 (80)	130 (108)	20 (20)
	Methylcyclohexane	20	17.2	ug/L	86	2		70 (72)	130 (115)	20 (20)
	Benzene	20	19.1	ug/L	96	1		70 (82)	130 (109)	20 (20)
	1,2-Dichloroethane	20	20.3	ug/L	102	3		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	18.6	ug/L	93	2		70 (77)	130 (113)	20 (20)
	1,2-Dichloropropane	20	18.7	ug/L	94	6		70 (83)	130 (111)	20 (20)
	Bromodichloromethane	20	19.9	ug/L	100	0		70 (83)	130 (110)	20 (20)
	4-Methyl-2-Pentanone	100	100	ug/L	100	0		40 (74)	160 (118)	20 (20)
	Toluene	20	19.1	ug/L	96	2		70 (82)	130 (110)	20 (20)
	t-1,3-Dichloropropene	20	18.8	ug/L	94	1		70 (79)	130 (110)	20 (20)
	cis-1,3-Dichloropropene	20	19.3	ug/L	97	2		70 (82)	130 (110)	20 (20)
	1,1,2-Trichloroethane	20	20.3	ug/L	102	0		70 (83)	130 (112)	20 (20)
	2-Hexanone	100	110	ug/L	110	10		40 (73)	160 (117)	20 (20)
	Dibromochloromethane	20	19.7	ug/L	99	1		70 (82)	130 (110)	20 (20)
	1,2-Dibromoethane	20	20.2	ug/L	101	3		70 (81)	130 (110)	20 (20)
	Tetrachloroethene	20	18.2	ug/L	91	1		70 (67)	130 (123)	20 (20)
	Chlorobenzene	20	18.7	ug/L	94	1		70 (82)	130 (109)	20 (20)
	Ethyl Benzene	20	19.0	ug/L	95	0		70 (83)	130 (109)	20 (20)
	m/p-Xylenes	40	38.2	ug/L	96	3		70 (82)	130 (110)	20 (20)
	o-Xylene	20	19.1	ug/L	96	1		70 (83)	130 (109)	20 (20)
	Styrene	20	19.8	ug/L	99	0		70 (80)	130 (111)	20 (20)
	Bromoform	20	19.2	ug/L	96	1		70 (79)	130 (109)	20 (20)
	Isopropylbenzene	20	19.2	ug/L	96	2		70 (83)	130 (112)	20 (20)
	1,1,2,2-Tetrachloroethane	20	19.5	ug/L	98	1		70 (76)	130 (118)	20 (20)
	1,3-Dichlorobenzene	20	18.8	ug/L	94	0		70 (82)	130 (108)	20 (20)
	1,4-Dichlorobenzene	20	18.8	ug/L	94	3		70 (82)	130 (107)	20 (20)
	1,2-Dichlorobenzene	20	19.8	ug/L	99	2		70 (82)	130 (109)	20 (20)

() = LABORATORY INHOUSE LIMIT

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2073

Client: G Environmental

Analytical Method: SW8260-Low

Datafile : VX046259.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0519WBSD01	1,2-Dibromo-3-Chloropropane	20	20.0	ug/L	100	5		40 (68)	160 (112)	20 (20)
	1,2,4-Trichlorobenzene	20	18.6	ug/L	93	4		70 (75)	130 (113)	20 (20)
	1,2,3-Trichlorobenzene	20	19.4	ug/L	97	6		70 (76)	130 (114)	20 (20)

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VX0519WBL01

Lab Name: CHEMTECH

Contract: GENV01

Lab Code: CHEM Case No.: Q2073

SAS No.: Q2073 SDG NO.: Q2073

Lab File ID: VX046257.D

Lab Sample ID: VX0519WBL01

Date Analyzed: 05/19/2025

Time Analyzed: 11:02

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_X

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VX0519WBS01	VX0519WBS01	VX046258.D	05/19/2025
VX0519WBSD01	VX0519WBSD01	VX046259.D	05/19/2025
GDW1	Q2073-01	VX046280.D	05/19/2025
GDW2	Q2073-02	VX046281.D	05/19/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG NO.: Q2073
Lab File ID: VX046038.D BFB Injection Date: 05/05/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:37
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	56.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.4
173	Less than 2.0% of mass 174	0.5 (0.7) 1
174	50.0 - 100.0% of mass 95	68.8
175	5.0 - 9.0% of mass 174	5 (7.3) 1
176	95.0 - 101.0% of mass 174	66.7 (97) 1
177	5.0 - 9.0% of mass 176	4.6 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC020	VSTDIC020	VX046041.D	05/05/2025	11:35
VSTDIC050	VSTDIC050	VX046042.D	05/05/2025	11:58
VSTDIC100	VSTDIC100	VX046043.D	05/05/2025	12:21
VSTDIC150	VSTDIC150	VX046044.D	05/05/2025	12:45
VSTDIC005	VSTDIC005	VX046046.D	05/05/2025	16:04
VSTDIC001	VSTDIC001	VX046047.D	05/05/2025	16:27



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG NO.: Q2073
Lab File ID: VX046254.D BFB Injection Date: 05/19/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:33
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.1
75	30.0 - 60.0% of mass 95	58.5
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.7 (1.1) 1
174	50.0 - 100.0% of mass 95	67.9
175	5.0 - 9.0% of mass 174	5.5 (8.1) 1
176	95.0 - 101.0% of mass 174	67.3 (99.1) 1
177	5.0 - 9.0% of mass 176	4.7 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX046255.D	05/19/2025	10:05
VX0519WBL01	VX0519WBL01	VX046257.D	05/19/2025	11:02
VX0519WBS01	VX0519WBS01	VX046258.D	05/19/2025	11:25
VX0519WBSD01	VX0519WBSD01	VX046259.D	05/19/2025	11:53
GDW1	Q2073-01	VX046280.D	05/19/2025	20:06
GDW2	Q2073-02	VX046281.D	05/19/2025	20:29

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG NO.: Q2073
 Lab File ID: VX046255.D Date Analyzed: 05/19/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:05
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	97192	5.54	163845	6.76	142712	10.05
UPPER LIMIT	194384	6.038	327690	7.257	285424	10.549
LOWER LIMIT	48596	5.038	81922.5	6.257	71356	9.549
EPA SAMPLE NO.						
GDW1	62035	5.54	123496	6.76	115828	10.05
GDW2	61746	5.54	125249	6.76	113628	10.05
VX0519WBL01	62503	5.54	124684	6.76	117918	10.05
VX0519WBS01	91585	5.54	158835	6.76	142753	10.05
VX0519WBSD01	87263	5.54	150978	6.76	132315	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG NO.: Q2073
 Lab File ID: VX046255.D Date Analyzed: 05/19/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:05
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	70270	12.018				
UPPER LIMIT	140540	12.518				
LOWER LIMIT	35135	11.518				
EPA SAMPLE NO.						
GDW1	49714	12.02				
GDW2	45883	12.02				
VX0519WBL01	48295	12.02				
VX0519WBS01	66879	12.02				
VX0519WBSD01	63585	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.



SAMPLE DATA

Report of Analysis

Client:	G Environmental		Date Collected:	05/16/25	
Project:	Nelson		Date Received:	05/16/25	
Client Sample ID:	GDW1		SDG No.:	Q2073	
Lab Sample ID:	Q2073-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046280.D	1		05/19/25 20:06	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	6.80		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	3.50		0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	05/16/25	
Project:	Nelson		Date Received:	05/16/25	
Client Sample ID:	GDW1		SDG No.:	Q2073	
Lab Sample ID:	Q2073-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046280.D	1		05/19/25 20:06	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	53.5		70 (74) - 130 (125)	107%	SPK: 50
1868-53-7	Dibromofluoromethane	50.8		70 (75) - 130 (124)	102%	SPK: 50
2037-26-5	Toluene-d8	51.2		70 (86) - 130 (113)	102%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.5		70 (77) - 130 (121)	101%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	62000	5.544			
540-36-3	1,4-Difluorobenzene	123000	6.757			
3114-55-4	Chlorobenzene-d5	116000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	49700	12.018			

Report of Analysis

Client:	G Environmental		Date Collected:	05/16/25	
Project:	Nelson		Date Received:	05/16/25	
Client Sample ID:	GDW1		SDG No.:	Q2073	
Lab Sample ID:	Q2073-01		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046280.D	1		05/19/25 20:06	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046280.D
 Acq On : 19 May 2025 20:06
 Operator : JC/MD
 Sample : Q2073-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GDW1

Quant Time: May 19 23:18:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	62035	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	123496	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	115828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49714	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	61896	53.519	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.040%			
35) Dibromofluoromethane	5.379	113	45187	50.812	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.620%			
50) Toluene-d8	8.647	98	157467	51.159	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.320%			
62) 4-Bromofluorobenzene	11.079	95	59592	50.473	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 100.940%			
Target Compounds						
16) Acetone	2.380	43	3138	6.767	ug/l	95
18) Methyl Acetate	2.709	43	3775	3.508	ug/l	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046280.D
Acq On : 19 May 2025 20:06
Operator : JC/MD
Sample : Q2073-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GDW1

Quant Time: May 19 23:18:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046280.D
Acq: 19 May 2025 20:06




Instrument :
MSVOA_X
ClientSampleId :
GDW1


Tgt Ion:168 Resp: 62035
Ion Ratio Lower Upper
168 100
99 73.4 54.9 82.3



#16
Acetone
Concen: 6.767 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.006 min
Lab File: VX046280.D
Acq: 19 May 2025 20:06

Tgt Ion: 43 Resp: 3138
Ion Ratio Lower Upper
43 100
58 24.1 21.2 31.8





#18

Methyl Acetate

Concen: 3.508 ug/l

RT: 2.709 min Scan# 2

Delta R.T. 0.006 min

Lab File: VX046280.D

Acq: 19 May 2025 20:06

Instrument :

MSVOA_X

ClientSampleId :





GDW1

Tgt Ion: 43 Resp: 3775

Ion Ratio Lower Upper

43 100

74 20.8 16.7 25.1



#33

1,2-Dichloroethane-d4

Concen: 53.519 ug/l

RT: 5.946 min Scan# 797

Delta R.T. -0.006 min

Lab File: VX046280.D




Acq: 19 May 2025 20:06


Tgt Ion: 65 Resp: 61896

Ion Ratio Lower Upper

65 100

67 49.1 0.0 99.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX046280.D
Acq: 19 May 2025 20:06




Instrument :
MSVOA_X
ClientSampleId :
GDW1


Tgt Ion	Ratio	Lower	Upper
114	100		
63	23.8	0.0	49.2
88	17.5	0.0	33.6



#35
Dibromofluoromethane
Concen: 50.812 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX046280.D
Acq: 19 May 2025 20:06

Tgt Ion	Ratio	Lower	Upper
113	100		
111	104.2	83.1	124.7
192	17.6	13.3	19.9





#50

Toluene-d8

Concen: 51.159 ug/l

RT: 8.647 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046280.D

Acq: 19 May 2025 20:06

Instrument :

MSVOA_X

ClientSampleId :





GDW1

Tgt Ion: 98 Resp: 157467

Ion Ratio Lower Upper

98 100

100 65.1 53.5 80.3



#62

4-Bromofluorobenzene

Concen: 50.473 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046280.D

Acq: 19 May 2025 20:06




Tgt Ion: 95 Resp: 59592


Ion Ratio Lower Upper

95 100

174 67.8 0.0 135.8

176 65.1 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046280.D
Acq: 19 May 2025 20:06




Instrument :
MSVOA_X
ClientSampleId :
GDW1

Tgt Ion:117 Resp: 115828
Ion Ratio Lower Upper
117 100
82 64.7 50.6 76.0
119 33.0 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046280.D
Acq: 19 May 2025 20:06

Tgt Ion:152 Resp: 49714
Ion Ratio Lower Upper
152 100
115 63.9 46.9 140.7
150 154.3 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046280.D
 Acq On : 19 May 2025 20:06
 Operator : JC/MD
 Sample : Q2073-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GDW1

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Title : SW846 8260

Signal : TIC: VX046280.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.575	70	80	81	rBV6	4766	9881	2.23%	0.421%
2	2.703	260	265	274	rVB	2720	4952	1.12%	0.211%
3	5.379	693	704	721	rBV	50865	153795	34.73%	6.556%
4	5.544	721	731	745	rVV	72177	208011	46.98%	8.867%
5	5.952	789	798	813	rBV	58306	160630	36.28%	6.847%
6	6.757	921	930	944	rBV	129296	316511	71.48%	13.491%
7	7.068	977	981	989	rVB4	3277	8214	1.85%	0.350%
8	8.647	1233	1240	1251	rBV	287005	442805	100.00%	18.875%
9	10.049	1465	1470	1483	rBV	282161	396414	89.52%	16.897%
10	11.079	1634	1639	1659	rBV	231665	293628	66.31%	12.516%
11	12.018	1788	1793	1807	rBV	261653	331270	74.81%	14.121%
12	12.219	1822	1826	1836	rBV3	7716	13591	3.07%	0.579%
13	15.670	2386	2392	2400	rBV5	2815	6306	1.42%	0.269%




Sum of corrected areas: 2346008

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046280.D
Acq On : 19 May 2025 20:06
Operator : JC/MD
Sample : Q2073-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GDW1

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046280.D
Acq On : 19 May 2025 20:06
Operator : JC/MD
Sample : Q2073-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GDW1

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046280.D
Acq On : 19 May 2025 20:06
Operator : JC/MD
Sample : Q2073-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GDW1

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc
------------------	----	---------	-------	----------	---	----	------	------

Report of Analysis

Client:	G Environmental	Date Collected:	05/16/25
Project:	Nelson	Date Received:	05/16/25
Client Sample ID:	GDW2	SDG No.:	Q2073
Lab Sample ID:	Q2073-02	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046281.D	1		05/19/25 20:29	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	1.10		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	5.20		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	1.40		0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	05/16/25	
Project:	Nelson		Date Received:	05/16/25	
Client Sample ID:	GDW2		SDG No.:	Q2073	
Lab Sample ID:	Q2073-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046281.D	1		05/19/25 20:29	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.3		70 (74) - 130 (125)	109%	SPK: 50
1868-53-7	Dibromofluoromethane	50.1		70 (75) - 130 (124)	100%	SPK: 50
2037-26-5	Toluene-d8	49.8		70 (86) - 130 (113)	100%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.7		70 (77) - 130 (121)	97%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	61700	5.544			
540-36-3	1,4-Difluorobenzene	125000	6.757			
3114-55-4	Chlorobenzene-d5	114000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	45900	12.018			

Report of Analysis

Client:	G Environmental		Date Collected:	05/16/25	
Project:	Nelson		Date Received:	05/16/25	
Client Sample ID:	GDW2		SDG No.:	Q2073	
Lab Sample ID:	Q2073-02		Matrix:	Water	
Analytical Method:	8260D		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	DB-624UI	ID : 0.18	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046281.D	1		05/19/25 20:29	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046281.D
 Acq On : 19 May 2025 20:29
 Operator : JC/MD
 Sample : Q2073-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GDW2

Quant Time: May 19 23:19:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.544	168	61746	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125249	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	113628	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	45883	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	62475	54.272	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.540%	
35) Dibromofluoromethane	5.379	113	45215	50.132	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.260%	
50) Toluene-d8	8.647	98	155594	49.843	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.680%	
62) 4-Bromofluorobenzene	11.079	95	58373	48.748	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 97.500%	
Target Compounds						
						Qvalue
3) Chloromethane	1.307	50	1006	1.098	ug/l	95
16) Acetone	2.386	43	2378	5.152	ug/l	94
18) Methyl Acetate	2.709	43	1455	1.358	ug/l #	80


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046281.D
Acq On : 19 May 2025 20:29
Operator : JC/MD
Sample : Q2073-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GDW2

Quant Time: May 19 23:19:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration





#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX046281.D

Acq: 19 May 2025 20:29

Instrument :

MSVOA_X

ClientSampleId :





GDW2

Tgt Ion:168 Resp: 61746

Ion Ratio Lower Upper

168 100

99 71.6 54.9 82.3



#3

Chloromethane

Concen: 1.098 ug/l

RT: 1.307 min Scan# 36

Delta R.T. -0.000 min

Lab File: VX046281.D




Acq: 19 May 2025 20:29


Tgt Ion: 50 Resp: 1006

Ion Ratio Lower Upper

50 100

52 34.6 25.4 38.2









#16
Acetone
Concen: 5.152 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX046281.D
Acq: 19 May 2025 20:29




Instrument :
MSVOA_X
ClientSampleId :
GDW2


Tgt Ion: 43 Resp: 2378
Ion Ratio Lower Upper
43 100
58 23.2 21.2 31.8



#18
Methyl Acetate
Concen: 1.358 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.006 min
Lab File: VX046281.D
Acq: 19 May 2025 20:29

Tgt Ion: 43 Resp: 1455
Ion Ratio Lower Upper
43 100
74 11.3 16.7 25.1#





#33

1,2-Dichloroethane-d4

Concen: 54.272 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046281.D

Acq: 19 May 2025 20:29

Instrument :

MSVOA_X

ClientSampleId :





GDW2

Tgt Ion: 65 Resp: 62475

Ion Ratio Lower Upper

65 100

67 49.3 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX046281.D

Acq: 19 May 2025 20:29




Tgt Ion: 114 Resp: 125249


Ion Ratio Lower Upper

114 100

63 24.6 0.0 49.2

88 16.9 0.0 33.6





#35

Dibromofluoromethane

Concen: 50.132 ug/l

RT: 5.379 min Scan# 704

Delta R.T. -0.000 min

Lab File: VX046281.D


Acq: 19 May 2025 20:29

Instrument :

MSVOA_X




ClientSampleId :

GDW2



Tgt Ion:113 Resp: 45215

Ion	Ratio	Lower	Upper
113	100		
111	104.3	83.1	124.7
192	16.1	13.3	19.9



#50

Toluene-d8

Concen: 49.843 ug/l

RT: 8.647 min Scan# 1240




Delta R.T. -0.000 min


Lab File: VX046281.D

Acq: 19 May 2025 20:29

Tgt Ion: 98 Resp: 155594

Ion	Ratio	Lower	Upper
98	100		
100	65.7	53.5	80.3









#62
4-Bromofluorobenzene
Concen: 48.748 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX046281.D
Acq: 19 May 2025 20:29




Instrument :
MSVOA_X
ClientSampleId :
GDW2


Tgt Ion: 95 Resp: 58373
Ion Ratio Lower Upper
95 100
174 67.0 0.0 135.8
176 63.2 0.0 131.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046281.D
Acq: 19 May 2025 20:29

Tgt Ion: 117 Resp: 113628
Ion Ratio Lower Upper
117 100
82 62.6 50.6 76.0
119 33.0 25.8 38.6





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046281.D

Acq: 19 May 2025 20:29

Instrument :

MSVOA_X

ClientSampleId :

GDW2




Tgt Ion:152 Resp: 45883

Ion Ratio Lower Upper

152 100

115 66.3 46.9 140.7

150 159.9 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046281.D
Acq On : 19 May 2025 20:29
Operator : JC/MD
Sample : Q2073-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GDW2

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046281.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.575	69	80	81	rBV5	5212	10979	2.51%	0.480%
2	5.379	693	704	720	rBV	50868	153216	34.99%	6.694%
3	5.544	722	731	745	rVV2	71389	205979	47.04%	8.999%
4	5.952	788	798	814	rBV	60732	162871	37.20%	7.116%
5	6.757	921	930	943	rBV	130019	318648	72.77%	13.922%
6	8.647	1232	1240	1250	rBV	279893	437863	100.00%	19.130%
7	10.049	1465	1470	1487	rBV	280096	390775	89.25%	17.073%
8	11.079	1634	1639	1652	rBV	218484	284311	64.93%	12.422%
9	12.018	1788	1793	1806	rBV	247776	314680	71.87%	13.748%
10	12.219	1820	1826	1835	rBV3	4650	9540	2.18%	0.417%




Sum of corrected areas: 2288862

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046281.D
Acq On : 19 May 2025 20:29
Operator : JC/MD
Sample : Q2073-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GDW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046281.D
Acq On : 19 May 2025 20:29
Operator : JC/MD
Sample : Q2073-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GDW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046281.D
Acq On : 19 May 2025 20:29
Operator : JC/MD
Sample : Q2073-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GDW2

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	--Internal Standard--			
					#	RT	Resp	Conc



CALIBRATION SUMMARY



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG No.: Q2073
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D			
	RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D			
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD
Dichlorodifluoromethane	0.697	0.864	0.859	0.875	0.639	0.658	0.765	14.6
Chloromethane	0.727	0.775	0.787	0.791	0.679	0.694	0.742	6.6
Vinyl Chloride	0.660	0.710	0.727	0.755	0.619	0.673	0.691	7.2
Bromomethane	0.296	0.326	0.340	0.334	0.305		0.320	5.8
Chloroethane	0.354	0.378	0.329	0.317	0.368	0.467	0.369	14.4
Trichlorofluoromethane	1.035	1.068	0.983	0.985	0.990	1.064	1.021	3.9
1,1,2-Trichlorotrifluoroethane	0.628	0.641	0.629	0.648	0.610	0.633	0.632	2.1
1,1-Dichloroethene	0.565	0.601	0.607	0.625	0.567	0.594	0.593	3.9
Acetone	0.361	0.362	0.361	0.370	0.408	0.380	0.374	4.9
Carbon Disulfide	1.295	1.455	1.522	1.597	1.141	1.423	1.406	11.7
Methyl tert-butyl Ether	2.044	2.160	2.172	2.239	1.908	1.949	2.079	6.4
Methyl Acetate	0.814	0.848	0.845	0.875	0.816	1.006	0.867	8.3
Methylene Chloride	0.689	0.684	0.691	0.691	0.689	0.853	0.716	9.4
trans-1,2-Dichloroethene	0.573	0.610	0.612	0.622	0.557	0.604	0.596	4.3
1,1-Dichloroethane	1.233	1.263	1.263	1.286	1.154	1.116	1.219	5.6
Cyclohexane	1.090	1.128	1.128	1.150	1.059		1.111	3.3
2-Butanone	0.540	0.555	0.558	0.569	0.539	0.495	0.543	4.8
Carbon Tetrachloride	0.528	0.558	0.552	0.577	0.505	0.541	0.544	4.6
cis-1,2-Dichloroethene	0.716	0.737	0.738	0.755	0.642	0.719	0.718	5.5
Bromochloromethane	0.628	0.578	0.595	0.590	0.553	0.576	0.587	4.3
Chloroform	1.287	1.296	1.277	1.300	1.199	1.265	1.271	3
1,1,1-Trichloroethane	1.106	1.131	1.155	1.188	1.013	1.015	1.101	6.6
Methylcyclohexane	0.596	0.641	0.627	0.658	0.587	0.627	0.623	4.3
Benzene	1.426	1.474	1.441	1.477	1.337	1.348	1.417	4.3
1,2-Dichloroethane	0.632	0.627	0.611	0.625	0.594	0.579	0.612	3.5
Trichloroethene	0.344	0.355	0.345	0.362	0.315	0.324	0.341	5.3
1,2-Dichloropropane	0.356	0.371	0.368	0.378	0.324	0.317	0.352	7.4
Bromodichloromethane	0.557	0.577	0.573	0.594	0.498	0.485	0.547	8.2
4-Methyl-2-Pentanone	0.620	0.634	0.630	0.631	0.555	0.561	0.605	6
Toluene	0.884	0.898	0.885	0.904	0.838	0.803	0.869	4.5

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG No.: Q2073
 Instrument ID: MSVOA_X Calibration Date(s): 05/05/2025 05/05/2025
 Heated Purge: (Y/N) N Calibration Time(s): 11:35 16:27
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF020 = VX046041.D		RRF050 = VX046042.D		RRF100 = VX046043.D		RRF150 = VX046044.D		RRF005 = VX046046.D		RRF001 = VX046047.D	
COMPOUND	RRF020	RRF050	RRF100	RRF150	RRF005	RRF001	RRF	% RSD				
t-1,3-Dichloropropene	0.468	0.528	0.555	0.591	0.406	0.371	0.487	17.9				
cis-1,3-Dichloropropene	0.531	0.578	0.602	0.623	0.469	0.423	0.538	14.6				
1,1,2-Trichloroethane	0.349	0.354	0.351	0.356	0.337	0.308	0.343	5.3				
2-Hexanone	0.466	0.473	0.477	0.473	0.414	0.385	0.448	8.7				
Dibromochloromethane	0.378	0.400	0.415	0.431	0.326	0.306	0.376	13.3				
1,2-Dibromoethane	0.359	0.373	0.368	0.381	0.333	0.322	0.356	6.5				
Tetrachloroethene	0.390	0.375	0.345	0.344	0.323	0.347	0.354	6.8				
Chlorobenzene	1.093	1.098	1.085	1.114	1.046	1.131	1.094	2.7				
Ethyl Benzene	1.919	2.022	1.979	2.036	1.816	1.803	1.929	5.2				
m/p-Xylenes	0.706	0.740	0.721	0.740	0.678	0.648	0.706	5.2				
o-Xylene	0.688	0.727	0.706	0.726	0.639	0.642	0.688	5.7				
Styrene	1.135	1.219	1.214	1.230	1.012	0.951	1.127	10.6				
Bromoform	0.270	0.304	0.312	0.327	0.236	0.234	0.281	14.2				
Isopropylbenzene	3.843	4.130	3.876	4.156	3.562	3.789	3.893	5.7				
1,1,2,2-Tetrachloroethane	1.315	1.338	1.284	1.345	1.350	1.552	1.364	7				
1,3-Dichlorobenzene	1.633	1.701	1.656	1.730	1.558	1.619	1.649	3.7				
1,4-Dichlorobenzene	1.629	1.693	1.639	1.722	1.606	1.817	1.684	4.6				
1,2-Dichlorobenzene	1.613	1.696	1.634	1.702	1.577	1.710	1.655	3.3				
1,2-Dibromo-3-Chloropropane	0.299	0.322	0.329	0.356	0.248	0.259	0.302	13.9				
1,2,4-Trichlorobenzene	0.861	0.981	1.035	1.123	0.842	0.862	0.951	12				
1,2,3-Trichlorobenzene	0.921	1.019	1.051	1.107	0.846	0.941	0.981	9.7				
1,2-Dichloroethane-d4	0.953	0.910	0.930	0.932	0.935		0.932	1.6				
Dibromofluoromethane	0.359	0.355	0.364	0.368	0.354		0.360	1.7				
Toluene-d8	1.246	1.223	1.266	1.275	1.221		1.246	2				
4-Bromofluorobenzene	0.455	0.470	0.500	0.500	0.464		0.478	4.4				

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X050525W.M
 Title : SW846 8260
 Last Update : Tue May 06 07:12:22 2025
 Response Via : Initial Calibration

Calibration Files

1 =VX046047.D 5 =VX046046.D 20 =VX046041.D 50 =VX046042.D 100 =VX046043.D 150 =VX046044.D

Compound		1	5	20	50	100	150	Avg	%RSD

1) I	Pentafluorobenzene	-----ISTD-----							
2) T	Dichlorodifluo...	0.658	0.639	0.697	0.864	0.859	0.875	0.765	14.62
3) P	Chloromethane	0.694	0.679	0.727	0.775	0.787	0.791	0.742	6.58
4) C	Vinyl Chloride	0.673	0.619	0.660	0.710	0.727	0.755	0.691	7.17#
5) T	Bromomethane		0.305	0.296	0.326	0.340	0.334	0.320	5.84
6) T	Chloroethane	0.467	0.368	0.354	0.378	0.329	0.317	0.369	14.44
7) T	Trichlorofluor...	1.064	0.990	1.035	1.068	0.983	0.985	1.021	3.92
8) T	Diethyl Ether	0.403	0.311	0.340	0.337	0.338	0.355	0.347	8.83
9) T	1,1,2-Trichlor...	0.633	0.610	0.628	0.641	0.629	0.648	0.632	2.10
10) T	Methyl Iodide		0.608	0.767	0.806	0.793	0.763	0.747	10.68
11) T	Tert butyl alc...		0.114	0.122	0.129	0.144	0.146	0.131	10.45
12) CM	1,1-Dichloroet...	0.594	0.567	0.565	0.601	0.607	0.625	0.593	3.94#
13) T	Acrolein		0.117	0.158	0.152	0.154	0.163	0.149	12.33
14) T	Allyl chloride	1.052	1.058	1.127	1.179	1.187	1.196	1.133	5.75
15) T	Acrylonitrile	0.363	0.345	0.378	0.388	0.381	0.390	0.374	4.52
16) T	Acetone	0.380	0.408	0.361	0.362	0.361	0.370	0.374	4.90
17) T	Carbon Disulfide	1.423	1.141	1.295	1.455	1.522	1.597	1.406	11.68
18) T	Methyl Acetate	1.006	0.816	0.814	0.848	0.845	0.875	0.867	8.27
19) T	Methyl tert-bu...	1.949	1.908	2.044	2.160	2.172	2.239	2.079	6.39
20) T	Methylene Chlo...	0.853	0.689	0.689	0.684	0.691	0.691	0.716	9.39
21) T	trans-1,2-Dich...	0.604	0.557	0.573	0.610	0.612	0.622	0.596	4.27
22) T	Diisopropyl ether	2.095	1.924	2.219	2.278	2.295	2.321	2.189	6.97
23) T	Vinyl Acetate	1.660	1.698	1.928	2.048	2.082	2.134	1.925	10.52
24) P	1,1-Dichloroet...	1.116	1.154	1.233	1.263	1.263	1.286	1.219	5.60
25) T	2-Butanone	0.495	0.539	0.540	0.555	0.558	0.569	0.543	4.77
26) T	2,2-Dichloropr...	0.965	0.850	0.910	0.957	1.003	1.039	0.954	7.03
27) T	cis-1,2-Dichlo...	0.719	0.642	0.716	0.737	0.738	0.755	0.718	5.55
28) T	Bromochloromet...	0.576	0.553	0.628	0.578	0.595	0.590	0.587	4.28
29) T	Tetrahydrofuran	0.318	0.318	0.340	0.350	0.351	0.362	0.340	5.36
30) C	Chloroform	1.265	1.199	1.287	1.296	1.277	1.300	1.271	2.95#
31) T	Cyclohexane		1.059	1.090	1.128	1.128	1.150	1.111	3.26
32) T	1,1,1-Trichlor...	1.015	1.013	1.106	1.131	1.155	1.188	1.101	6.60
33) S	1,2-Dichloroet...		0.935	0.953	0.910	0.930	0.932	0.932	1.65
34) I	1,4-Difluorobenzene	-----ISTD-----							
35) S	Dibromofluorom...		0.354	0.359	0.355	0.364	0.368	0.360	1.70
36) T	1,1-Dichloropr...	0.493	0.462	0.463	0.495	0.483	0.505	0.484	3.68
37) T	Ethyl Acetate	0.586	0.582	0.569	0.611	0.609	0.631	0.598	3.83
38) T	Carbon Tetrach...	0.541	0.505	0.528	0.558	0.552	0.577	0.544	4.61
39) T	Methylcyclohexane	0.627	0.587	0.596	0.641	0.627	0.658	0.623	4.34
40) TM	Benzene	1.348	1.337	1.426	1.474	1.441	1.477	1.417	4.30
41) T	Methacrylonitrile	0.233	0.288	0.318	0.346	0.343	0.348	0.313	14.50
42) TM	1,2-Dichloroet...	0.579	0.594	0.632	0.627	0.611	0.625	0.612	3.45
43) T	Isopropyl Acetate	0.764	0.826	0.905	0.963	0.982	1.030	0.912	11.05
44) TM	Trichloroethene	0.324	0.315	0.344	0.355	0.345	0.362	0.341	5.28
45) C	1,2-Dichloropr...	0.317	0.324	0.356	0.371	0.368	0.378	0.352	7.37#
46) T	Dibromomethane	0.263	0.262	0.285	0.287	0.280	0.289	0.278	4.35
47) T	Bromodichlorom...	0.485	0.498	0.557	0.577	0.573	0.594	0.547	8.22
48) T	Methyl methacr...	0.370	0.426	0.465	0.502	0.500	0.531	0.466	12.75
49) T	1,4-Dioxane	0.007	0.009	0.009	0.009	0.009	0.010	0.009	8.45
50) S	Toluene-d8		1.221	1.246	1.223	1.266	1.275	1.246	1.95
51) T	4-Methyl-2-Pen...	0.561	0.555	0.620	0.634	0.630	0.631	0.605	6.04
52) CM	Toluene	0.803	0.838	0.884	0.898	0.885	0.904	0.869	4.54#
53) T	t-1,3-Dichloro...	0.371	0.406	0.468	0.528	0.555	0.591	0.487	17.86
54) T	cis-1,3-Dichlo...	0.423	0.469	0.531	0.578	0.602	0.623	0.538	14.61
55) T	1,1,2-Trichlor...	0.308	0.337	0.349	0.354	0.351	0.356	0.343	5.30
56) T	Ethyl methacry...	0.377	0.508	0.540	0.595	0.617	0.639	0.546	17.58

Method Path : Z:\voasrv\HPCHEM1\MSVOA_X\Method\
 Method File : 82X050525W.M

57)	T	1,3-Dichloropr...	0.610	0.601	0.618	0.623	0.613	0.627	0.615	1.53
58)	T	2-Chloroethyl ...	0.230	0.247	0.270	0.307	0.303	0.313	0.278	12.41
59)	T	2-Hexanone	0.385	0.414	0.466	0.473	0.477	0.473	0.448	8.69
60)	T	Dibromochlorom...	0.306	0.326	0.378	0.400	0.415	0.431	0.376	13.28
61)	T	1,2-Dibromoethane	0.322	0.333	0.359	0.373	0.368	0.381	0.356	6.54
62)	S	4-Bromofluorob...		0.464	0.455	0.470	0.500	0.500	0.478	4.41
63)	I	Chlorobenzene-d5	-----ISTD-----							
64)	T	Tetrachloroethene	0.347	0.323	0.390	0.375	0.345	0.344	0.354	6.84
65)	PM	Chlorobenzene	1.131	1.046	1.093	1.098	1.085	1.114	1.094	2.65
66)	T	1,1,1,2-Tetrac...	0.369	0.341	0.365	0.390	0.382	0.395	0.374	5.22
67)	C	Ethyl Benzene	1.803	1.816	1.919	2.022	1.979	2.036	1.929	5.24#
68)	T	m/p-Xylenes	0.648	0.678	0.706	0.740	0.721	0.740	0.706	5.21
69)	T	o-Xylene	0.642	0.639	0.688	0.727	0.706	0.726	0.688	5.75
70)	T	Styrene	0.951	1.012	1.135	1.219	1.214	1.230	1.127	10.56
71)	P	Bromoform	0.234	0.236	0.270	0.304	0.312	0.327	0.281	14.17
72)	I	1,4-Dichlorobenzen...	-----ISTD-----							
73)	T	Isopropylbenzene	3.789	3.562	3.843	4.130	3.876	4.156	3.893	5.72
74)	T	N-amyl acetate	1.715	1.652	1.846	2.067	2.068	2.192	1.924	11.32
75)	P	1,1,2,2-Tetrac...	1.552	1.350	1.315	1.338	1.284	1.345	1.364	6.97
76)	T	1,2,3-Trichlor...	1.405	1.167	1.151	1.187	1.131	1.181	1.204	8.36
77)	T	Bromobenzene	0.926	0.862	0.896	0.928	0.883	0.928	0.904	3.08
78)	T	n-propylbenzene	4.272	4.186	4.394	4.854	4.583	4.868	4.526	6.45
79)	T	2-Chlorotoluene	3.184	2.748	2.832	2.994	2.805	2.953	2.919	5.45
80)	T	1,3,5-Trimethy...	3.036	3.053	3.275	3.487	3.255	3.405	3.252	5.60
81)	T	trans-1,4-Dich...		0.269	0.335	0.385	0.410	0.449	0.370	18.91
82)	T	4-Chlorotoluene	3.226	2.939	3.196	3.430	3.255	3.379	3.238	5.32
83)	T	tert-Butylbenzene	3.341	3.098	3.115	3.435	3.255	3.411	3.276	4.44
84)	T	1,2,4-Trimethy...	3.150	3.034	3.274	3.522	3.335	3.444	3.293	5.52
85)	T	sec-Butylbenzene	3.708	3.767	3.937	4.282	4.095	4.343	4.022	6.55
86)	T	p-Isopropyltol...	3.025	3.084	3.206	3.555	3.450	3.599	3.320	7.44
87)	T	1,3-Dichlorobe...	1.619	1.558	1.633	1.701	1.656	1.729	1.649	3.71
88)	T	1,4-Dichlorobe...	1.817	1.606	1.629	1.693	1.639	1.722	1.684	4.64
89)	T	n-Butylbenzene	2.443	2.650	2.748	3.147	3.139	3.346	2.912	12.00
90)	T	Hexachloroethane	0.511	0.523	0.551	0.622	0.622	0.680	0.585	11.44
91)	T	1,2-Dichlorobe...	1.710	1.577	1.613	1.696	1.634	1.702	1.655	3.34
92)	T	1,2-Dibromo-3-...	0.259	0.248	0.299	0.322	0.329	0.356	0.302	13.89
93)	T	1,2,4-Trichlor...	0.862	0.842	0.861	0.981	1.035	1.123	0.951	12.03
94)	T	Hexachlorobuta...	0.393	0.414	0.394	0.427	0.418	0.445	0.415	4.79
95)	T	Naphthalene	3.499	2.929	3.204	3.613	3.690	3.984	3.487	10.69
96)	T	1,2,3-Trichlor...	0.941	0.846	0.921	1.019	1.051	1.107	0.981	9.74

(#) = Out of Range

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	83671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	147096	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	127829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60503	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	31909	12.820	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	25.640%#		
35) Dibromofluoromethane	5.385	113	21112	12.804	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	25.600%#		
50) Toluene-d8	8.646	98	73333	13.200	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	26.400%#		
62) 4-Bromofluorobenzene	11.079	95	26760	13.236	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	26.480%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	23341	12.638	ug/l	96
3) Chloromethane	1.306	50	24316	13.371	ug/l	98
4) Vinyl Chloride	1.373	62	22096	13.868	ug/l	94
5) Bromomethane	1.593	94	9920	12.279	ug/l	93
6) Chloroethane	1.666	64	11832	14.414	ug/l	95
7) Trichlorofluoromethane	1.873	101	34636	14.271	ug/l	100
8) Diethyl Ether	2.129	74	11392	14.278	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.324	101	21027	14.499	ug/l	98
10) Methyl Iodide	2.446	142	25661	15.070	ug/l	99
11) Tert butyl alcohol	2.971	59	20370	68.662	ug/l	97
12) 1,1-Dichloroethene	2.312	96	18899	13.385	ug/l	98
13) Acrolein	2.233	56	26427	76.689	ug/l	95
14) Allyl chloride	2.660	41	37708	14.133	ug/l	99
15) Acrylonitrile	3.062	53	63290	71.312	ug/l	98
16) Acetone	2.379	43	60363	70.884	ug/l	99
17) Carbon Disulfide	2.501	76	43348	13.251	ug/l	100
18) Methyl Acetate	2.702	43	27234	13.397	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	68402	13.890	ug/l	100
20) Methylene Chloride	2.782	84	23050	13.362	ug/l	96
21) trans-1,2-Dichloroethene	3.086	96	19161	13.294	ug/l	96
22) Diisopropyl ether	3.763	45	74257	14.896	ug/l	95
23) Vinyl Acetate	3.720	43	322598	73.230	ug/l	100
24) 1,1-Dichloroethane	3.605	63	41257	14.082	ug/l	98
25) 2-Butanone	4.556	43	90331	73.489	ug/l	98
26) 2,2-Dichloropropane	4.470	77	30472	13.729	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	23970	13.791	ug/l	98
28) Bromochloromethane	4.897	49	21028	13.497	ug/l	99
29) Tetrahydrofuran	5.007	42	56819	71.175	ug/l	100
30) Chloroform	5.086	83	43065	14.144	ug/l	95
31) Cyclohexane	5.458	56	36470	14.769	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	37003	14.092	ug/l	98
36) 1,1-Dichloropropene	5.684	75	27242	14.231	ug/l	98
37) Ethyl Acetate	4.720	43	33458	13.931	ug/l	98
38) Carbon Tetrachloride	5.671	117	31091	14.137	ug/l	98
39) Methylcyclohexane	7.372	83	35077	14.623	ug/l	99
40) Benzene	6.031	78	83898	14.140	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046041.D
 Acq On : 05 May 2025 11:35
 Operator : JC/MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC020

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.915	41	18732	14.038	ug/l	95
42) 1,2-Dichloroethane	6.080	62	37203	15.165	ug/l	100
43) Isopropyl Acetate	6.342	43	53275	14.512	ug/l	99
44) Trichloroethene	7.116	130	20239	14.429	ug/l	99
45) 1,2-Dichloropropane	7.427	63	20940	14.204	ug/l	98
46) Dibromomethane	7.580	93	16781	14.439	ug/l	99
47) Bromodichloromethane	7.817	83	32762	14.545	ug/l	94
48) Methyl methacrylate	7.689	41	27374	14.521	ug/l	100
49) 1,4-Dioxane	7.659	88	10585	282.442	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	182455	75.775	ug/l	98
52) Toluene	8.713	92	52030	14.788	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	27549	14.566	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	31226	14.154	ug/l	98
55) 1,1,2-Trichloroethane	9.152	97	20536	14.442	ug/l	96
56) Ethyl methacrylate	9.116	69	31787	14.425	ug/l	98
57) 1,3-Dichloropropane	9.305	76	36341	14.417	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	79531	80.312	ug/l	100
59) 2-Hexanone	9.427	43	137054	74.997	ug/l	99
60) Dibromochloromethane	9.518	129	22267	14.387	ug/l	99
61) 1,2-Dibromoethane	9.610	107	21113	14.458	ug/l	96
64) Tetrachloroethene	9.268	164	19924	15.263	ug/l	97
65) Chlorobenzene	10.079	112	55863	14.202	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.158	131	18671	14.231	ug/l	98
67) Ethyl Benzene	10.189	91	98113	14.631	ug/l	100
68) m/p-Xylenes	10.299	106	72202	29.806	ug/l	99
69) o-Xylene	10.640	106	35178	14.408	ug/l	96
70) Styrene	10.652	104	58034	14.853	ug/l	98
71) Bromoform	10.799	173	13825	14.030	ug/l #	100
73) Isopropylbenzene	10.957	105	93013	14.096	ug/l	99
74) N-amyl acetate	10.841	43	44684	13.581	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	31830	13.560	ug/l	99
76) 1,2,3-Trichloropropane	11.237	75	27845m	11.039	ug/l	
77) Bromobenzene	11.195	156	21695	14.212	ug/l	96
78) n-propylbenzene	11.298	91	106337	14.376	ug/l	99
79) 2-Chlorotoluene	11.359	91	68534	13.748	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	79261	14.468	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	8104	13.371	ug/l	91
82) 4-Chlorotoluene	11.451	91	77338	14.116	ug/l	100
83) tert-Butylbenzene	11.713	119	75382	14.011	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	79232	14.480	ug/l	99
85) sec-Butylbenzene	11.890	105	95271	14.263	ug/l	100
86) p-Isopropyltoluene	12.006	119	77599	14.387	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	39520	13.898	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	39419	14.146	ug/l	99
89) n-Butylbenzene	12.329	91	66506	14.184	ug/l	99
90) Hexachloroethane	12.536	117	13332	13.287	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	39031	14.092	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.938	75	7232	13.535	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	20846	13.745	ug/l	99
94) Hexachlorobutadiene	13.725	225	9524	13.426	ug/l	97
95) Naphthalene	13.774	128	77542	14.040	ug/l	99
96) 1,2,3-Trichlorobenzene	13.956	180	22301	14.023	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046041.D
Acq On : 05 May 2025 11:35
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046041.D
Acq On : 05 May 2025 11:35
Operator : JC/MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC020

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:08:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	97076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	146257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67976	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88371	30.601	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	61.200%#		
35) Dibromofluoromethane	5.379	113	59550	31.633	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	63.260%#		
50) Toluene-d8	8.647	98	205479	32.395	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	64.780%#		
62) 4-Bromofluorobenzene	11.079	95	79012	34.229	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	68.460%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	83826	39.120	ug/l	100
3) Chloromethane	1.307	50	75238	35.658	ug/l	100
4) Vinyl Chloride	1.374	62	68901	37.273	ug/l	100
5) Bromomethane	1.593	94	31648	33.765	ug/l	100
6) Chloroethane	1.666	64	36722	38.558	ug/l	100
7) Trichlorofluoromethane	1.874	101	103693	36.824	ug/l	100
8) Diethyl Ether	2.136	74	32724	35.350	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	62251	36.997	ug/l	100
10) Methyl Iodide	2.447	142	78277	39.622	ug/l	100
11) Tert butyl alcohol	2.977	59	62557	181.745	ug/l	100
12) 1,1-Dichloroethene	2.313	96	58334	35.610	ug/l	100
13) Acrolein	2.233	56	73927	184.907	ug/l	100
14) Allyl chloride	2.654	41	114445	36.971	ug/l	100
15) Acrylonitrile	3.062	53	188304	182.874	ug/l	100
16) Acetone	2.386	43	175777	177.910	ug/l	100
17) Carbon Disulfide	2.502	76	141281	37.223	ug/l	100
18) Methyl Acetate	2.703	43	82347	34.915	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	209665	36.695	ug/l	100
20) Methylene Chloride	2.782	84	66412	33.182	ug/l	100
21) trans-1,2-Dichloroethene	3.087	96	59234	35.421	ug/l	100
22) Diisopropyl ether	3.757	45	221172	38.241	ug/l	100
23) Vinyl Acetate	3.721	43	994236	194.528	ug/l	100
24) 1,1-Dichloroethane	3.605	63	122601	36.067	ug/l	100
25) 2-Butanone	4.556	43	269224	188.783	ug/l	100
26) 2,2-Dichloropropane	4.465	77	92857	36.059	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71518	35.466	ug/l	100
28) Bromochloromethane	4.898	49	56091	31.031	ug/l	100
29) Tetrahydrofuran	5.001	42	170124	183.679	ug/l	100
30) Chloroform	5.087	83	125850	35.626	ug/l	100
31) Cyclohexane	5.464	56	109459	38.205	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	109781	36.034	ug/l	100
36) 1,1-Dichloropropene	5.684	75	83215	38.073	ug/l	100
37) Ethyl Acetate	4.715	43	102537	37.394	ug/l	100
38) Carbon Tetrachloride	5.672	117	93712	37.320	ug/l	100
39) Methylcyclohexane	7.373	83	107651	39.307	ug/l	100
40) Benzene	6.031	78	247476	36.530	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	58082	38.123	ug/l	100
42) 1,2-Dichloroethane	6.080	62	105332	37.605	ug/l	100
43) Isopropyl Acetate	6.336	43	161787	38.598	ug/l	100
44) Trichloroethene	7.123	130	59623	37.230	ug/l	100
45) 1,2-Dichloropropane	7.428	63	62387	37.064	ug/l	100
46) Dibromomethane	7.574	93	48201	36.326	ug/l	100
47) Bromodichloromethane	7.818	83	96916	37.685	ug/l	100
48) Methyl methacrylate	7.690	41	84277	39.157	ug/l	100
49) 1,4-Dioxane	7.659	88	30529	713.477	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	532388	193.653	ug/l	100
52) Toluene	8.714	92	150757	37.530	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	88655	41.056	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	97031	38.521	ug/l	100
55) 1,1,2-Trichloroethane	9.147	97	59499	36.648	ug/l	100
56) Ethyl methacrylate	9.116	69	99947	39.726	ug/l	100
57) 1,3-Dichloropropane	9.305	76	104606	36.347	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	257678	227.902	ug/l	100
59) 2-Hexanone	9.427	43	396784	190.168	ug/l	100
60) Dibromochloromethane	9.519	129	67198	38.027	ug/l	100
61) 1,2-Dibromoethane	9.604	107	62633	37.567	ug/l	100
64) Tetrachloroethene	9.269	164	54853	36.726	ug/l	100
65) Chlorobenzene	10.079	112	160600	35.686	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	57066	38.017	ug/l	100
67) Ethyl Benzene	10.189	91	295712	38.543	ug/l	100
68) m/p-Xylenes	10.299	106	216578	78.142	ug/l	100
69) o-Xylene	10.640	106	106335	38.064	ug/l	100
70) Styrene	10.653	104	178313	39.888	ug/l	100
71) Bromoform	10.799	173	44486	39.457	ug/l #	100
73) Isopropylbenzene	10.957	105	280719	37.865	ug/l	100
74) N-amyl acetate	10.842	43	140509	38.012	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	90975	34.495	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	80678m	28.468	ug/l	
77) Bromobenzene	11.195	156	63053	36.763	ug/l	100
78) n-propylbenzene	11.299	91	329981	39.707	ug/l	100
79) 2-Chlorotoluene	11.360	91	203509	36.336	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	237066	38.516	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	26179	38.445	ug/l	100
82) 4-Chlorotoluene	11.451	91	233172	37.880	ug/l	100
83) tert-Butylbenzene	11.713	119	233468	38.622	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	239433	38.946	ug/l	100
85) sec-Butylbenzene	11.890	105	291074	38.786	ug/l	100
86) p-Isopropyltoluene	12.006	119	241658	39.879	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	115651	36.199	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	115097	36.763	ug/l	100
89) n-Butylbenzene	12.329	91	213899	40.603	ug/l	100
90) Hexachloroethane	12.536	117	42314	37.534	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	115286	37.047	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21909	36.496	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	66655	39.117	ug/l	100
94) Hexachlorobutadiene	13.725	225	29007	36.395	ug/l	100
95) Naphthalene	13.774	128	245568	39.575	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	69262	38.765	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046042.D
Acq On : 05 May 2025 11:58
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	87028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	152958	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	135214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	161877	62.526	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 125.060%#			
35) Dibromofluoromethane	5.379	113	111455	65.006	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 130.020%#			
50) Toluene-d8	8.647	98	387230	67.031	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 134.060%#			
62) 4-Bromofluorobenzene	11.079	95	153085	72.818	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 145.640%#			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	149508	77.828	ug/l	99
3) Chloromethane	1.307	50	136937	72.392	ug/l	98
4) Vinyl Chloride	1.374	62	126609	76.398	ug/l	98
5) Bromomethane	1.593	94	59246	70.507	ug/l	99
6) Chloroethane	1.660	64	57332	67.148	ug/l	99
7) Trichlorofluoromethane	1.867	101	171042	67.754	ug/l	99
8) Diethyl Ether	2.136	74	58762	70.807	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	109520	72.604	ug/l	99
10) Methyl Iodide	2.440	142	138045	77.943	ug/l	100
11) Tert butyl alcohol	2.995	59	125094	405.394	ug/l	100
12) 1,1-Dichloroethene	2.306	96	105566	71.884	ug/l	96
13) Acrolein	2.239	56	134456	375.132	ug/l	99
14) Allyl chloride	2.660	41	206519	74.419	ug/l	98
15) Acrylonitrile	3.068	53	331142	358.723	ug/l	97
16) Acetone	2.386	43	314383	354.936	ug/l	98
17) Carbon Disulfide	2.501	76	264956	77.867	ug/l	100
18) Methyl Acetate	2.703	43	147014	69.531	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	378113	73.818	ug/l	100
20) Methylene Chloride	2.782	84	120270	67.030	ug/l	96
21) trans-1,2-Dichloroethene	3.087	96	106606	71.109	ug/l	99
22) Diisopropyl ether	3.763	45	399494	77.048	ug/l	90
23) Vinyl Acetate	3.721	43	1811550	395.363	ug/l	100
24) 1,1-Dichloroethane	3.605	63	219759	72.114	ug/l	99
25) 2-Butanone	4.556	43	485448	379.704	ug/l	100
26) 2,2-Dichloropropane	4.464	77	174579	75.621	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	128454	71.056	ug/l	99
28) Bromochloromethane	4.891	49	103602	63.932	ug/l	100
29) Tetrahydrofuran	5.001	42	305757	368.234	ug/l	99
30) Chloroform	5.086	83	222253	70.180	ug/l	95
31) Cyclohexane	5.458	56	196302	76.428	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	201112	73.635	ug/l	98
36) 1,1-Dichloropropene	5.690	75	147873	74.285	ug/l	99
37) Ethyl Acetate	4.714	43	186215	74.565	ug/l	99
38) Carbon Tetrachloride	5.672	117	168946	73.875	ug/l	99
39) Methylcyclohexane	7.372	83	191941	76.952	ug/l	96
40) Benzene	6.031	78	440885	71.457	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046043.D
 Acq On : 05 May 2025 12:21
 Operator : JC/MD
 Sample : VSTDICC100
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC100

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	104863	75.574	ug/l	98
42) 1,2-Dichloroethane	6.080	62	187036	73.318	ug/l	99
43) Isopropyl Acetate	6.342	43	300491	78.714	ug/l	100
44) Trichloroethene	7.123	130	105628	72.420	ug/l	98
45) 1,2-Dichloropropane	7.427	63	112554	73.420	ug/l	100
46) Dibromomethane	7.574	93	85728	70.938	ug/l	100
47) Bromodichloromethane	7.818	83	175301	74.845	ug/l	99
48) Methyl methacrylate	7.689	41	153093	78.101	ug/l	99
49) 1,4-Dioxane	7.659	88	57313	1470.687	ug/l	99
51) 4-Methyl-2-Pentanone	8.573	43	963319	384.740	ug/l	100
52) Toluene	8.714	92	270763	74.010	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	169863	86.372	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	184134	80.264	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	107264	72.543	ug/l	99
56) Ethyl methacrylate	9.116	69	188717	82.359	ug/l	99
57) 1,3-Dichloropropane	9.305	76	187649	71.590	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	463169	449.791	ug/l	100
59) 2-Hexanone	9.433	43	729602	383.945	ug/l	100
60) Dibromochloromethane	9.518	129	127012	78.918	ug/l	98
61) 1,2-Dibromoethane	9.610	107	112669	74.200	ug/l	98
64) Tetrachloroethene	9.268	164	93195	67.493	ug/l	96
65) Chlorobenzene	10.079	112	293284	70.490	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	103177	74.349	ug/l	99
67) Ethyl Benzene	10.189	91	535122	75.444	ug/l	99
68) m/p-Xylenes	10.299	106	389935	152.181	ug/l	99
69) o-Xylene	10.640	106	190833	73.890	ug/l	97
70) Styrene	10.652	104	328195	79.411	ug/l	100
71) Bromoform	10.799	173	84353	80.927	ug/l #	98
73) Isopropylbenzene	10.957	105	514528	71.083	ug/l	100
74) N-amyl acetate	10.841	43	274553	76.073	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	170481	66.207	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	150182m	54.277	ug/l	
77) Bromobenzene	11.195	156	117151	69.959	ug/l	99
78) n-propylbenzene	11.305	91	608332	74.974	ug/l	100
79) 2-Chlorotoluene	11.360	91	372392	68.100	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	432096	71.903	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	54398	81.820	ug/l	93
82) 4-Chlorotoluene	11.451	91	432053	71.890	ug/l	100
83) tert-Butylbenzene	11.713	119	432011	73.197	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	442728	73.758	ug/l	100
85) sec-Butylbenzene	11.890	105	543597	74.188	ug/l	100
86) p-Isopropyltoluene	12.006	119	457898	77.393	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	219753	70.449	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	217613	71.190	ug/l	99
89) n-Butylbenzene	12.329	91	416602	80.996	ug/l	100
90) Hexachloroethane	12.536	117	82528	74.978	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	216834	71.367	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	43673	74.512	ug/l	98
93) 1,2,4-Trichlorobenzene	13.585	180	137430	82.605	ug/l	98
94) Hexachlorobutadiene	13.725	225	55494	71.314	ug/l	98
95) Naphthalene	13.774	128	489778	80.842	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	139525	79.981	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC100

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046043.D
Acq On : 05 May 2025 12:21
Operator : JC/MD
Sample : VSTDIC100
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDIC100

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/06/2025
Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:10:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC150

Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	82963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	144975	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	128557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	60345	50.000	ug/l	# 0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.952	65	231952	93.983	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 187.960%#	
35) Dibromofluoromethane	5.385	113	160168	98.562	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 197.120%#	
50) Toluene-d8	8.647	98	554390	101.252	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 202.500%#	
62) 4-Bromofluorobenzene	11.079	95	217536	109.174	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 218.340%#	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	217793	118.929	ug/l	98
3) Chloromethane	1.307	50	196843	109.161	ug/l	99
4) Vinyl Chloride	1.374	62	187924	118.953	ug/l	100
5) Bromomethane	1.593	94	83008	103.625	ug/l	95
6) Chloroethane	1.660	64	78806	96.822	ug/l	98
7) Trichlorofluoromethane	1.868	101	245108	101.851	ug/l	97
8) Diethyl Ether	2.136	74	88441	111.791	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.313	101	161404	112.243	ug/l	99
10) Methyl Iodide	2.441	142	189843	112.441	ug/l	100
11) Tert butyl alcohol	3.002	59	181238	616.119	ug/l	100
12) 1,1-Dichloroethene	2.307	96	155451	111.039	ug/l	99
13) Acrolein	2.239	56	203257	594.872	ug/l	98
14) Allyl chloride	2.654	41	297776	112.561	ug/l	98
15) Acrylonitrile	3.069	53	484854	550.973	ug/l	98
16) Acetone	2.392	43	460820	545.754	ug/l	98
17) Carbon Disulfide	2.502	76	397353	122.499	ug/l	100
18) Methyl Acetate	2.709	43	217885	108.099	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	557210	114.112	ug/l	100
20) Methylene Chloride	2.782	84	172035	100.579	ug/l	96
21) trans-1,2-Dichloroethene	3.081	96	154694	108.241	ug/l	96
22) Diisopropyl ether	3.764	45	577610	116.858	ug/l	90
23) Vinyl Acetate	3.721	43	2655751	608.005	ug/l	100
24) 1,1-Dichloroethane	3.605	63	320083	110.182	ug/l	99
25) 2-Butanone	4.562	43	708463	581.292	ug/l	98
26) 2,2-Dichloropropane	4.471	77	258698	117.548	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	187835	108.994	ug/l	99
28) Bromochloromethane	4.898	49	146923	95.108	ug/l	100
29) Tetrahydrofuran	5.007	42	450456	569.081	ug/l	99
30) Chloroform	5.087	83	323659	107.208	ug/l	95
31) Cyclohexane	5.458	56	286312	116.934	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	295702	113.572	ug/l	98
36) 1,1-Dichloropropene	5.684	75	219731	116.462	ug/l	99
37) Ethyl Acetate	4.715	43	274312	115.889	ug/l	99
38) Carbon Tetrachloride	5.666	117	250916	115.760	ug/l	99
39) Methylcyclohexane	7.373	83	286311	121.107	ug/l	97
40) Benzene	6.031	78	642178	109.813	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046044.D
 Acq On : 05 May 2025 12:45
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :

MSVOA_X

ClientSampleId :

VSTDICC150

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Quant Title : SW846 8260

QLast Update : Tue May 06 06:04:56 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	151424	115.139	ug/l	98
42) 1,2-Dichloroethane	6.080	62	271992	112.492	ug/l	99
43) Isopropyl Acetate	6.342	43	447871	123.780	ug/l	100
44) Trichloroethene	7.117	130	157641	114.032	ug/l	96
45) 1,2-Dichloropropane	7.428	63	164505	113.217	ug/l	99
46) Dibromomethane	7.574	93	125846	109.869	ug/l	100
47) Bromodichloromethane	7.818	83	258326	116.365	ug/l	98
48) Methyl methacrylate	7.690	41	231040	124.356	ug/l	99
49) 1,4-Dioxane	7.665	88	82800	2241.695	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	1371275	577.831	ug/l	100
52) Toluene	8.714	92	393323	113.430	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	257233	138.001	ug/l	99
54) cis-1,3-Dichloropropene	8.366	75	271054	124.658	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	154932	110.551	ug/l	97
56) Ethyl methacrylate	9.116	69	277921	127.968	ug/l	99
57) 1,3-Dichloropropane	9.305	76	272814	109.813	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.244	63	680608	697.344	ug/l	100
59) 2-Hexanone	9.433	43	1028597	571.094	ug/l	99
60) Dibromochloromethane	9.519	129	187497	122.915	ug/l	98
61) 1,2-Dibromoethane	9.604	107	165634	115.087	ug/l	97
64) Tetrachloroethene	9.269	164	132488	100.918	ug/l	96
65) Chlorobenzene	10.080	112	429656	108.615	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	152173	115.333	ug/l	99
67) Ethyl Benzene	10.189	91	785121	116.421	ug/l	99
68) m/p-Xylenes	10.299	106	570913	234.349	ug/l	100
69) o-Xylene	10.640	106	279906	113.990	ug/l	99
70) Styrene	10.653	104	474254	120.694	ug/l	99
71) Bromoform	10.799	173	125986	127.128	ug/l #	99
73) Isopropylbenzene	10.957	105	752337	114.312	ug/l	99
74) N-amyl acetate	10.842	43	396877	120.944	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	243532	104.018	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	213744m	84.960	ug/l	
77) Bromobenzene	11.195	156	167934	110.297	ug/l	99
78) n-propylbenzene	11.305	91	881218	119.448	ug/l	100
79) 2-Chlorotoluene	11.360	91	534560	107.514	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	616477	112.825	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	81372	134.609	ug/l	93
82) 4-Chlorotoluene	11.451	91	611736	111.948	ug/l	100
83) tert-Butylbenzene	11.713	119	617473	115.065	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	623456	114.236	ug/l	99
85) sec-Butylbenzene	11.890	105	786293	118.023	ug/l	99
86) p-Isopropyltoluene	12.006	119	651531	121.113	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	313100	110.394	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	311752	112.167	ug/l	100
89) n-Butylbenzene	12.329	91	605783	129.533	ug/l	100
90) Hexachloroethane	12.536	117	123188	123.091	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	308039	111.506	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	64438	120.915	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	203255	134.366	ug/l	98
94) Hexachlorobutadiene	13.719	225	80558	113.857	ug/l	99
95) Naphthalene	13.774	128	721236	130.929	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	200379	126.331	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046044.D
Acq On : 05 May 2025 12:45
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046044.D
Acq On : 05 May 2025 12:45
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC150

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:11:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	96964	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	168484	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	147167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67939	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	9067	3.143	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	6.280%#		
35) Dibromofluoromethane	5.373	113	5969	3.161	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	6.320%#		
50) Toluene-d8	8.647	98	20566	3.232	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	6.460%#		
62) 4-Bromofluorobenzene	11.079	95	7822	3.378	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	6.760%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	6195	2.894	ug/l	95
3) Chloromethane	1.307	50	6587	3.125	ug/l	99
4) Vinyl Chloride	1.374	62	5999	3.249	ug/l	96
5) Bromomethane	1.593	94	2962	3.164	ug/l	88
6) Chloroethane	1.666	64	3567	3.750	ug/l	94
7) Trichlorofluoromethane	1.874	101	9599	3.413	ug/l	88
8) Diethyl Ether	2.130	74	3020	3.266	ug/l	95
9) 1,1,2-Trichlorotrifluo...	2.325	101	5911	3.517	ug/l	99
10) Methyl Iodide	2.441	142	5900	2.990	ug/l	99
11) Tert butyl alcohol	2.965	59	5539	16.111	ug/l	98
12) 1,1-Dichloroethene	2.306	96	5496	3.359	ug/l	97
13) Acrolein	2.233	56	5673	14.206	ug/l	95
14) Allyl chloride	2.654	41	10262	3.319	ug/l	98
15) Acrylonitrile	3.062	53	16746	16.282	ug/l	97
16) Acetone	2.380	43	19773	20.036	ug/l	98
17) Carbon Disulfide	2.501	76	11068	2.919	ug/l	96
18) Methyl Acetate	2.703	43	7909	3.357	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	18497	3.241	ug/l	97
20) Methylene Chloride	2.782	84	6680	3.341	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	5402	3.234	ug/l	100
22) Diisopropyl ether	3.751	45	18657	3.230	ug/l #	65
23) Vinyl Acetate	3.715	43	82328	16.127	ug/l	99
24) 1,1-Dichloroethane	3.605	63	11193	3.297	ug/l	99
25) 2-Butanone	4.562	43	26125	18.340	ug/l	98
26) 2,2-Dichloropropane	4.465	77	8245	3.205	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	6222	3.089	ug/l	95
28) Bromochloromethane	4.891	49	5360	2.969	ug/l #	100
29) Tetrahydrofuran	5.001	42	15431	16.680	ug/l	98
30) Chloroform	5.080	83	11624	3.294	ug/l	89
31) Cyclohexane	5.458	56	10269	3.588	ug/l	96
32) 1,1,1-Trichloroethane	5.367	97	9827	3.229	ug/l	97
36) 1,1-Dichloropropene	5.678	75	7787	3.551	ug/l	98
37) Ethyl Acetate	4.721	43	9806m	3.565	ug/l	
38) Carbon Tetrachloride	5.659	117	8505	3.376	ug/l	97
39) Methylcyclohexane	7.373	83	9882	3.597	ug/l	93
40) Benzene	6.031	78	22528	3.315	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046046.D
 Acq On : 05 May 2025 16:04
 Operator : JC/MD
 Sample : VSTDICC005
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC005

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	4851	3.174	ug/l #	65
42) 1,2-Dichloroethane	6.080	62	10011	3.563	ug/l	100
43) Isopropyl Acetate	6.348	43	13914	3.309	ug/l	96
44) Trichloroethene	7.123	130	5313	3.307	ug/l	94
45) 1,2-Dichloropropane	7.427	63	5451	3.228	ug/l #	89
46) Dibromomethane	7.580	93	4417	3.318	ug/l	98
47) Bromodichloromethane	7.818	83	8394	3.254	ug/l	98
48) Methyl methacrylate	7.696	41	7171	3.321	ug/l	97
49) 1,4-Dioxane	7.659	88	2907	67.721	ug/l	96
51) 4-Methyl-2-Pentanone	8.567	43	46794	16.967	ug/l	98
52) Toluene	8.714	92	14126	3.505	ug/l	97
53) t-1,3-Dichloropropene	8.976	75	6834	3.155	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	7908	3.129	ug/l	91
55) 1,1,2-Trichloroethane	9.153	97	5685	3.490	ug/l	98
56) Ethyl methacrylate	9.116	69	8555	3.390	ug/l	98
57) 1,3-Dichloropropane	9.305	76	10127	3.508	ug/l	94
58) 2-Chloroethyl Vinyl ether	8.238	63	20809	18.346	ug/l	99
59) 2-Hexanone	9.427	43	34853	16.651	ug/l	97
60) Dibromochloromethane	9.518	129	5500	3.102	ug/l	98
61) 1,2-Dibromoethane	9.604	107	5614	3.356	ug/l	99
64) Tetrachloroethene	9.269	164	4752	3.162	ug/l	96
65) Chlorobenzene	10.079	112	15391	3.399	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	5024	3.326	ug/l	99
67) Ethyl Benzene	10.195	91	26732	3.463	ug/l	99
68) m/p-Xylenes	10.299	106	19957	7.156	ug/l	98
69) o-Xylene	10.640	106	9398	3.343	ug/l	95
70) Styrene	10.652	104	14900	3.312	ug/l	99
71) Bromoform	10.799	173	3470	3.059	ug/l #	99
73) Isopropylbenzene	10.957	105	24201	3.266	ug/l	98
74) N-amyl acetate	10.841	43	11224	3.038	ug/l	96
75) 1,1,2,2-Tetrachloroethane	11.207	83	9170	3.479	ug/l	97
76) 1,2,3-Trichloropropane	11.238	75	7928m	2.799	ug/l	
77) Bromobenzene	11.195	156	5857	3.417	ug/l	96
78) n-propylbenzene	11.299	91	28436	3.424	ug/l	100
79) 2-Chlorotoluene	11.360	91	18672	3.336	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	20742	3.372	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	1828	2.686	ug/l	99
82) 4-Chlorotoluene	11.451	91	19966	3.245	ug/l	98
83) tert-Butylbenzene	11.713	119	21049	3.484	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	20613	3.355	ug/l	97
85) sec-Butylbenzene	11.890	105	25590	3.412	ug/l	99
86) p-Isopropyltoluene	12.006	119	20955	3.460	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	10585	3.315	ug/l	97
88) 1,4-Dichlorobenzene	12.036	146	10908	3.486	ug/l	90
89) n-Butylbenzene	12.329	91	18004	3.419	ug/l	94
90) Hexachloroethane	12.536	117	3553	3.153	ug/l	94
91) 1,2-Dichlorobenzene	12.335	146	10712	3.444	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	1687	2.812	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	5720	3.359	ug/l	96
94) Hexachlorobutadiene	13.719	225	2816	3.535	ug/l	96
95) Naphthalene	13.774	128	19902	3.209	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	5745	3.217	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On : 05 May 2025 16:04
Operator : JC/MD
Sample : VSTDICC005
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


```
Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046046.D
Acq On    : 05 May 2025  16:04
Operator  : JC/MD
Sample    : VSTDIC005
Misc      : 5.0mL/MSVOA_X/WATER
ALS Vial  : 10   Sample Multiplier: 1
```

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC005

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:12:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92040	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167022	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	140108	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59123	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	0.000	65	0d	0.000	ug/l	
Spiked Amount	50.000	Range	74 - 125	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0d	0.000	ug/l	
Spiked Amount	50.000	Range	75 - 124	Recovery	=	0.000%#
50) Toluene-d8	0.000	98	0d	0.000	ug/l	
Spiked Amount	50.000	Range	86 - 113	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	0.000	95	0d	0.000	ug/l	
Spiked Amount	50.000	Range	77 - 121	Recovery	=	0.000%#
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1211	0.596	ug/l	98
3) Chloromethane	1.307	50	1278	0.639	ug/l	97
4) Vinyl Chloride	1.374	62	1239	0.707	ug/l #	82
6) Chloroethane	1.678	64	859	0.951	ug/l	91
7) Trichlorofluoromethane	1.880	101	1959	0.734	ug/l	95
8) Diethyl Ether	2.130	74	742	0.845	ug/l	87
9) 1,1,2-Trichlorotrifluo...	2.319	101	1166	0.731	ug/l	96
12) 1,1-Dichloroethene	2.313	96	1093	0.704	ug/l #	87
14) Allyl chloride	2.654	41	1936	0.660	ug/l	89
15) Acrylonitrile	3.075	53	3343	3.424	ug/l	98
16) Acetone	2.380	43	3500	3.736	ug/l	97
17) Carbon Disulfide	2.508	76	2619	0.728	ug/l	98
18) Methyl Acetate	2.709	43	1852	0.828	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	3588	0.662	ug/l	96
20) Methylene Chloride	2.782	84	1571	0.828	ug/l #	90
21) trans-1,2-Dichloroethene	3.081	96	1111	0.701	ug/l #	71
22) Diisopropyl ether	3.757	45	3857	0.703	ug/l #	26
23) Vinyl Acetate	3.727	43	15283	3.154	ug/l	96
24) 1,1-Dichloroethane	3.605	63	2054	0.637	ug/l	97
25) 2-Butanone	4.593	43	4558	3.371	ug/l	92
26) 2,2-Dichloropropane	4.458	77	1777	0.728	ug/l	84
27) cis-1,2-Dichloroethene	4.483	96	1324	0.693	ug/l	91
28) Bromochloromethane	4.910	49	1061m	0.619	ug/l	
29) Tetrahydrofuran	5.019	42	2931	3.338	ug/l	99
30) Chloroform	5.093	83	2328	0.695	ug/l	98
32) 1,1,1-Trichloroethane	5.379	97	1869	0.647	ug/l #	51
36) 1,1-Dichloropropene	5.684	75	1648	0.758	ug/l #	83
37) Ethyl Acetate	4.733	43	1956m	0.717	ug/l	
38) Carbon Tetrachloride	5.678	117	1807	0.724	ug/l #	78
39) Methylcyclohexane	7.373	83	2096	0.770	ug/l	97
40) Benzene	6.044	78	4502	0.668	ug/l #	89
41) Methacrylonitrile	4.946	41	778	0.513	ug/l #	56
42) 1,2-Dichloroethane	6.092	62	1934	0.694	ug/l	81
43) Isopropyl Acetate	6.354	43	2553	0.612	ug/l #	81
44) Trichloroethene	7.135	130	1083	0.680	ug/l	85
45) 1,2-Dichloropropane	7.440	63	1059	0.633	ug/l	89

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046047.D
 Acq On : 05 May 2025 16:27
 Operator : JC/MD
 Sample : VSTDICC001
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICC001

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:13:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) Dibromomethane	7.586	93	880	0.667	ug/l	97
47) Bromodichloromethane	7.818	83	1620	0.633	ug/l #	91
48) Methyl methacrylate	7.714	41	1235m	0.577	ug/l	
49) 1,4-Dioxane	7.677	88	498	11.703	ug/l #	66
51) 4-Methyl-2-Pentanone	8.580	43	9378	3.430	ug/l	97
52) Toluene	8.720	92	2684	0.672	ug/l	96
53) t-1,3-Dichloropropene	8.988	75	1238	0.576	ug/l #	79
54) cis-1,3-Dichloropropene	8.379	75	1414	0.564	ug/l #	87
55) 1,1,2-Trichloroethane	9.153	97	1029	0.637	ug/l	92
56) Ethyl methacrylate	9.128	69	1260	0.504	ug/l #	76
57) 1,3-Dichloropropane	9.311	76	2036	0.711	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.244	63	3846	3.420	ug/l	94
59) 2-Hexanone	9.439	43	6424	3.096	ug/l	88
60) Dibromochloromethane	9.525	129	1023	0.582	ug/l	93
61) 1,2-Dibromoethane	9.610	107	1077	0.650	ug/l	98
64) Tetrachloroethene	9.275	164	972	0.679	ug/l	95
65) Chlorobenzene	10.079	112	3170	0.735	ug/l	98
66) 1,1,1,2-Tetrachloroethane	10.159	131	1035	0.720	ug/l #	67
67) Ethyl Benzene	10.195	91	5052	0.687	ug/l	94
68) m/p-Xylenes	10.299	106	3630	1.367	ug/l	97
69) o-Xylene	10.640	106	1799	0.672	ug/l	94
70) Styrene	10.659	104	2664	0.622	ug/l	98
71) Bromoform	10.799	173	657	0.608	ug/l #	100
73) Isopropylbenzene	10.963	105	4480	0.695	ug/l	96
74) N-amyl acetate	10.854	43	2028	0.631	ug/l #	91
75) 1,1,2,2-Tetrachloroethane	11.213	83	1835	0.800	ug/l	94
76) 1,2,3-Trichloropropane	11.244	75	1661m	0.674	ug/l	
77) Bromobenzene	11.201	156	1095	0.734	ug/l	96
78) n-propylbenzene	11.305	91	5052	0.699	ug/l	97
79) 2-Chlorotoluene	11.360	91	3765	0.773	ug/l	94
80) 1,3,5-Trimethylbenzene	11.451	105	3590	0.671	ug/l	96
82) 4-Chlorotoluene	11.457	91	3815	0.713	ug/l	97
83) tert-Butylbenzene	11.713	119	3951	0.751	ug/l	97
84) 1,2,4-Trimethylbenzene	11.756	105	3725	0.697	ug/l	97
85) sec-Butylbenzene	11.890	105	4385	0.672	ug/l	100
86) p-Isopropyltoluene	12.006	119	3577	0.679	ug/l	94
87) 1,3-Dichlorobenzene	11.969	146	1914	0.689	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	2149m	0.789	ug/l	
89) n-Butylbenzene	12.335	91	2889	0.631	ug/l	92
90) Hexachloroethane	12.536	117	604	0.616	ug/l	99
91) 1,2-Dichlorobenzene	12.341	146	2022	0.747	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.945	75	306	0.586	ug/l	83
93) 1,2,4-Trichlorobenzene	13.591	180	1019	0.688	ug/l	94
94) Hexachlorobutadiene	13.725	225	465	0.671	ug/l	89
95) Naphthalene	13.780	128	4138	0.767	ug/l	97
96) 1,2,3-Trichlorobenzene	13.963	180	1113	0.716	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Instrument :
MSVOA_X
ClientSampleId :
VSTDICC001

Manual Integrations APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	92588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	162651	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	145078	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	68229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83733	48.509	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.020%		
35) Dibromofluoromethane	5.379	113	58192	49.683	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	99.360%		
50) Toluene-d8	8.647	98	198575	48.984	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.960%		
62) 4-Bromofluorobenzene	11.079	95	75568	48.596	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	97.200%		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	78783	55.594	ug/l	99
3) Chloromethane	1.307	50	72275	52.592	ug/l	100
4) Vinyl Chloride	1.374	62	65994	51.598	ug/l	98
5) Bromomethane	1.593	94	30033	50.627	ug/l	97
6) Chloroethane	1.666	64	35690	52.271	ug/l	94
7) Trichlorofluoromethane	1.874	101	99352	52.560	ug/l	99
8) Diethyl Ether	2.130	74	31875	49.536	ug/l	97
9) 1,1,2-Trichlorotrifluo...	2.319	101	59451	50.822	ug/l	99
10) Methyl Iodide	2.447	142	73940	53.419	ug/l	99
11) Tert butyl alcohol	2.977	59	61065	252.025	ug/l	99
12) 1,1-Dichloroethene	2.313	96	56097	51.097	ug/l	98
13) Acrolein	2.233	56	64157	232.505	ug/l	99
14) Allyl chloride	2.660	41	108757	51.833	ug/l	100
15) Acrylonitrile	3.063	53	181808	262.412	ug/l	99
16) Acetone	2.380	43	169923	245.518	ug/l	98
17) Carbon Disulfide	2.502	76	132695	50.981	ug/l	99
18) Methyl Acetate	2.703	43	80251	49.968	ug/l	97
19) Methyl tert-butyl Ether	3.111	73	195442	50.777	ug/l	99
20) Methylene Chloride	2.782	84	63369	47.779	ug/l	99
21) trans-1,2-Dichloroethene	3.087	96	55770	50.513	ug/l	99
22) Diisopropyl ether	3.757	45	209946	51.800	ug/l	97
23) Vinyl Acetate	3.715	43	946746	265.585	ug/l	100
24) 1,1-Dichloroethane	3.605	63	114733	50.825	ug/l	98
25) 2-Butanone	4.556	43	258339	257.110	ug/l	97
26) 2,2-Dichloropropane	4.471	77	88229	49.934	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	67727	50.956	ug/l	100
28) Bromochloromethane	4.891	49	55390	50.975	ug/l	98
29) Tetrahydrofuran	5.001	42	163675	259.958	ug/l	99
30) Chloroform	5.093	83	120782	51.332	ug/l	97
31) Cyclohexane	5.458	56	103158	50.147	ug/l	96
32) 1,1,1-Trichloroethane	5.373	97	104243	51.108	ug/l	100
36) 1,1-Dichloropropene	5.690	75	77931	49.521	ug/l	99
37) Ethyl Acetate	4.715	43	98604	50.714	ug/l	100
38) Carbon Tetrachloride	5.666	117	88100	49.825	ug/l	97
39) Methylcyclohexane	7.373	83	100808	49.757	ug/l	98
40) Benzene	6.031	78	233977	50.759	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	56017	55.076	ug/l	100
42) 1,2-Dichloroethane	6.080	62	101185	50.861	ug/l	98
43) Isopropyl Acetate	6.336	43	154818	52.195	ug/l	99
44) Trichloroethene	7.123	130	55491	50.018	ug/l	100
45) 1,2-Dichloropropane	7.428	63	59582	51.983	ug/l	98
46) Dibromomethane	7.580	93	45181	49.978	ug/l	99
47) Bromodichloromethane	7.818	83	92388	51.889	ug/l	100
48) Methyl methacrylate	7.690	41	82144	54.226	ug/l	98
49) 1,4-Dioxane	7.659	88	30477	1059.599	ug/l	97
51) 4-Methyl-2-Pentanone	8.574	43	510206	259.133	ug/l	100
52) Toluene	8.714	92	144574	51.150	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	83751	52.920	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	92517	52.892	ug/l	98
55) 1,1,2-Trichloroethane	9.147	97	56713	50.887	ug/l	96
56) Ethyl methacrylate	9.116	69	95772	53.918	ug/l	99
57) 1,3-Dichloropropane	9.305	76	99663	49.793	ug/l	97
58) 2-Chloroethyl Vinyl ether	8.238	63	240681	265.780	ug/l	99
59) 2-Hexanone	9.427	43	390701	268.218	ug/l	100
60) Dibromochloromethane	9.519	129	63738	52.075	ug/l	99
61) 1,2-Dibromoethane	9.610	107	59134	51.050	ug/l	98
64) Tetrachloroethene	9.269	164	48034	46.795	ug/l	98
65) Chlorobenzene	10.073	112	154640	48.699	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	53735	49.557	ug/l	100
67) Ethyl Benzene	10.189	91	281239	50.245	ug/l	100
68) m/p-Xylenes	10.299	106	206813	101.022	ug/l	99
69) o-Xylene	10.640	106	101469	50.841	ug/l	99
70) Styrene	10.653	104	169703	51.907	ug/l	99
71) Bromoform	10.799	173	40364	49.583	ug/l #	97
73) Isopropylbenzene	10.957	105	269484	50.733	ug/l	100
74) N-amyl acetate	10.842	43	137559	52.407	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	89027	47.827	ug/l	98
76) 1,2,3-Trichloropropane	11.238	75	79763m	48.568	ug/l	
77) Bromobenzene	11.195	156	59871	48.549	ug/l	98
78) n-propylbenzene	11.299	91	312285	50.562	ug/l	100
79) 2-Chlorotoluene	11.360	91	195630	49.107	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	228538	51.500	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	24857	49.276	ug/l	98
82) 4-Chlorotoluene	11.451	91	223811	50.661	ug/l	100
83) tert-Butylbenzene	11.713	119	225783	50.511	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	231371	51.485	ug/l	100
85) sec-Butylbenzene	11.890	105	280578	51.122	ug/l	99
86) p-Isopropyltoluene	12.006	119	231942	51.198	ug/l	100
87) 1,3-Dichlorobenzene	11.963	146	113146	50.272	ug/l	99
88) 1,4-Dichlorobenzene	12.037	146	111173	48.368	ug/l	99
89) n-Butylbenzene	12.329	91	205858	51.804	ug/l	100
90) Hexachloroethane	12.536	117	39823	49.895	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	111482	49.361	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21220	51.458	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	64421	49.662	ug/l	100
94) Hexachlorobutadiene	13.719	225	27425	48.408	ug/l	98
95) Naphthalene	13.774	128	239805	50.404	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	64950	48.525	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Manual Integrations APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICSVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.765	0.851	-11.2	94	0.00
3 P	Chloromethane	0.742	0.781	-5.3	96	0.00
4 C	Vinyl Chloride	0.691	0.713	-3.2#	96	0.00
5 T	Bromomethane	0.320	0.324	-1.3	95	0.00
6 T	Chloroethane	0.369	0.385	-4.3	97	0.00
7 T	Trichlorofluoromethane	1.021	1.073	-5.1	96	0.00
8 T	Diethyl Ether	0.347	0.344	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.642	-1.6	96	0.00
10 T	Methyl Iodide	0.747	0.799	-7.0	94	0.00
11 T	Tert butyl alcohol	0.131	0.132	-0.8	98	0.00
12 CM	1,1-Dichloroethene	0.593	0.606	-2.2#	96	0.00
13 T	Acrolein	0.149	0.139	6.7	87	0.00
14 T	Allyl chloride	1.133	1.175	-3.7	95	0.00
15 T	Acrylonitrile	0.374	0.393	-5.1	97	0.00
16 T	Acetone	0.374	0.367	1.9	97	0.00
17 T	Carbon Disulfide	1.406	1.433	-1.9	94	0.00
18 T	Methyl Acetate	0.867	0.867	0.0	97	0.00
19 T	Methyl tert-butyl Ether	2.079	2.111	-1.5	93	0.00
20 T	Methylene Chloride	0.716	0.684	4.5	95	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.602	-1.0	94	0.00
22 T	Diisopropyl ether	2.189	2.268	-3.6	95	0.00
23 T	Vinyl Acetate	1.925	2.045	-6.2	95	0.00
24 P	1,1-Dichloroethane	1.219	1.239	-1.6	94	0.00
25 T	2-Butanone	0.543	0.558	-2.8	96	0.00
26 T	2,2-Dichloropropane	0.954	0.953	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.731	-1.8	95	0.00
28 T	Bromochloromethane	0.587	0.598	-1.9	99	0.00
29 T	Tetrahydrofuran	0.340	0.354	-4.1	96	0.00
30 C	Chloroform	1.271	1.305	-2.7#	96	0.00
31 T	Cyclohexane	1.111	1.114	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	1.101	1.126	-2.3	95	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.904	3.0	95	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.360	0.358	0.6	98	0.00
36 T	1,1-Dichloropropene	0.484	0.479	1.0	94	0.00
37 T	Ethyl Acetate	0.598	0.606	-1.3	96	0.00
38 T	Carbon Tetrachloride	0.544	0.542	0.4	94	0.00
39 T	Methylcyclohexane	0.623	0.620	0.5	94	0.00
40 TM	Benzene	1.417	1.439	-1.6	95	0.00
41 T	Methacrylonitrile	0.313	0.344	-9.9	96	0.00
42 TM	1,2-Dichloroethane	0.612	0.622	-1.6	96	0.00
43 T	Isopropyl Acetate	0.912	0.952	-4.4	96	0.00
44 TM	Trichloroethene	0.341	0.341	0.0	93	0.00
45 C	1,2-Dichloropropane	0.352	0.366	-4.0#	96	0.00
46 T	Dibromomethane	0.278	0.278	0.0	94	0.00
47 T	Bromodichloromethane	0.547	0.568	-3.8	95	0.00
48 T	Methyl methacrylate	0.466	0.505	-8.4	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	100	0.00
50 S	Toluene-d8	1.246	1.221	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.627	-3.6	96	0.00
52 CM	Toluene	0.869	0.889	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	0.487	0.515	-5.7	94	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.569	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	0.343	0.349	-1.7	95	0.00
56 T	Ethyl methacrylate	0.546	0.589	-7.9	96	0.00
57 T	1,3-Dichloropropane	0.615	0.613	0.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.296	-6.5	93	0.00
59 T	2-Hexanone	0.448	0.480	-7.1	98	0.00
60 T	Dibromochloromethane	0.376	0.392	-4.3	95	0.00
61 T	1,2-Dibromoethane	0.356	0.364	-2.2	94	0.00
62 S	4-Bromofluorobenzene	0.478	0.465	2.7	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.354	0.331	6.5	88	0.00
65 PM	Chlorobenzene	1.094	1.066	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.370	1.1	94	0.00
67 C	Ethyl Benzene	1.929	1.939	-0.5#	95	0.00
68 T	m/p-Xylenes	0.706	0.713	-1.0	95	0.00
69 T	o-Xylene	0.688	0.699	-1.6	95	0.00
70 T	Styrene	1.127	1.170	-3.8	95	0.00
71 P	Bromoform	0.281	0.278	1.1	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.893	3.950	-1.5	96	0.00
74 T	N-amyl acetate	1.924	2.016	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.305	4.3	98	0.00
76 T	1,2,3-Trichloropropane	1.204	1.169	2.9	99	0.00
77 T	Bromobenzene	0.904	0.878	2.9	95	0.00
78 T	n-propylbenzene	4.526	4.577	-1.1	95	0.00
79 T	2-Chlorotoluene	2.919	2.867	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.350	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.364	1.6	95	0.00
82 T	4-Chlorotoluene	3.238	3.280	-1.3	96	0.00
83 T	tert-Butylbenzene	3.276	3.309	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.391	-3.0	97	0.00
85 T	sec-Butylbenzene	4.022	4.112	-2.2	96	0.00
86 T	p-Isopropyltoluene	3.320	3.399	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	1.649	1.658	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	1.684	1.629	3.3	97	0.00
89 T	n-Butylbenzene	2.912	3.017	-3.6	96	0.00
90 T	Hexachloroethane	0.585	0.584	0.2	94	0.00
91 T	1,2-Dichlorobenzene	1.655	1.634	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.311	-3.0	97	0.00
93 T	1,2,4-Trichlorobenzene	0.951	0.944	0.7	97	0.00
94 T	Hexachlorobutadiene	0.415	0.402	3.1	95	0.00
95 T	Naphthalene	3.487	3.515	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	0.952	3.0	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	55.594	-11.2	94	0.00
3 P	Chloromethane	50.000	52.592	-5.2	96	0.00
4 C	Vinyl Chloride	50.000	51.598	-3.2#	96	0.00
5 T	Bromomethane	50.000	50.627	-1.3	95	0.00
6 T	Chloroethane	50.000	52.271	-4.5	97	0.00
7 T	Trichlorofluoromethane	50.000	52.560	-5.1	96	0.00
8 T	Diethyl Ether	50.000	49.536	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.822	-1.6	96	0.00
10 T	Methyl Iodide	50.000	53.419	-6.8	94	0.00
11 T	Tert butyl alcohol	250.000	252.025	-0.8	98	0.00
12 CM	1,1-Dichloroethene	50.000	51.097	-2.2#	96	0.00
13 T	Acrolein	250.000	232.505	7.0	87	0.00
14 T	Allyl chloride	50.000	51.833	-3.7	95	0.00
15 T	Acrylonitrile	250.000	262.412	-5.0	97	0.00
16 T	Acetone	250.000	245.518	1.8	97	0.00
17 T	Carbon Disulfide	50.000	50.981	-2.0	94	0.00
18 T	Methyl Acetate	50.000	49.968	0.1	97	0.00
19 T	Methyl tert-butyl Ether	50.000	50.777	-1.6	93	0.00
20 T	Methylene Chloride	50.000	47.779	4.4	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.513	-1.0	94	0.00
22 T	Diisopropyl ether	50.000	51.800	-3.6	95	0.00
23 T	Vinyl Acetate	250.000	265.585	-6.2	95	0.00
24 P	1,1-Dichloroethane	50.000	50.825	-1.7	94	0.00
25 T	2-Butanone	250.000	257.110	-2.8	96	0.00
26 T	2,2-Dichloropropane	50.000	49.934	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.956	-1.9	95	0.00
28 T	Bromochloromethane	50.000	50.975	-2.0	99	0.00
29 T	Tetrahydrofuran	250.000	259.958	-4.0	96	0.00
30 C	Chloroform	50.000	51.332	-2.7#	96	0.00
31 T	Cyclohexane	50.000	50.147	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	50.000	51.108	-2.2	95	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.509	3.0	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	49.683	0.6	98	0.00
36 T	1,1-Dichloropropene	50.000	49.521	1.0	94	0.00
37 T	Ethyl Acetate	50.000	50.714	-1.4	96	0.00
38 T	Carbon Tetrachloride	50.000	49.825	0.3	94	0.00
39 T	Methylcyclohexane	50.000	49.757	0.5	94	0.00
40 TM	Benzene	50.000	50.759	-1.5	95	0.00
41 T	Methacrylonitrile	50.000	55.076	-10.2	96	0.00
42 TM	1,2-Dichloroethane	50.000	50.861	-1.7	96	0.00
43 T	Isopropyl Acetate	50.000	52.195	-4.4	96	0.00
44 TM	Trichloroethene	50.000	50.018	-0.0	93	0.00
45 C	1,2-Dichloropropane	50.000	51.983	-4.0#	96	0.00
46 T	Dibromomethane	50.000	49.978	0.0	94	0.00
47 T	Bromodichloromethane	50.000	51.889	-3.8	95	0.00
48 T	Methyl methacrylate	50.000	54.226	-8.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1059.599	-6.0	100	0.00
50 S	Toluene-d8	50.000	48.984	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	259.133	-3.7	96	0.00
52 CM	Toluene	50.000	51.150	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	52.920	-5.8	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.892	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	50.000	50.887	-1.8	95	0.00
56 T	Ethyl methacrylate	50.000	53.918	-7.8	96	0.00
57 T	1,3-Dichloropropane	50.000	49.793	0.4	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	265.780	-6.3	93	0.00
59 T	2-Hexanone	250.000	268.218	-7.3	98	0.00
60 T	Dibromochloromethane	50.000	52.075	-4.2	95	0.00
61 T	1,2-Dibromoethane	50.000	51.050	-2.1	94	0.00
62 S	4-Bromofluorobenzene	50.000	48.596	2.8	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	46.795	6.4	88	0.00
65 PM	Chlorobenzene	50.000	48.699	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.557	0.9	94	0.00
67 C	Ethyl Benzene	50.000	50.245	-0.5#	95	0.00
68 T	m/p-Xylenes	100.000	101.022	-1.0	95	0.00
69 T	o-Xylene	50.000	50.841	-1.7	95	0.00
70 T	Styrene	50.000	51.907	-3.8	95	0.00
71 P	Bromoform	50.000	49.583	0.8	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	50.733	-1.5	96	0.00
74 T	N-ethyl acetate	50.000	52.407	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.827	4.3	98	0.00
76 T	1,2,3-Trichloropropane	50.000	48.568	2.9	99	0.00
77 T	Bromobenzene	50.000	48.549	2.9	95	0.00
78 T	n-propylbenzene	50.000	50.562	-1.1	95	0.00
79 T	2-Chlorotoluene	50.000	49.107	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.500	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.276	1.4	95	0.00
82 T	4-Chlorotoluene	50.000	50.661	-1.3	96	0.00
83 T	tert-Butylbenzene	50.000	50.511	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.485	-3.0	97	0.00
85 T	sec-Butylbenzene	50.000	51.122	-2.2	96	0.00
86 T	p-Isopropyltoluene	50.000	51.198	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	50.000	50.272	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	50.000	48.368	3.3	97	0.00
89 T	n-Butylbenzene	50.000	51.804	-3.6	96	0.00
90 T	Hexachloroethane	50.000	49.895	0.2	94	0.00
91 T	1,2-Dichlorobenzene	50.000	49.361	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.458	-2.9	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.662	0.7	97	0.00
94 T	Hexachlorobutadiene	50.000	48.408	3.2	95	0.00
95 T	Naphthalene	50.000	50.404	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.525	3.0	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG No.: Q2073

Instrument ID: MSVOA_X Calibration Date/Time: 05/19/2025 10:05

Lab File ID: VX046255.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.765	0.801		4.71	20
Chloromethane	0.742	0.743	0.1	0.14	20
Vinyl Chloride	0.691	0.673		-2.61	20
Bromomethane	0.320	0.297		-7.19	20
Chloroethane	0.369	0.377		2.17	20
Trichlorofluoromethane	1.021	1.050		2.84	20
1,1,2-Trichlorotrifluoroethane	0.632	0.646		2.21	20
1,1-Dichloroethene	0.593	0.573		-3.37	20
Acetone	0.374	0.373		-0.27	20
Carbon Disulfide	1.406	1.339		-4.76	20
Methyl tert-butyl Ether	2.079	2.149		3.37	20
Methyl Acetate	0.867	0.995		14.76	20
Methylene Chloride	0.716	0.670		-6.43	20
trans-1,2-Dichloroethene	0.596	0.585		-1.85	20
1,1-Dichloroethane	1.219	1.233	0.1	1.15	20
Cyclohexane	1.111	1.043		-6.12	20
2-Butanone	0.543	0.550		1.29	20
Carbon Tetrachloride	0.544	0.577		6.07	20
cis-1,2-Dichloroethene	0.718	0.699		-2.65	20
Bromochloromethane	0.587	0.572		-2.56	20
Chloroform	1.271	1.295		1.89	20
1,1,1-Trichloroethane	1.101	1.133		2.91	20
Methylcyclohexane	0.623	0.602		-3.37	20
Benzene	1.417	1.454		2.61	20
1,2-Dichloroethane	0.612	0.624		1.96	20
Trichloroethene	0.341	0.349		2.35	20
1,2-Dichloropropane	0.352	0.380		7.95	20
Bromodichloromethane	0.547	0.604		10.42	20
4-Methyl-2-Pentanone	0.605	0.650		7.44	20
Toluene	0.869	0.887		2.07	20
t-1,3-Dichloropropene	0.487	0.537		10.27	20
cis-1,3-Dichloropropene	0.538	0.589		9.48	20
1,1,2-Trichloroethane	0.343	0.363		5.83	20
2-Hexanone	0.448	0.480		7.14	20
Dibromochloromethane	0.376	0.429		14.1	20
1,2-Dibromoethane	0.356	0.374		5.06	20
Tetrachloroethene	0.354	0.357		0.85	20
Chlorobenzene	1.094	1.113	0.3	1.74	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GENV01

Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG No.: Q2073

Instrument ID: MSVOA_X Calibration Date/Time: 05/19/2025 10:05

Lab File ID: VX046255.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.929	2.022		4.82	20
m/p-Xylenes	0.706	0.743		5.24	20
o-Xylene	0.688	0.721		4.8	20
Styrene	1.127	1.265		12.24	20
Bromoform	0.281	0.322	0.1	14.59	20
Isopropylbenzene	3.893	4.008		2.95	20
1,1,2,2-Tetrachloroethane	1.364	1.325	0.3	-2.86	20
1,3-Dichlorobenzene	1.649	1.676		1.64	20
1,4-Dichlorobenzene	1.684	1.694		0.59	20
1,2-Dichlorobenzene	1.655	1.683		1.69	20
1,2-Dibromo-3-Chloropropane	0.302	0.319		5.63	20
1,2,4-Trichlorobenzene	0.951	0.986		3.68	20
1,2,3-Trichlorobenzene	0.981	0.997		1.63	20
1,2-Dichloroethane-d4	0.932	0.903		-3.11	20
Dibromofluoromethane	0.360	0.383		6.39	20
Toluene-d8	1.246	1.254		0.64	20
4-Bromofluorobenzene	0.478	0.493		3.14	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046255.D
 Acq On : 19 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/21/2025
 Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:05:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	97192	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	163845	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	142712	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	70270	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.946	65	87750	48.428	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.860%
35) Dibromofluoromethane	5.379	113	62751	53.185	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.380%
50) Toluene-d8	8.641	98	205467	50.315	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.620%
62) 4-Bromofluorobenzene	11.079	95	80710	51.525	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.040%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	77853	52.335	ug/l	100
3) Chloromethane	1.301	50	72198	50.048	ug/l	96
4) Vinyl Chloride	1.374	62	65451	48.749	ug/l	98
5) Bromomethane	1.593	94	28871	46.363	ug/l	96
6) Chloroethane	1.666	64	36632	51.109	ug/l	93
7) Trichlorofluoromethane	1.874	101	102085	51.448	ug/l	97
8) Diethyl Ether	2.130	74	32915	48.729	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	62772	51.119	ug/l	99
10) Methyl Iodide	2.441	142	65860	45.327	ug/l	99
11) Tert butyl alcohol	2.977	59	62364	245.194	ug/l	100
12) 1,1-Dichloroethene	2.306	96	55656	48.293	ug/l	98
13) Acrolein	2.233	56	92840	320.514	ug/l	99
14) Allyl chloride	2.654	41	114233	51.864	ug/l	98
15) Acrylonitrile	3.062	53	189093	259.998	ug/l	98
16) Acetone	2.380	43	181372	249.646	ug/l	100
17) Carbon Disulfide	2.502	76	130127	47.627	ug/l	100
18) Methyl Acetate	2.703	43	96692	57.353	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	208892	51.700	ug/l	99
20) Methylene Chloride	2.782	84	65150	46.795	ug/l	97
21) trans-1,2-Dichloroethene	3.081	96	56851	49.053	ug/l	99
22) Diisopropyl ether	3.757	45	220467	51.819	ug/l	95
23) Vinyl Acetate	3.715	43	960660	256.722	ug/l	100
24) 1,1-Dichloroethane	3.599	63	119876	50.587	ug/l	98
25) 2-Butanone	4.550	43	267145	253.279	ug/l	100
26) 2,2-Dichloropropane	4.465	77	97036	52.317	ug/l	100
27) cis-1,2-Dichloroethene	4.477	96	67956	48.707	ug/l	98
28) Bromochloromethane	4.891	49	55621	48.763	ug/l	99
29) Tetrahydrofuran	4.995	42	167256	253.062	ug/l	98
30) Chloroform	5.080	83	125854	50.954	ug/l	96
31) Cyclohexane	5.458	56	101388	46.952	ug/l	99
32) 1,1,1-Trichloroethane	5.373	97	110084	51.415	ug/l	100
36) 1,1-Dichloropropene	5.684	75	78170	49.311	ug/l	97
37) Ethyl Acetate	4.708	43	99787	50.949	ug/l	99
38) Carbon Tetrachloride	5.666	117	94579	53.099	ug/l	98
39) Methylcyclohexane	7.373	83	98620	48.322	ug/l	97
40) Benzene	6.025	78	238301	51.321	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046255.D
 Acq On : 19 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/21/2025
 Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:05:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	57196	55.825	ug/l	98
42) 1,2-Dichloroethane	6.080	62	102281	51.038	ug/l	100
43) Isopropyl Acetate	6.336	43	158764	53.135	ug/l	99
44) Trichloroethene	7.117	130	57119	51.110	ug/l	98
45) 1,2-Dichloropropane	7.421	63	62232	53.899	ug/l	100
46) Dibromomethane	7.574	93	46193	50.726	ug/l	98
47) Bromodichloromethane	7.812	83	98944	55.166	ug/l	98
48) Methyl methacrylate	7.684	41	83134	54.480	ug/l	99
49) 1,4-Dioxane	7.659	88	28914	997.932	ug/l	97
51) 4-Methyl-2-Pentanone	8.568	43	532634	268.553	ug/l	100
52) Toluene	8.714	92	145397	51.066	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	87979	55.186	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	96528	54.783	ug/l	97
55) 1,1,2-Trichloroethane	9.147	97	59460	52.963	ug/l	96
56) Ethyl methacrylate	9.110	69	98149	54.853	ug/l	98
57) 1,3-Dichloropropane	9.305	76	102717	50.945	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	246924	270.687	ug/l	100
59) 2-Hexanone	9.427	43	392889	267.754	ug/l	100
60) Dibromochloromethane	9.519	129	70302	57.019	ug/l	99
61) 1,2-Dibromoethane	9.604	107	61232	52.476	ug/l	97
64) Tetrachloroethene	9.269	164	50976	50.484	ug/l	98
65) Chlorobenzene	10.073	112	158879	50.864	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	57442	53.854	ug/l	100
67) Ethyl Benzene	10.189	91	288514	52.399	ug/l	99
68) m/p-Xylenes	10.299	106	212192	105.368	ug/l	99
69) o-Xylene	10.640	106	102943	52.434	ug/l	96
70) Styrene	10.653	104	180462	56.113	ug/l	99
71) Bromoform	10.799	173	45893	57.309	ug/l #	97
73) Isopropylbenzene	10.957	105	281669	51.487	ug/l	99
74) N-amyl acetate	10.842	43	137605	50.902	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	93098	48.561	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	83305m	49.251	ug/l	
77) Bromobenzene	11.195	156	64940	51.130	ug/l	97
78) n-propylbenzene	11.299	91	332490	52.270	ug/l	100
79) 2-Chlorotoluene	11.360	91	204053	49.734	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	238771	52.243	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	27314	52.574	ug/l	97
82) 4-Chlorotoluene	11.451	91	239800	52.703	ug/l	99
83) tert-Butylbenzene	11.713	119	236790	51.435	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	241446	52.167	ug/l	99
85) sec-Butylbenzene	11.890	105	297240	52.585	ug/l	100
86) p-Isopropyltoluene	12.006	119	248673	53.297	ug/l	99
87) 1,3-Dichlorobenzene	11.963	146	117789	50.815	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	119045	50.288	ug/l	99
89) n-Butylbenzene	12.329	91	224545	54.865	ug/l	100
90) Hexachloroethane	12.536	117	44103	53.653	ug/l	99
91) 1,2-Dichlorobenzene	12.329	146	118255	50.839	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	22414	52.775	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	69267	51.847	ug/l	99
94) Hexachlorobutadiene	13.725	225	30637	52.507	ug/l	97
95) Naphthalene	13.774	128	242964	49.585	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	70028	50.799	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046255.D
Acq On : 19 May 2025 10:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/21/2025
Supervised By :Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:05:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046255.D
Acq On : 19 May 2025 10:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Manual Integrations APPROVED

Reviewed By :John Carlone 05/21/2025
Supervised By :Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:05:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046255.D
 Acq On : 19 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 19 23:05:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.765	0.801	-4.7	93	0.00
3 P	Chloromethane	0.742	0.743	-0.1	96	0.00
4 C	Vinyl Chloride	0.691	0.673	2.6#	95	0.00
5 T	Bromomethane	0.320	0.297	7.2	91	0.00
6 T	Chloroethane	0.369	0.377	-2.2	100	0.00
7 T	Trichlorofluoromethane	1.021	1.050	-2.8	98	0.00
8 T	Diethyl Ether	0.347	0.339	2.3	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.632	0.646	-2.2	101	0.00
10 T	Methyl Iodide	0.747	0.678	9.2	84	0.00
11 T	Tert butyl alcohol	0.131	0.128	2.3	100	0.00
12 CM	1,1-Dichloroethene	0.593	0.573	3.4#	95	0.00
13 T	Acrolein	0.149	0.191	-28.2#	126	0.00
14 T	Allyl chloride	1.133	1.175	-3.7	100	0.00
15 T	Acrylonitrile	0.374	0.389	-4.0	100	0.00
16 T	Acetone	0.374	0.373	0.3	103	0.00
17 T	Carbon Disulfide	1.406	1.339	4.8	92	0.00
18 T	Methyl Acetate	0.867	0.995	-14.8	117	0.00
19 T	Methyl tert-butyl Ether	2.079	2.149	-3.4	100	0.00
20 T	Methylene Chloride	0.716	0.670	6.4	98	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.585	1.8	96	0.00
22 T	Diisopropyl ether	2.189	2.268	-3.6	100	0.00
23 T	Vinyl Acetate	1.925	1.977	-2.7	97	0.00
24 P	1,1-Dichloroethane	1.219	1.233	-1.1	98	0.00
25 T	2-Butanone	0.543	0.550	-1.3	99	0.00
26 T	2,2-Dichloropropane	0.954	0.998	-4.6	105	0.00
27 T	cis-1,2-Dichloroethene	0.718	0.699	2.6	95	0.00
28 T	Bromochloromethane	0.587	0.572	2.6	99	0.00
29 T	Tetrahydrofuran	0.340	0.344	-1.2	98	0.00
30 C	Chloroform	1.271	1.295	-1.9#	100	0.00
31 T	Cyclohexane	1.111	1.043	6.1	93	0.00
32 T	1,1,1-Trichloroethane	1.101	1.133	-2.9	100	0.00
33 S	1,2-Dichloroethane-d4	0.932	0.903	3.1	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.360	0.383	-6.4	105	0.00
36 T	1,1-Dichloropropene	0.484	0.477	1.4	94	0.00
37 T	Ethyl Acetate	0.598	0.609	-1.8	97	0.00
38 T	Carbon Tetrachloride	0.544	0.577	-6.1	101	0.00
39 T	Methylcyclohexane	0.623	0.602	3.4	92	0.00
40 TM	Benzene	1.417	1.454	-2.6	96	0.00
41 T	Methacrylonitrile	0.313	0.349	-11.5	98	0.00
42 TM	1,2-Dichloroethane	0.612	0.624	-2.0	97	0.00
43 T	Isopropyl Acetate	0.912	0.969	-6.2	98	0.00
44 TM	Trichloroethene	0.341	0.349	-2.3	96	0.00
45 C	1,2-Dichloropropane	0.352	0.380	-8.0#	100	0.00
46 T	Dibromomethane	0.278	0.282	-1.4	96	0.00
47 T	Bromodichloromethane	0.547	0.604	-10.4	102	0.00
48 T	Methyl methacrylate	0.466	0.507	-8.8	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046255.D
 Acq On : 19 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 19 23:05:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	95	0.00
50 S	Toluene-d8	1.246	1.254	-0.6	100	0.00
51 T	4-Methyl-2-Pentanone	0.605	0.650	-7.4	100	0.00
52 CM	Toluene	0.869	0.887	-2.1#	96	0.00
53 T	t-1,3-Dichloropropene	0.487	0.537	-10.3	99	0.00
54 T	cis-1,3-Dichloropropene	0.538	0.589	-9.5	99	0.00
55 T	1,1,2-Trichloroethane	0.343	0.363	-5.8	100	0.00
56 T	Ethyl methacrylate	0.546	0.599	-9.7	98	0.00
57 T	1,3-Dichloropropane	0.615	0.627	-2.0	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.278	0.301	-8.3	96	0.00
59 T	2-Hexanone	0.448	0.480	-7.1	99	0.00
60 T	Dibromochloromethane	0.376	0.429	-14.1	105	0.00
61 T	1,2-Dibromoethane	0.356	0.374	-5.1	98	0.00
62 S	4-Bromofluorobenzene	0.478	0.493	-3.1	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.354	0.357	-0.8	93	0.00
65 PM	Chlorobenzene	1.094	1.113	-1.7	99	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.403	-7.8	101	0.00
67 C	Ethyl Benzene	1.929	2.022	-4.8#	98	0.00
68 T	m/p-Xylenes	0.706	0.743	-5.2	98	0.00
69 T	o-Xylene	0.688	0.721	-4.8	97	0.00
70 T	Styrene	1.127	1.265	-12.2	101	0.00
71 P	Bromoform	0.281	0.322	-14.6	103	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	103	0.00
73 T	Isopropylbenzene	3.893	4.008	-3.0	100	0.00
74 T	N-amyl acetate	1.924	1.958	-1.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	1.364	1.325	2.9	102	0.00
76 T	1,2,3-Trichloropropane	1.204	1.185	1.6	103	0.00
77 T	Bromobenzene	0.904	0.924	-2.2	103	0.00
78 T	n-propylbenzene	4.526	4.732	-4.6	101	0.00
79 T	2-Chlorotoluene	2.919	2.904	0.5	100	0.00
80 T	1,3,5-Trimethylbenzene	3.252	3.398	-4.5	101	0.00
81 T	trans-1,4-Dichloro-2-butene	0.370	0.389	-5.1	104	0.00
82 T	4-Chlorotoluene	3.238	3.413	-5.4	103	0.00
83 T	tert-Butylbenzene	3.276	3.370	-2.9	101	0.00
84 T	1,2,4-Trimethylbenzene	3.293	3.436	-4.3	101	0.00
85 T	sec-Butylbenzene	4.022	4.230	-5.2	102	0.00
86 T	p-Isopropyltoluene	3.320	3.539	-6.6	103	0.00
87 T	1,3-Dichlorobenzene	1.649	1.676	-1.6	102	0.00
88 T	1,4-Dichlorobenzene	1.684	1.694	-0.6	103	0.00
89 T	n-Butylbenzene	2.912	3.195	-9.7	105	0.00
90 T	Hexachloroethane	0.585	0.628	-7.4	104	0.00
91 T	1,2-Dichlorobenzene	1.655	1.683	-1.7	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.302	0.319	-5.6	102	0.00
93 T	1,2,4-Trichlorobenzene	0.951	0.986	-3.7	104	0.00
94 T	Hexachlorobutadiene	0.415	0.436	-5.1	106	0.00
95 T	Naphthalene	3.487	3.458	0.8	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046255.D
Acq On : 19 May 2025 10:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 19 23:05:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.981	0.997	-1.6	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046255.D
 Acq On : 19 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 19 23:05:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	50.000	52.335	-4.7	93	0.00
3 P	Chloromethane	50.000	50.048	-0.1	96	0.00
4 C	Vinyl Chloride	50.000	48.749	2.5#	95	0.00
5 T	Bromomethane	50.000	46.363	7.3	91	0.00
6 T	Chloroethane	50.000	51.109	-2.2	100	0.00
7 T	Trichlorofluoromethane	50.000	51.448	-2.9	98	0.00
8 T	Diethyl Ether	50.000	48.729	2.5	101	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.119	-2.2	101	0.00
10 T	Methyl Iodide	50.000	45.327	9.3	84	0.00
11 T	Tert butyl alcohol	250.000	245.194	1.9	100	0.00
12 CM	1,1-Dichloroethene	50.000	48.293	3.4#	95	0.00
13 T	Acrolein	250.000	320.514	-28.2#	126	0.00
14 T	Allyl chloride	50.000	51.864	-3.7	100	0.00
15 T	Acrylonitrile	250.000	259.998	-4.0	100	0.00
16 T	Acetone	250.000	249.646	0.1	103	0.00
17 T	Carbon Disulfide	50.000	47.627	4.7	92	0.00
18 T	Methyl Acetate	50.000	57.353	-14.7	117	0.00
19 T	Methyl tert-butyl Ether	50.000	51.700	-3.4	100	0.00
20 T	Methylene Chloride	50.000	46.795	6.4	98	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.053	1.9	96	0.00
22 T	Diisopropyl ether	50.000	51.819	-3.6	100	0.00
23 T	Vinyl Acetate	250.000	256.722	-2.7	97	0.00
24 P	1,1-Dichloroethane	50.000	50.587	-1.2	98	0.00
25 T	2-Butanone	250.000	253.279	-1.3	99	0.00
26 T	2,2-Dichloropropane	50.000	52.317	-4.6	105	0.00
27 T	cis-1,2-Dichloroethene	50.000	48.707	2.6	95	0.00
28 T	Bromochloromethane	50.000	48.763	2.5	99	0.00
29 T	Tetrahydrofuran	250.000	253.062	-1.2	98	0.00
30 C	Chloroform	50.000	50.954	-1.9#	100	0.00
31 T	Cyclohexane	50.000	46.952	6.1	93	0.00
32 T	1,1,1-Trichloroethane	50.000	51.415	-2.8	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.428	3.1	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	98	0.00
35 S	Dibromofluoromethane	50.000	53.185	-6.4	105	0.00
36 T	1,1-Dichloropropene	50.000	49.311	1.4	94	0.00
37 T	Ethyl Acetate	50.000	50.949	-1.9	97	0.00
38 T	Carbon Tetrachloride	50.000	53.099	-6.2	101	0.00
39 T	Methylcyclohexane	50.000	48.322	3.4	92	0.00
40 TM	Benzene	50.000	51.321	-2.6	96	0.00
41 T	Methacrylonitrile	50.000	55.825	-11.7	98	0.00
42 TM	1,2-Dichloroethane	50.000	51.038	-2.1	97	0.00
43 T	Isopropyl Acetate	50.000	53.135	-6.3	98	0.00
44 TM	Trichloroethene	50.000	51.110	-2.2	96	0.00
45 C	1,2-Dichloropropane	50.000	53.899	-7.8#	100	0.00
46 T	Dibromomethane	50.000	50.726	-1.5	96	0.00
47 T	Bromodichloromethane	50.000	55.166	-10.3	102	0.00
48 T	Methyl methacrylate	50.000	54.480	-9.0	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046255.D
 Acq On : 19 May 2025 10:05
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: May 19 23:05:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	997.932	0.2	95	0.00
50 S	Toluene-d8	50.000	50.315	-0.6	100	0.00
51 T	4-Methyl-2-Pentanone	250.000	268.553	-7.4	100	0.00
52 CM	Toluene	50.000	51.066	-2.1#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	55.186	-10.4	99	0.00
54 T	cis-1,3-Dichloropropene	50.000	54.783	-9.6	99	0.00
55 T	1,1,2-Trichloroethane	50.000	52.963	-5.9	100	0.00
56 T	Ethyl methacrylate	50.000	54.853	-9.7	98	0.00
57 T	1,3-Dichloropropane	50.000	50.945	-1.9	98	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	270.687	-8.3	96	0.00
59 T	2-Hexanone	250.000	267.754	-7.1	99	0.00
60 T	Dibromochloromethane	50.000	57.019	-14.0	105	0.00
61 T	1,2-Dibromoethane	50.000	52.476	-5.0	98	0.00
62 S	4-Bromofluorobenzene	50.000	51.525	-3.0	102	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	98	0.00
64 T	Tetrachloroethene	50.000	50.484	-1.0	93	0.00
65 PM	Chlorobenzene	50.000	50.864	-1.7	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.854	-7.7	101	0.00
67 C	Ethyl Benzene	50.000	52.399	-4.8#	98	0.00
68 T	m/p-Xylenes	100.000	105.368	-5.4	98	0.00
69 T	o-Xylene	50.000	52.434	-4.9	97	0.00
70 T	Styrene	50.000	56.113	-12.2	101	0.00
71 P	Bromoform	50.000	57.309	-14.6	103	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	103	0.00
73 T	Isopropylbenzene	50.000	51.487	-3.0	100	0.00
74 T	N-amyl acetate	50.000	50.902	-1.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	48.561	2.9	102	0.00
76 T	1,2,3-Trichloropropane	50.000	49.251	1.5	103	0.00
77 T	Bromobenzene	50.000	51.130	-2.3	103	0.00
78 T	n-propylbenzene	50.000	52.270	-4.5	101	0.00
79 T	2-Chlorotoluene	50.000	49.734	0.5	100	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.243	-4.5	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	52.574	-5.1	104	0.00
82 T	4-Chlorotoluene	50.000	52.703	-5.4	103	0.00
83 T	tert-Butylbenzene	50.000	51.435	-2.9	101	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.167	-4.3	101	0.00
85 T	sec-Butylbenzene	50.000	52.585	-5.2	102	0.00
86 T	p-Isopropyltoluene	50.000	53.297	-6.6	103	0.00
87 T	1,3-Dichlorobenzene	50.000	50.815	-1.6	102	0.00
88 T	1,4-Dichlorobenzene	50.000	50.288	-0.6	103	0.00
89 T	n-Butylbenzene	50.000	54.865	-9.7	105	0.00
90 T	Hexachloroethane	50.000	53.653	-7.3	104	0.00
91 T	1,2-Dichlorobenzene	50.000	50.839	-1.7	103	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.775	-5.5	102	0.00
93 T	1,2,4-Trichlorobenzene	50.000	51.847	-3.7	104	0.00
94 T	Hexachlorobutadiene	50.000	52.507	-5.0	106	0.00
95 T	Naphthalene	50.000	49.585	0.8	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046255.D
Acq On : 19 May 2025 10:05
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: May 19 23:05:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.799	-1.6	101	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6





QC SAMPLE DATA

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046038.D
Acq On : 05 May 2025 09:37
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631


Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.1	8676	PASS
75	95	30	60	56.2	22052	PASS
95	95	100	100	100.0	39213	PASS
96	95	5	9	6.4	2505	PASS
173	174	0.00	2	0.7	185	PASS
174	95	50	100	68.8	26963	PASS
175	174	5	9	7.3	1980	PASS
176	174	95	101	97.0	26147	PASS
177	176	5	9	6.9	1802	PASS

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046254.D
Acq On : 19 May 2025 09:33
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Title : SW846 8260
Last Update : Tue May 06 07:12:22 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.1	8149	PASS
75	95	30	60	58.5	21593	PASS
95	95	100	100	100.0	36885	PASS
96	95	5	9	6.6	2433	PASS
173	174	0.00	2	1.1	273	PASS
174	95	50	100	67.9	25035	PASS
175	174	5	9	8.1	2023	PASS
176	174	95	101	99.1	24816	PASS
177	176	5	9	6.9	1720	PASS

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Nelson		Date Received:	
Client Sample ID:	VX0519WBL01		SDG No.:	Q2073
Lab Sample ID:	VX0519WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046257.D	1		05/19/25 11:02	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	0.22	U	0.22	1.00	ug/L
74-87-3	Chloromethane	0.32	U	0.32	1.00	ug/L
75-01-4	Vinyl Chloride	0.26	U	0.26	1.00	ug/L
74-83-9	Bromomethane	1.40	U	1.40	5.00	ug/L
75-00-3	Chloroethane	0.47	U	0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.33	U	0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.25	U	0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.23	U	0.23	1.00	ug/L
67-64-1	Acetone	1.50	U	1.50	5.00	ug/L
75-15-0	Carbon Disulfide	0.21	U	0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.16	U	0.16	1.00	ug/L
79-20-9	Methyl Acetate	0.27	U	0.27	1.00	ug/L
75-09-2	Methylene Chloride	0.28	U	0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.23	U	0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.23	U	0.23	1.00	ug/L
110-82-7	Cyclohexane	1.50	U	1.50	5.00	ug/L
78-93-3	2-Butanone	0.98	U	0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.25	U	0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.19	U	0.19	1.00	ug/L
74-97-5	Bromochloromethane	0.22	U	0.22	1.00	ug/L
67-66-3	Chloroform	0.25	U	0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.20	U	0.20	1.00	ug/L
108-87-2	Methylcyclohexane	0.16	U	0.16	1.00	ug/L
71-43-2	Benzene	0.15	U	0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.22	U	0.22	1.00	ug/L
79-01-6	Trichloroethene	0.090	U	0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.20	U	0.20	1.00	ug/L
75-27-4	Bromodichloromethane	0.22	U	0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	0.68	U	0.68	5.00	ug/L
108-88-3	Toluene	0.14	U	0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental	Date Collected:	
Project:	Nelson	Date Received:	
Client Sample ID:	VX0519WBL01	SDG No.:	Q2073
Lab Sample ID:	VX0519WBL01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046257.D	1		05/19/25 11:02	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.17	U	0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.16	U	0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.21	U	0.21	1.00	ug/L
591-78-6	2-Hexanone	0.89	U	0.89	5.00	ug/L
124-48-1	Dibromochloromethane	0.18	U	0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.15	U	0.15	1.00	ug/L
127-18-4	Tetrachloroethene	0.23	U	0.23	1.00	ug/L
108-90-7	Chlorobenzene	0.12	U	0.12	1.00	ug/L
100-41-4	Ethyl Benzene	0.13	U	0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	0.24	U	0.24	2.00	ug/L
95-47-6	o-Xylene	0.12	U	0.12	1.00	ug/L
100-42-5	Styrene	0.15	U	0.15	1.00	ug/L
75-25-2	Bromoform	0.19	U	0.19	1.00	ug/L
98-82-8	Isopropylbenzene	0.12	U	0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.26	U	0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.19	U	0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.16	U	0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.53	U	0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.20	U	0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	54.1		70 (74) - 130 (125)	108%	SPK: 50
1868-53-7	Dibromofluoromethane	51.4		70 (75) - 130 (124)	103%	SPK: 50
2037-26-5	Toluene-d8	50.4		70 (86) - 130 (113)	101%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.6		70 (77) - 130 (121)	99%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	62500	5.544			
540-36-3	1,4-Difluorobenzene	125000	6.757			
3114-55-4	Chlorobenzene-d5	118000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	48300	12.018			

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Nelson		Date Received:	
Client Sample ID:	VX0519WBL01		SDG No.:	Q2073
Lab Sample ID:	VX0519WBL01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046257.D	1		05/19/25 11:02	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046257.D
 Acq On : 19 May 2025 11:02
 Operator : JC/MD
 Sample : VX0519WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0519WBL01

Quant Time: May 19 23:06:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124684	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	117918	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48295	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63090	54.143	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.280%	
35) Dibromofluoromethane	5.385	113	46152	51.402	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.800%	
50) Toluene-d8	8.647	98	156579	50.386	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.780%	
62) 4-Bromofluorobenzene	11.079	95	59083	49.565	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 99.120%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046257.D
Acq On : 19 May 2025 11:02
Operator : JC/MD
Sample : VX0519WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519WBL01

Quant Time: May 19 23:06:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration



#1
Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.544 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX046257.D

Acq: 19 May 2025 11:02

Instrument :

MSVOA_X

ClientSampleId :





VX0519WBL01

Tgt Ion:168 Resp: 62503

Ion Ratio Lower Upper

168 100

99 74.3 54.9 82.3



#33

1,2-Dichloroethane-d4

Concen: 54.143 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX046257.D




Acq: 19 May 2025 11:02


Tgt Ion: 65 Resp: 63090

Ion Ratio Lower Upper

65 100

67 49.5 0.0 99.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX046257.D
Acq: 19 May 2025 11:02




Instrument :
MSVOA_X
ClientSampleId :
VX0519WBL01


Tgt Ion:114 Resp: 124684
Ion Ratio Lower Upper
114 100
63 23.7 0.0 49.2
88 18.5 0.0 33.6



#35
Dibromofluoromethane
Concen: 51.402 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX046257.D
Acq: 19 May 2025 11:02

Tgt Ion:113 Resp: 46152
Ion Ratio Lower Upper
113 100
111 103.3 83.1 124.7
192 17.0 13.3 19.9





#50

Toluene-d8

Concen: 50.386 ug/l

RT: 8.647 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX046257.D

Acq: 19 May 2025 11:02

Instrument :

MSVOA_X

ClientSampleId :





VX0519WBL01

Tgt Ion: 98 Resp: 156579

Ion Ratio Lower Upper

98 100

100 65.9 53.5 80.3



#62

4-Bromofluorobenzene

Concen: 49.565 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.000 min

Lab File: VX046257.D

Acq: 19 May 2025 11:02




Tgt Ion: 95 Resp: 59083


Ion Ratio Lower Upper

95 100

174 67.5 0.0 135.8

176 64.9 0.0 131.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. -0.000 min
Lab File: VX046257.D
Acq: 19 May 2025 11:02




Instrument :
MSVOA_X
ClientSampleId :
VX0519WBL01

Tgt Ion:117 Resp: 117918
Ion Ratio Lower Upper
117 100
82 62.0 50.6 76.0
119 31.6 25.8 38.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VX046257.D
Acq: 19 May 2025 11:02

Tgt Ion:152 Resp: 48295
Ion Ratio Lower Upper
152 100
115 64.3 46.9 140.7
150 156.2 0.0 351.0



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046257.D
 Acq On : 19 May 2025 11:02
 Operator : JC/MD
 Sample : VX0519WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0519WBL01

Integration Parameters: RTEINT.P

Integrator: RTE

Smoothing : ON

Filtering: 5

Sampling : 1

Min Area: 3 % of largest Peak

Start Thrs: 0.2

Max Peaks: 100

Stop Thrs : 0

Peak Location: TOP

If leading or trailing edge < 100 prefer < Baseline drop else tangent >

Peak separation: 5

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M

Title : SW846 8260

Signal : TIC: VX046257.D\data.ms

peak #	R.T. min	first scan	max scan	last scan	PK TY	peak height	corr. area	corr. % max.	% of total
1	1.258	23	28	38	rBV2	5281	11818	2.69%	0.506%
2	1.575	71	80	81	rBV2	4413	9560	2.18%	0.410%
3	1.611	85	86	93	rVB4	6517	10154	2.31%	0.435%
4	5.385	693	705	721	rBV	52968	157953	35.98%	6.768%
5	5.544	721	731	751	rVB	73875	210790	48.02%	9.031%
6	5.952	788	798	814	rBV	60739	165458	37.69%	7.089%
7	6.757	921	930	945	rBV	129851	316599	72.12%	13.565%
8	8.647	1233	1240	1254	rBV	282503	438960	100.00%	18.807%
9	10.049	1465	1470	1485	rBV	287220	396808	90.40%	17.001%
10	11.079	1634	1639	1652	rBV	226830	288494	65.72%	12.361%
11	12.018	1788	1793	1806	rBV	263994	327397	74.58%	14.027%


Sum of corrected areas: 2333991

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046257.D
Acq On : 19 May 2025 11:02
Operator : JC/MD
Sample : VX0519WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046257.D
Acq On : 19 May 2025 11:02
Operator : JC/MD
Sample : VX0519WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

No Library Search Compounds Detected

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046257.D
Acq On : 19 May 2025 11:02
Operator : JC/MD
Sample : VX0519WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519WBL01

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260

TIC Library : C:\Database\NIST20.L
TIC Integration Parameters: LSCINT.P

TIC Top Hit name	RT	EstConc	Units	Response	#	RT	Resp	Conc
------------------	----	---------	-------	----------	---	----	------	------

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Nelson		Date Received:	
Client Sample ID:	VX0519WBS01		SDG No.:	Q2073
Lab Sample ID:	VX0519WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046258.D	1		05/19/25 11:25	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	18.7		0.22	1.00	ug/L
74-87-3	Chloromethane	17.7		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	17.2		0.26	1.00	ug/L
74-83-9	Bromomethane	18.6		1.40	5.00	ug/L
75-00-3	Chloroethane	19.3		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.8		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	19.0		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	18.4		0.23	1.00	ug/L
67-64-1	Acetone	95.5		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	15.7		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.0		0.16	1.00	ug/L
79-20-9	Methyl Acetate	22.6		0.27	1.00	ug/L
75-09-2	Methylene Chloride	18.1		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.2		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.3		0.23	1.00	ug/L
110-82-7	Cyclohexane	17.8		1.50	5.00	ug/L
78-93-3	2-Butanone	98.6		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	19.1		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	18.4		0.19	1.00	ug/L
74-97-5	Bromochloromethane	21.4		0.22	1.00	ug/L
67-66-3	Chloroform	19.7		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.1		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	17.6		0.16	1.00	ug/L
71-43-2	Benzene	19.0		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	19.8		0.22	1.00	ug/L
79-01-6	Trichloroethene	18.2		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	19.9		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	20.0		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	99.9		0.68	5.00	ug/L
108-88-3	Toluene	19.5		0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental		Date Collected:	
Project:	Nelson		Date Received:	
Client Sample ID:	VX0519WBS01		SDG No.:	Q2073
Lab Sample ID:	VX0519WBS01		Matrix:	Water
Analytical Method:	8260D		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046258.D	1		05/19/25 11:25	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.5		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.0		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.4		0.21	1.00	ug/L
591-78-6	2-Hexanone	100		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	19.9		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	19.6		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	17.9		0.23	1.00	ug/L
108-90-7	Chlorobenzene	18.5		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	18.9		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	37.1		0.24	2.00	ug/L
95-47-6	o-Xylene	19.0		0.12	1.00	ug/L
100-42-5	Styrene	19.7		0.15	1.00	ug/L
75-25-2	Bromoform	19.0		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	19.5		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.3		0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.8		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.2		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.3		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	19.0		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	17.8		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	18.1		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	51.6		70 (74) - 130 (125)	103%	SPK: 50
1868-53-7	Dibromofluoromethane	53.1		70 (75) - 130 (124)	106%	SPK: 50
2037-26-5	Toluene-d8	51.6		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.4		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	91600	5.544			
540-36-3	1,4-Difluorobenzene	159000	6.757			
3114-55-4	Chlorobenzene-d5	143000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	66900	12.018			

Report of Analysis

Client:	G Environmental			Date Collected:	
Project:	Nelson			Date Received:	
Client Sample ID:	VX0519WBS01			SDG No.:	Q2073
Lab Sample ID:	VX0519WBS01			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046258.D	1		05/19/25 11:25	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046258.D
 Acq On : 19 May 2025 11:25
 Operator : JC/MD
 Sample : VX0519WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0519WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/21/2025
 Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:06:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	91585	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	158835	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	142753	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	66879	50.000	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	5.952	65	88116	51.607	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 103.220%	
35) Dibromofluoromethane	5.385	113	60766	53.127	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.260%	
50) Toluene-d8	8.647	98	204172	51.575	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.140%	
62) 4-Bromofluorobenzene	11.079	95	78102	51.433	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.860%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.166	85	26163	18.664	ug/l	97
3) Chloromethane	1.307	50	24022	17.672	ug/l	99
4) Vinyl Chloride	1.374	62	21728	17.174	ug/l	99
5) Bromomethane	1.599	94	10888	18.555	ug/l	99
6) Chloroethane	1.672	64	13012	19.266	ug/l	99
7) Trichlorofluoromethane	1.880	101	35167	18.808	ug/l	96
8) Diethyl Ether	2.136	74	11436	17.967	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	21986	19.001	ug/l	98
10) Methyl Iodide	2.447	142	22257	16.256	ug/l	99
11) Tert butyl alcohol	2.971	59	23079	96.294	ug/l	100
12) 1,1-Dichloroethene	2.312	96	19928	18.350	ug/l	99
13) Acrolein	2.239	56	21625	79.227	ug/l	99
14) Allyl chloride	2.660	41	39571	19.066	ug/l	97
15) Acrylonitrile	3.062	53	68193	99.504	ug/l	99
16) Acetone	2.386	43	65401	95.531	ug/l	100
17) Carbon Disulfide	2.508	76	40417	15.698	ug/l	98
18) Methyl Acetate	2.703	43	35904	22.601	ug/l	100
19) Methyl tert-butyl Ether	3.117	73	72372	19.009	ug/l	98
20) Methylene Chloride	2.782	84	23718	18.079	ug/l	94
21) trans-1,2-Dichloroethene	3.087	96	19873	18.197	ug/l	99
22) Diisopropyl ether	3.757	45	77609	19.358	ug/l	89
23) Vinyl Acetate	3.721	43	330626	93.764	ug/l	99
24) 1,1-Dichloroethane	3.605	63	43022	19.267	ug/l	98
25) 2-Butanone	4.556	43	98037	98.639	ug/l	98
26) 2,2-Dichloropropane	4.471	77	32249	18.451	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	24145	18.365	ug/l	98
28) Bromochloromethane	4.897	49	23030	21.426	ug/l	96
29) Tetrahydrofuran	5.007	42	61341	98.492	ug/l	99
30) Chloroform	5.086	83	45790	19.674	ug/l	97
31) Cyclohexane	5.458	56	36164	17.773	ug/l	93
32) 1,1,1-Trichloroethane	5.373	97	38499	19.082	ug/l	99
36) 1,1-Dichloropropene	5.690	75	27229	17.718	ug/l	97
37) Ethyl Acetate	4.714	43	35800	18.855	ug/l	99
38) Carbon Tetrachloride	5.672	117	32935	19.074	ug/l	96
39) Methylcyclohexane	7.372	83	34876	17.628	ug/l	96
40) Benzene	6.031	78	85691	19.037	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046258.D
 Acq On : 19 May 2025 11:25
 Operator : JC/MD
 Sample : VX0519WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0519WBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/21/2025
 Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:06:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	19793	19.928	ug/l	97
42) 1,2-Dichloroethane	6.080	62	38460	19.797	ug/l	100
43) Isopropyl Acetate	6.342	43	56874	19.635	ug/l	98
44) Trichloroethene	7.123	130	19711	18.194	ug/l	90
45) 1,2-Dichloropropane	7.427	63	22322	19.943	ug/l	98
46) Dibromomethane	7.580	93	17167	19.446	ug/l	98
47) Bromodichloromethane	7.818	83	34706	19.960	ug/l	96
48) Methyl methacrylate	7.690	41	29632	20.031	ug/l	98
49) 1,4-Dioxane	7.659	88	11445	407.470	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	192103	99.913	ug/l	99
52) Toluene	8.714	92	53924	19.536	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	28589	18.499	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	32448	18.996	ug/l	94
55) 1,1,2-Trichloroethane	9.147	97	22179	20.379	ug/l	96
56) Ethyl methacrylate	9.116	69	33552	19.343	ug/l	98
57) 1,3-Dichloropropane	9.305	76	38236	19.562	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	82645	93.456	ug/l	100
59) 2-Hexanone	9.427	43	143585	100.940	ug/l	99
60) Dibromochloromethane	9.518	129	23814	19.924	ug/l	100
61) 1,2-Dibromoethane	9.610	107	22227	19.649	ug/l	98
64) Tetrachloroethene	9.268	164	18066	17.886	ug/l	96
65) Chlorobenzene	10.073	112	57845	18.513	ug/l	97
66) 1,1,1,2-Tetrachloroethane	10.159	131	20506	19.220	ug/l	100
67) Ethyl Benzene	10.189	91	104062	18.894	ug/l	99
68) m/p-Xylenes	10.299	106	74816	37.141	ug/l	96
69) o-Xylene	10.640	106	37343	19.015	ug/l	97
70) Styrene	10.652	104	63479	19.732	ug/l	99
71) Bromoform	10.799	173	15204	18.981	ug/l #	97
73) Isopropylbenzene	10.957	105	101409	19.477	ug/l	100
74) N-amyl acetate	10.841	43	48129	18.706	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.207	83	35171	19.276	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	30324m	18.837	ug/l	
77) Bromobenzene	11.195	156	22145	18.320	ug/l	95
78) n-propylbenzene	11.299	91	113110	18.683	ug/l	100
79) 2-Chlorotoluene	11.360	91	72633	18.600	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	84589	19.446	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	8340	16.867	ug/l	93
82) 4-Chlorotoluene	11.451	91	82807	19.122	ug/l	100
83) tert-Butylbenzene	11.713	119	85062	19.414	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	85052	19.308	ug/l	100
85) sec-Butylbenzene	11.890	105	102527	19.058	ug/l	99
86) p-Isopropyltoluene	12.006	119	84578	19.046	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	41388	18.761	ug/l	99
88) 1,4-Dichlorobenzene	12.036	146	40939	18.171	ug/l	96
89) n-Butylbenzene	12.329	91	70827	18.183	ug/l	99
90) Hexachloroethane	12.536	117	14388	18.391	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	42690	19.283	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	7673	18.983	ug/l	96
93) 1,2,4-Trichlorobenzene	13.585	180	22616	17.786	ug/l	96
94) Hexachlorobutadiene	13.725	225	10164	18.303	ug/l	97
95) Naphthalene	13.774	128	79390	17.024	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	23795	18.136	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046258.D
Acq On : 19 May 2025 11:25
Operator : JC/MD
Sample : VX0519WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519WBS01

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/21/2025
Supervised By :Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:06:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046258.D
Acq On : 19 May 2025 11:25
Operator : JC/MD
Sample : VX0519WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519WBS01

Manual Integrations APPROVED

Reviewed By :John Carlone 05/21/2025
Supervised By :Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:06:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration



Report of Analysis

Client:	G Environmental	Date Collected:	
Project:	Nelson	Date Received:	
Client Sample ID:	VX0519WBSD01	SDG No.:	Q2073
Lab Sample ID:	VX0519WBSD01	Matrix:	Water
Analytical Method:	8260D	% Solid:	0
Sample Wt/Vol:	5 Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:	uL	Test:	VOCMS Group1
GC Column:	DB-624UI ID : 0.18	Level :	LOW
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046259.D	1		05/19/25 11:53	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
75-71-8	Dichlorodifluoromethane	18.0		0.22	1.00	ug/L
74-87-3	Chloromethane	17.3		0.32	1.00	ug/L
75-01-4	Vinyl Chloride	16.7		0.26	1.00	ug/L
74-83-9	Bromomethane	17.2		1.40	5.00	ug/L
75-00-3	Chloroethane	17.8		0.47	1.00	ug/L
75-69-4	Trichlorofluoromethane	18.4		0.33	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	18.5		0.25	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.9		0.23	1.00	ug/L
67-64-1	Acetone	98.7		1.50	5.00	ug/L
75-15-0	Carbon Disulfide	15.1		0.21	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.5		0.16	1.00	ug/L
79-20-9	Methyl Acetate	23.4		0.27	1.00	ug/L
75-09-2	Methylene Chloride	17.8		0.28	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	18.0		0.23	1.00	ug/L
75-34-3	1,1-Dichloroethane	19.0		0.23	1.00	ug/L
110-82-7	Cyclohexane	17.6		1.50	5.00	ug/L
78-93-3	2-Butanone	100		0.98	5.00	ug/L
56-23-5	Carbon Tetrachloride	18.8		0.25	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	19.0		0.19	1.00	ug/L
74-97-5	Bromochloromethane	20.9		0.22	1.00	ug/L
67-66-3	Chloroform	19.9		0.25	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	19.2		0.20	1.00	ug/L
108-87-2	Methylcyclohexane	17.2		0.16	1.00	ug/L
71-43-2	Benzene	19.1		0.15	1.00	ug/L
107-06-2	1,2-Dichloroethane	20.3		0.22	1.00	ug/L
79-01-6	Trichloroethene	18.6		0.090	1.00	ug/L
78-87-5	1,2-Dichloropropane	18.7		0.20	1.00	ug/L
75-27-4	Bromodichloromethane	19.9		0.22	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	100		0.68	5.00	ug/L
108-88-3	Toluene	19.1		0.14	1.00	ug/L

Report of Analysis

Client:	G Environmental			Date Collected:	
Project:	Nelson			Date Received:	
Client Sample ID:	VX0519WBSD01			SDG No.:	Q2073
Lab Sample ID:	VX0519WBSD01			Matrix:	Water
Analytical Method:	8260D			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOCMS Group1
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046259.D	1		05/19/25 11:53	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	18.8		0.17	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	19.3		0.16	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	20.3		0.21	1.00	ug/L
591-78-6	2-Hexanone	110		0.89	5.00	ug/L
124-48-1	Dibromochloromethane	19.7		0.18	1.00	ug/L
106-93-4	1,2-Dibromoethane	20.2		0.15	1.00	ug/L
127-18-4	Tetrachloroethene	18.2		0.23	1.00	ug/L
108-90-7	Chlorobenzene	18.7		0.12	1.00	ug/L
100-41-4	Ethyl Benzene	19.0		0.13	1.00	ug/L
179601-23-1	m/p-Xylenes	38.2		0.24	2.00	ug/L
95-47-6	o-Xylene	19.1		0.12	1.00	ug/L
100-42-5	Styrene	19.8		0.15	1.00	ug/L
75-25-2	Bromoform	19.2		0.19	1.00	ug/L
98-82-8	Isopropylbenzene	19.2		0.12	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	19.5		0.26	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	18.8		0.16	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	18.8		0.19	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	19.8		0.16	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	20.0		0.53	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	18.6		0.20	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	19.4		0.20	1.00	ug/L
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	52.0		70 (74) - 130 (125)	104%	SPK: 50
1868-53-7	Dibromofluoromethane	52.9		70 (75) - 130 (124)	106%	SPK: 50
2037-26-5	Toluene-d8	51.4		70 (86) - 130 (113)	103%	SPK: 50
460-00-4	4-Bromofluorobenzene	51.5		70 (77) - 130 (121)	103%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	87300	5.544			
540-36-3	1,4-Difluorobenzene	151000	6.757			
3114-55-4	Chlorobenzene-d5	132000	10.049			
3855-82-1	1,4-Dichlorobenzene-d4	63600	12.018			

Report of Analysis

Client:	G Environmental			Date Collected:			
Project:	Nelson			Date Received:			
Client Sample ID:	VX0519WBSD01			SDG No.:	Q2073		
Lab Sample ID:	VX0519WBSD01			Matrix:	Water		
Analytical Method:	8260D			% Solid:	0		
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000	uL	
Soil Aliquot Vol:			uL	Test:	VOCMS Group1		
GC Column:	DB-624UI	ID :	0.18	Level :	LOW		
Prep Method :							

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX046259.D	1		05/19/25 11:53	VX051925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
LOQ = Limit of Quantitation
MDL = Method Detection Limit
LOD = Limit of Detection
E = Value Exceeds Calibration Range
Q = indicates LCS control criteria did not meet requirements
M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
B = Analyte Found in Associated Method Blank
N = Presumptive Evidence of a Compound
* = Values outside of QC limits
D = Dilution
() = Laboratory InHouse Limit
A = Aldol-Condensation Reaction Products

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046259.D
 Acq On : 19 May 2025 11:53
 Operator : JC/MD
 Sample : VX0519WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0519WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/21/2025
 Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:07:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	87263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	150978	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	132315	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	63585	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	84643	52.028	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 104.060%			
35) Dibromofluoromethane	5.379	113	57543	52.928	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.860%			
50) Toluene-d8	8.647	98	193249	51.356	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.720%			
62) 4-Bromofluorobenzene	11.079	95	74331	51.497	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 103.000%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.167	85	24053	18.009	ug/l	96
3) Chloromethane	1.307	50	22394	17.290	ug/l	99
4) Vinyl Chloride	1.374	62	20162	16.726	ug/l	96
5) Bromomethane	1.593	94	9591	17.154	ug/l	100
6) Chloroethane	1.673	64	11432	17.765	ug/l	100
7) Trichlorofluoromethane	1.880	101	32823	18.424	ug/l	99
8) Diethyl Ether	2.136	74	11155	18.393	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	20365	18.472	ug/l	99
10) Methyl Iodide	2.447	142	21735	16.661	ug/l	99
11) Tert butyl alcohol	2.971	59	23837	104.382	ug/l	97
12) 1,1-Dichloroethene	2.313	96	18561	17.938	ug/l	99
13) Acrolein	2.233	56	20993	80.721	ug/l	100
14) Allyl chloride	2.654	41	37096	18.759	ug/l	97
15) Acrylonitrile	3.063	53	66231	101.428	ug/l	99
16) Acetone	2.380	43	64397	98.724	ug/l	99
17) Carbon Disulfide	2.502	76	36970	15.071	ug/l	97
18) Methyl Acetate	2.703	43	35356	23.358	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	70635	19.471	ug/l	95
20) Methylene Chloride	2.782	84	22283	17.826	ug/l	98
21) trans-1,2-Dichloroethene	3.087	96	18725	17.995	ug/l	95
22) Diisopropyl ether	3.758	45	75925	19.876	ug/l	93
23) Vinyl Acetate	3.721	43	324003	96.437	ug/l	99
24) 1,1-Dichloroethane	3.599	63	40455	19.014	ug/l	98
25) 2-Butanone	4.556	43	96498	101.899	ug/l	98
26) 2,2-Dichloropropane	4.465	77	29528	17.731	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	23862	19.049	ug/l	97
28) Bromochloromethane	4.891	49	21443	20.938	ug/l	99
29) Tetrahydrofuran	5.001	42	61012	102.816	ug/l	98
30) Chloroform	5.087	83	44217	19.939	ug/l	100
31) Cyclohexane	5.465	56	34187	17.633	ug/l	97
32) 1,1,1-Trichloroethane	5.379	97	36865	19.177	ug/l	99
36) 1,1-Dichloropropene	5.684	75	25897	17.728	ug/l	98
37) Ethyl Acetate	4.715	43	34382	19.051	ug/l	99
38) Carbon Tetrachloride	5.672	117	30875	18.811	ug/l	99
39) Methylcyclohexane	7.373	83	32313	17.182	ug/l	96
40) Benzene	6.032	78	81819	19.122	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
 Data File : VX046259.D
 Acq On : 19 May 2025 11:53
 Operator : JC/MD
 Sample : VX0519WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0519WBSD01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/21/2025
 Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:07:49 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	20373	21.579	ug/l	99
42) 1,2-Dichloroethane	6.080	62	37427	20.267	ug/l	99
43) Isopropyl Acetate	6.336	43	54907	19.942	ug/l	100
44) Trichloroethene	7.123	130	19111	18.558	ug/l	97
45) 1,2-Dichloropropane	7.428	63	19869	18.675	ug/l	98
46) Dibromomethane	7.580	93	16651	19.843	ug/l	100
47) Bromodichloromethane	7.818	83	32824	19.861	ug/l	99
48) Methyl methacrylate	7.690	41	29191	20.760	ug/l	95
49) 1,4-Dioxane	7.659	88	11265	421.933	ug/l	97
51) 4-Methyl-2-Pentanone	8.568	43	190473	104.221	ug/l	100
52) Toluene	8.714	92	50238	19.148	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	27640	18.815	ug/l	99
54) cis-1,3-Dichloropropene	8.360	75	31384	19.329	ug/l	95
55) 1,1,2-Trichloroethane	9.147	97	21003	20.302	ug/l	97
56) Ethyl methacrylate	9.116	69	32978	20.001	ug/l	98
57) 1,3-Dichloropropane	9.305	76	37223	20.035	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	81477	96.930	ug/l	98
59) 2-Hexanone	9.427	43	143472	106.109	ug/l	100
60) Dibromochloromethane	9.519	129	22336	19.660	ug/l	98
61) 1,2-Dibromoethane	9.604	107	21729	20.209	ug/l	98
64) Tetrachloroethene	9.269	164	17073	18.237	ug/l	90
65) Chlorobenzene	10.073	112	54256	18.735	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	19365	19.582	ug/l	99
67) Ethyl Benzene	10.189	91	96865	18.975	ug/l	99
68) m/p-Xylenes	10.299	106	71386	38.234	ug/l	97
69) o-Xylene	10.640	106	34722	19.075	ug/l	95
70) Styrene	10.653	104	59040	19.800	ug/l	99
71) Bromoform	10.799	173	14287	19.243	ug/l #	96
73) Isopropylbenzene	10.957	105	94808	19.152	ug/l	100
74) N-amyl acetate	10.842	43	47522	19.427	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.207	83	33849	19.512	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	29880m	19.523	ug/l	
77) Bromobenzene	11.195	156	22274	19.381	ug/l	97
78) n-propylbenzene	11.299	91	110079	19.125	ug/l	99
79) 2-Chlorotoluene	11.360	91	69282	18.661	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	79492	19.221	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	7819	16.632	ug/l	88
82) 4-Chlorotoluene	11.451	91	79386	19.282	ug/l	100
83) tert-Butylbenzene	11.713	119	78714	18.896	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	81650	19.496	ug/l	98
85) sec-Butylbenzene	11.890	105	98747	19.306	ug/l	100
86) p-Isopropyltoluene	12.006	119	81895	19.398	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	39428	18.798	ug/l	99
88) 1,4-Dichlorobenzene	12.037	146	40234	18.783	ug/l	98
89) n-Butylbenzene	12.329	91	69047	18.645	ug/l	99
90) Hexachloroethane	12.536	117	14001	18.824	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	41580	19.755	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	7677	19.976	ug/l	99
93) 1,2,4-Trichlorobenzene	13.585	180	22494	18.607	ug/l	97
94) Hexachlorobutadiene	13.719	225	10367	19.635	ug/l	98
95) Naphthalene	13.774	128	81954	18.484	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	24173	19.379	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046259.D
Acq On : 19 May 2025 11:53
Operator : JC/MD
Sample : VX0519WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519WBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/21/2025
Supervised By :Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:07:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX051925\
Data File : VX046259.D
Acq On : 19 May 2025 11:53
Operator : JC/MD
Sample : VX0519WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0519WBSD01

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/21/2025
Supervised By : Mahesh Dadoda 05/21/2025

Quant Time: May 19 23:07:49 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration



Manual Integration Report

Sequence:	VX050525	Instrument	MSVOA_x
-----------	----------	------------	---------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDICC020	VX046041.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:13 AM	MMDadoda	5/6/2025 12:42:46 PM	Peak Integrated by Software
VSTDICCC050	VX046042.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:18 AM	MMDadoda	5/6/2025 12:42:48 PM	Peak Integrated by Software
VSTDICC100	VX046043.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:22 AM	MMDadoda	5/6/2025 12:42:50 PM	Peak Integrated by Software
VSTDICC150	VX046044.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:27 AM	MMDadoda	5/6/2025 12:42:53 PM	Peak Integrated by Software
VSTDICC005	VX046046.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:32 AM	MMDadoda	5/6/2025 12:42:56 PM	Peak Integrated by Software
VSTDICC005	VX046046.D	Ethyl Acetate	JOHN	5/6/2025 9:53:32 AM	MMDadoda	5/6/2025 12:42:56 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	1,4-Dichlorobenzene	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Bromochloromethane	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Ethyl Acetate	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICC001	VX046047.D	Methyl methacrylate	JOHN	5/6/2025 9:53:38 AM	MMDadoda	5/6/2025 12:41:35 PM	Peak Integrated by Software
VSTDICV050	VX046048.D	1,2,3-Trichloropropane	JOHN	5/6/2025 9:53:45 AM	MMDadoda	5/6/2025 12:41:37 PM	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

VX050525

Instrument

MSVOA_x

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

vx051925

Instrument

MSVOA_x

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
VSTDCCC050	VX046255.D	1,2,3-Trichloropropane	JOHN	5/21/2025 9:53:07 AM	MMDadoda	5/21/2025 3:45:24 PM	Peak Integrated by Software
VX0519WBS01	VX046258.D	1,2,3-Trichloropropane	JOHN	5/21/2025 9:53:11 AM	MMDadoda	5/21/2025 3:45:27 PM	Peak Integrated by Software
VX0519WBSD0 1	VX046259.D	1,2,3-Trichloropropane	JOHN	5/21/2025 9:53:16 AM	MMDadoda	5/21/2025 3:45:32 PM	Peak Integrated by Software
VSTDCCC050	VX046282.D	1,2,3-Trichloropropane	JOHN	5/21/2025 9:54:11 AM	MMDadoda	5/21/2025 3:46:14 PM	Peak Integrated by Software

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Maresh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk	VP133811		
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837		
CCC			
Internal Standard/PEM			
ICV/I.BLK	VP133838		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046038.D	05 May 2025 09:37	JC/MD	Ok
2	VSTDIC001	VX046039.D	05 May 2025 10:49	JC/MD	Not Ok
3	VSTDIC005	VX046040.D	05 May 2025 11:12	JC/MD	Not Ok
4	VSTDIC020	VX046041.D	05 May 2025 11:35	JC/MD	Ok,M
5	VSTDIC050	VX046042.D	05 May 2025 11:58	JC/MD	Ok,M
6	VSTDIC100	VX046043.D	05 May 2025 12:21	JC/MD	Ok,M
7	VSTDIC150	VX046044.D	05 May 2025 12:45	JC/MD	Ok,M
8	IBLK	VX046045.D	05 May 2025 13:08	JC/MD	Ok
9	VSTDIC005	VX046046.D	05 May 2025 16:04	JC/MD	Ok,M
10	VSTDIC001	VX046047.D	05 May 2025 16:27	JC/MD	Ok,M
11	VSTDICV050	VX046048.D	05 May 2025 16:50	JC/MD	Ok,M

M : Manual Integration

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QC Batch ID # VX051925

Review By	John Carlone	Review On	5/21/2025 9:57:32 AM
Supervise By	Maresh Dadoda	Supervise On	5/21/2025 3:46:34 PM
SubDirectory	VX051925	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133957		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133958,VP133959		

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VX046254.D	19 May 2025 09:33	JC/MD	Ok
2	VSTDCCC050	VX046255.D	19 May 2025 10:05	JC/MD	Ok,M
3	VX0519MBL01	VX046256.D	19 May 2025 10:39	JC/MD	Ok
4	VX0519WBL01	VX046257.D	19 May 2025 11:02	JC/MD	Ok
5	VX0519WBS01	VX046258.D	19 May 2025 11:25	JC/MD	Ok,M
6	VX0519WBSD01	VX046259.D	19 May 2025 11:53	JC/MD	Ok,M
7	Q2053-11	VX046260.D	19 May 2025 12:17	JC/MD	Not Ok
8	Q2052-04	VX046261.D	19 May 2025 12:40	JC/MD	Ok
9	Q2057-04	VX046262.D	19 May 2025 13:03	JC/MD	Ok
10	Q2062-04	VX046263.D	19 May 2025 13:27	JC/MD	Ok
11	Q2062-08	VX046264.D	19 May 2025 13:50	JC/MD	Ok
12	Q2062-12	VX046265.D	19 May 2025 14:14	JC/MD	Ok
13	Q2062-16	VX046266.D	19 May 2025 14:37	JC/MD	Ok
14	Q2062-20	VX046267.D	19 May 2025 15:00	JC/MD	Ok
15	Q2062-24	VX046268.D	19 May 2025 15:24	JC/MD	Ok
16	Q2053-11	VX046269.D	19 May 2025 15:47	JC/MD	Ok,M
17	Q2053-01	VX046270.D	19 May 2025 16:11	JC/MD	Ok,M
18	Q2053-02	VX046271.D	19 May 2025 16:34	JC/MD	Ok,M
19	Q2053-03	VX046272.D	19 May 2025 16:58	JC/MD	Ok
20	Q2053-04	VX046273.D	19 May 2025 17:21	JC/MD	Ok
21	Q2053-05	VX046274.D	19 May 2025 17:45	JC/MD	Ok

Instrument ID: **MSVOA_X**

Daily Analysis Runlog For Sequence/QCBatch ID # VX051925

Review By	John Carlone	Review On	5/21/2025 9:57:32 AM
Supervise By	Mahesh Dadoda	Supervise On	5/21/2025 3:46:34 PM
SubDirectory	VX051925	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133957		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133958,VP133959		

22	Q2053-06	VX046275.D	19 May 2025 18:08	JC/MD	Ok,M
23	Q2053-07	VX046276.D	19 May 2025 18:32	JC/MD	Ok,M
24	Q2053-08	VX046277.D	19 May 2025 18:55	JC/MD	Ok,M
25	Q2053-09	VX046278.D	19 May 2025 19:19	JC/MD	Ok,M
26	Q2053-10	VX046279.D	19 May 2025 19:42	JC/MD	Ok,M
27	Q2073-01	VX046280.D	19 May 2025 20:06	JC/MD	Ok
28	Q2073-02	VX046281.D	19 May 2025 20:29	JC/MD	Ok
29	VSTDCCC050	VX046282.D	19 May 2025 20:52	JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX050525

Review By	John Carlone	Review On	5/6/2025 9:53:58 AM
Supervise By	Mahesh Dadoda	Supervise On	5/6/2025 12:43:00 PM
SubDirectory	VX050525	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk	VP133811
Initial Calibration Stds	VP133832,VP133833,VP133834,VP133835,VP133836,VP133837
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP133838
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampled	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046038.D	05 May 2025 09:37		JC/MD	Ok
2	VSTDIC001	VSTDIC001	VX046039.D	05 May 2025 10:49	Not used	JC/MD	Not Ok
3	VSTDIC005	VSTDIC005	VX046040.D	05 May 2025 11:12	Not used	JC/MD	Not Ok
4	VSTDIC020	VSTDIC020	VX046041.D	05 May 2025 11:35		JC/MD	Ok,M
5	VSTDIC050	VSTDIC050	VX046042.D	05 May 2025 11:58		JC/MD	Ok,M
6	VSTDIC100	VSTDIC100	VX046043.D	05 May 2025 12:21		JC/MD	Ok,M
7	VSTDIC150	VSTDIC150	VX046044.D	05 May 2025 12:45		JC/MD	Ok,M
8	IBLK	IBLK	VX046045.D	05 May 2025 13:08		JC/MD	Ok
9	VSTDIC005	VSTDIC005	VX046046.D	05 May 2025 16:04		JC/MD	Ok,M
10	VSTDIC001	VSTDIC001	VX046047.D	05 May 2025 16:27		JC/MD	Ok,M
11	VSTDICV050	ICVVX050525	VX046048.D	05 May 2025 16:50		JC/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX051925

Review By	John Carlone	Review On	5/21/2025 9:57:32 AM
Supervise By	Maresh Dadoda	Supervise On	5/21/2025 3:46:34 PM
SubDirectory	VX051925	HP Acquire Method	HP Processing Method 82X050525W.M

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP133957
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133958,VP133959

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VX046254.D	19 May 2025 09:33		JC/MD	Ok
2	VSTDCCC050	VSTDCCC050	VX046255.D	19 May 2025 10:05	pH#Lot#V12668	JC/MD	Ok,M
3	VX0519MBL01	VX0519MBL01	VX046256.D	19 May 2025 10:39		JC/MD	Ok
4	VX0519WBL01	VX0519WBL01	VX046257.D	19 May 2025 11:02		JC/MD	Ok
5	VX0519WBS01	VX0519WBS01	VX046258.D	19 May 2025 11:25		JC/MD	Ok,M
6	VX0519WBSD01	VX0519WBSD01	VX046259.D	19 May 2025 11:53		JC/MD	Ok,M
7	Q2053-11	SB-11	VX046260.D	19 May 2025 12:17	Need Straight Run	JC/MD	Not Ok
8	Q2052-04	TP-B	VX046261.D	19 May 2025 12:40	vial A pH#5.0	JC/MD	Ok
9	Q2057-04	MH-L	VX046262.D	19 May 2025 13:03	vial A pH#5.0	JC/MD	Ok
10	Q2062-04	L1-WC-1	VX046263.D	19 May 2025 13:27	vial A pH#5.0	JC/MD	Ok
11	Q2062-08	L1-WC-2	VX046264.D	19 May 2025 13:50	vial A pH#5.0	JC/MD	Ok
12	Q2062-12	L1-WC-3	VX046265.D	19 May 2025 14:14	vial A pH#5.0	JC/MD	Ok
13	Q2062-16	L1-WC-4	VX046266.D	19 May 2025 14:37	vial A pH#5.0	JC/MD	Ok
14	Q2062-20	L1-WC-5	VX046267.D	19 May 2025 15:00	vial A pH#5.0	JC/MD	Ok
15	Q2062-24	L1-WC-6	VX046268.D	19 May 2025 15:24	vial A pH#5.0	JC/MD	Ok
16	Q2053-11	SB-11	VX046269.D	19 May 2025 15:47	vial A pH#5.0	JC/MD	Ok,M
17	Q2053-01	SB-1	VX046270.D	19 May 2025 16:11	vial A pH#5.0	JC/MD	Ok,M
18	Q2053-02	SB-2	VX046271.D	19 May 2025 16:34	vial A pH#5.0	JC/MD	Ok,M

Instrument ID: MSVOA_X

Daily Analysis Runlog For Sequence/QC Batch ID # VX051925

Review By	John Carlone	Review On	5/21/2025 9:57:32 AM
Supervise By	Maresh Dadoda	Supervise On	5/21/2025 3:46:34 PM
SubDirectory	VX051925	HP Acquire Method	HP Processing Method 82X050525W.M
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	VP133957		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133958,VP133959		

19	Q2053-03	SB-3	VX046272.D	19 May 2025 16:58	vial A pH#5.0	JC/MD	Ok
20	Q2053-04	SB-4	VX046273.D	19 May 2025 17:21	vial A pH#5.0	JC/MD	Ok
21	Q2053-05	SB-5	VX046274.D	19 May 2025 17:45	vial A pH#5.0	JC/MD	Ok
22	Q2053-06	SB-6	VX046275.D	19 May 2025 18:08	vial A pH#5.0	JC/MD	Ok,M
23	Q2053-07	SB-7	VX046276.D	19 May 2025 18:32	vial A pH#5.0	JC/MD	Ok,M
24	Q2053-08	SB-8	VX046277.D	19 May 2025 18:55	vial A pH#5.0	JC/MD	Ok,M
25	Q2053-09	SB-9	VX046278.D	19 May 2025 19:19	vial A pH#5.0	JC/MD	Ok,M
26	Q2053-10	SB-10	VX046279.D	19 May 2025 19:42	vial A pH#5.0	JC/MD	Ok,M
27	Q2073-01	GDW1	VX046280.D	19 May 2025 20:06	vial A pH<2	JC/MD	Ok
28	Q2073-02	GDW2	VX046281.D	19 May 2025 20:29	vial A pH<2	JC/MD	Ok
29	VSTDCCC050	VSTDCCC050EC	VX046282.D	19 May 2025 20:52		JC/MD	Ok,M

M : Manual Integration



SHIPPING DOCUMENTS

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: Environmental

ADDRESS: 8 CARRAGE

CITY: Success STATE: NY ZIP: 01876

ATTENTION:

PHONE: FAX:

PROJECT NAME: Nelson

PROJECT NO.: LOCATION:

PROJECT MANAGER: BL

e-mail:

PHONE: FAX:

BILL TO: Government PO#:

ADDRESS: 8 CARRAGE

CITY: Success STATE: NY ZIP:

ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

DATA DELIVERABLE INFORMATION

FAX (RUSH) Standard DAYS*

HARDCOPY (DATA PACKAGE) DAYS*

EDD: DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)

☐ Level 2 (Results + QC) ☒ NJ Reduced ☐ US EPA CLP

☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B

+ Raw Data) Other

☒ EDD FORMAT hardcopy, electronic

Self create box 15 (no pres)
Low level 15

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS
			COMP	GRAB	DATE	TIME		Hd	Re	Ice							
1.	GDW 1	GW	X		5/16/25	1200	4	X	X	X							
2.	GDW 2	GW	X		5/16/25	1315	4	X	X	X							
3.																	
4.																	
5.																	
6.																	
7.																	
8.																	
9.																	
10.																	

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☒ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>2.6</u> °C
1. <u>HT</u>	<u>5/16/25 1440</u>	1. <u>JT</u>	Comments:
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
2.		2.	
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3.		3.	

Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q2073	GENV01	Order Date : 5/16/2025 2:51:00 PM	Project Mgr :
Client Name : G Environmental		Project Name : Nelson	Report Type : Level 1 <i>NJ Reduce</i>
Client Contact : Gary Landis		Receive DateTime : 5/16/2025 2:40:00 PM	EDD Type : Excel NJ
Invoice Name : G Environmental		Purchase Order :	Hard Copy Date :
Invoice Contact : Gary Landis			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q2073-01	GDW1	Water	05/16/2025	12:00					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q2073-02	GDW2	Water	05/16/2025	13:15					
					VOCMS Group1		8260-Low	10 Bus. Days	

*Stored in 20A
Set # 04*

Relinquished By : *[Signature]*
Date / Time : 5/16/25 15:25

Received By : *[Signature]*
Date / Time : 5-16-25 15:25

Storage Area : VOA Refridgerator Room