

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG NO.: Q2073
Lab File ID: VX046255.D Date Analyzed: 05/19/2025
Instrument ID: MSVOA_X Time Analyzed: 10:05
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	97192	5.54	163845	6.76	142712	10.05
UPPER LIMIT	194384	6.038	327690	7.257	285424	10.549
LOWER LIMIT	48596	5.038	81922.5	6.257	71356	9.549
EPA SAMPLE NO.						
GDW1	62035	5.54	123496	6.76	115828	10.05
GDW2	61746	5.54	125249	6.76	113628	10.05
VX0519WBL01	62503	5.54	124684	6.76	117918	10.05
VX0519WBS01	91585	5.54	158835	6.76	142753	10.05
VX0519WBSD01	87263	5.54	150978	6.76	132315	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.