

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q2073 SAS No.: Q2073 SDG NO.: Q2073
 Lab File ID: VX046255.D Date Analyzed: 05/19/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:05
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	70270	12.018				
UPPER LIMIT	140540	12.518				
LOWER LIMIT	35135	11.518				
EPA SAMPLE NO.						
GDW1	49714	12.02				
GDW2	45883	12.02				
VX0519WBL01	48295	12.02				
VX0519WBS01	66879	12.02				
VX0519WBSD01	63585	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.