


LB135837

7471B

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000
 B= 1.9839e-004
 C= -3.3196e-004
 Rho= 0.9998888
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.013	-0.013	-65	0.000	-65					
0.2	0.200	0.174	-0.026	881	0.0 %	881					
2.5	2.500	2.478	-0.022	12491	0.0 %	12491					
5.0	5.000	5.121	0.121	25813	0.0 %	25813					
7.5	7.500	7.485	-0.015	37731	0.0 %	37731					
10.0	10.000	9.955	-0.045	50179	0.0 %	50179					

9/10
 73
 1
 2
 0
 0

LB135837 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 May 2025 12:40:19

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-65	-	0.0000	7471B	PPB	20 May 2025 12:43:40	S	Std
0.2 - 1	50.2	881	-	0.2000	7471B	PPB	20 May 2025 12:45:57	S	Std
2.5 - 1	52.5	12491	-	2.5000	7471B	PPB	20 May 2025 12:48:15	S	Std
5.0 - 1	55	25813	-	5.0000	7471B	PPB	20 May 2025 12:50:32	S	Std
7.5 - 1	55.5	37731	-	7.5000	7471B	PPB	20 May 2025 12:52:51	S	Std
10.0 - 1	56	50179	-	10.0000	7471B	PPB	20 May 2025 12:55:12	S	Std
ICV12 - 1	ICV12	20138	3.9949	-	7471B	PPB	20 May 2025 12:58:31	U	SMPL
ICB12 - 1	ICB12	-278	-0.0555	-	7471B	PPB	20 May 2025 13:00:46	U	SMPL
CCV45 - 1	CCV45	24685	4.8970	-	7471B	PPB	20 May 2025 13:03:04	U	SMPL
CCB45 - 1	CCB45	-327	-0.0652	-	7471B	PPB	20 May 2025 13:05:19	U	SMPL
CRA - 1	CRA	937	0.1856	-	7471B	PPB	20 May 2025 13:07:39	U	SMPL
HighStd - 1	HighStd	49997	9.9188	-	7471B	PPB	20 May 2025 13:09:54	U	SMPL
ChkStd - 1	ChkStd	34271	6.7988	-	7471B	PPB	20 May 2025 13:12:09	U	SMPL
PB168089BL - 1	PBS	-84	-0.0170	-	7471B	PPB	20 May 2025 13:17:23	U	SMPL
PB168089BS - 1	LCSS	20185	4.0043	-	7471B	PPB	20 May 2025 13:19:38	U	SMPL
Q2052-01 - 1	TP-B	5530	1.0968	-	7471B	PPB	20 May 2025 13:21:55	U	SMPL
Q2052-01DUP - 1	TP-BDUP	4819	0.9557	-	7471B	PPB	20 May 2025 13:24:14	U	SMPL
Q2052-01MS - 1	TP-BMS	27069	5.3700	-	7471B	PPB	20 May 2025 13:26:34	U	SMPL
Q2052-01MSD - 1	TP-BMSD	25608	5.0801	-	7471B	PPB	20 May 2025 13:28:52	U	SMPL
Q2055-01 - 1	OR-02-051525	533	0.1054	-	7471B	PPB	20 May 2025 13:31:13	U	SMPL
CCV46 - 1	CCV46	23846	4.7306	-	7471B	PPB	20 May 2025 13:33:36	U	SMPL
CCB46 - 1	CCB46	-384	-0.0765	-	7471B	PPB	20 May 2025 13:35:51	U	SMPL
Q2057-01 - 1	MH-L	330	0.0651	-	7471B	PPB	20 May 2025 13:38:12	U	SMPL
Q2062-01 - 1	L1-WC-1	1329	0.2633	-	7471B	PPB	20 May 2025 13:40:27	U	SMPL
Q2062-05 - 1	L1-WC-2	2213	0.4387	-	7471B	PPB	20 May 2025 13:42:43	U	SMPL
Q2062-09 - 1	L1-WC-3	5568	1.1043	-	7471B	PPB	20 May 2025 13:45:03	U	SMPL
Q2062-13 - 1	L1-WC-4	1155	0.2288	-	7471B	PPB	20 May 2025 13:47:20	U	SMPL
Q2062-17 - 1	L1-WC-5	2931	0.5812	-	7471B	PPB	20 May 2025 13:49:41	U	SMPL
Q2062-21 - 1	L1-WC-6	6767	1.3422	-	7471B	PPB	20 May 2025 13:51:57	U	SMPL
Q2068-01 - 1	LAW-25-0075	424	0.0838	-	7471B	PPB	20 May 2025 13:54:15	U	SMPL
PB168090BL - 1	PB168090BL	-106	-0.0214	-	7471B	PPB	20 May 2025 13:56:35	U	SMPL
PB168090BS - 1	PB168090BS	20217	4.0106	-	7471B	PPB	20 May 2025 13:58:53	U	SMPL
CCV47 - 1	CCV47	24196	4.8000	-	7471B	PPB	20 May 2025 14:01:09	U	SMPL
CCB47 - 1	CCB47	-364	-0.0725	-	7471B	PPB	20 May 2025 14:03:28	U	SMPL
Q2071-01 - 1	L1-WC-7	782	0.1548	-	7471B	PPB	20 May 2025 14:05:47	U	SMPL
Q2071-05 - 1	L1-WC-8	4600	0.9123	-	7471B	PPB	20 May 2025 14:08:06	U	SMPL
Q2071-09 - 1	L1-WC-9	4196	0.8321	-	7471B	PPB	20 May 2025 14:10:37	U	SMPL
Q2071-13 - 1	L2-WC-5	2509	0.4974	-	7471B	PPB	20 May 2025 14:13:02	U	SMPL
Q2071-17 - 1	L2-WC-6	1998	0.3961	-	7471B	PPB	20 May 2025 14:15:25	U	SMPL
Q2074-01 - 1	TP-12	974	0.1929	-	7471B	PPB	20 May 2025 14:17:47	U	SMPL
Q2074-02 - 1	TP-7	729	0.1443	-	7471B	PPB	20 May 2025 14:20:07	U	SMPL
Q2074-03 - 1	TP-15	4097	0.8125	-	7471B	PPB	20 May 2025 14:22:24	U	SMPL
Q2074-04 - 1	TP-20	2399	0.4756	-	7471B	PPB	20 May 2025 14:24:45	U	SMPL
Q2074-05 - 1	TP-38	64	0.0124	-	7471B	PPB	20 May 2025 14:27:02	U	SMPL
CCV48 - 1	CCV48	24723	4.9046	-	7471B	PPB	20 May 2025 14:29:18	U	SMPL
CCB48 - 1	CCB48	-327	-0.0652	-	7471B	PPB	20 May 2025 14:31:34	U	SMPL
Q2074-06 - 1	TP-19	3355	0.6653	-	7471B	PPB	20 May 2025 14:33:54	U	SMPL
Q2074-07 - 1	TP-40	2102	0.4167	-	7471B	PPB	20 May 2025 14:36:10	U	SMPL
Q2074-08 - 1	TP-18	1838	0.3643	-	7471B	PPB	20 May 2025 14:38:27	U	SMPL
Q2074-08DUP - 1	TP-18DUP	1508	0.2988	-	7471B	PPB	20 May 2025 14:40:43	U	SMPL
Q2074-08MS - 1	TP-18MS	18439	3.6579	-	7471B	PPB	20 May 2025 14:42:59	U	SMPL
Q2074-08MSD - 1	TP-18MSD	21097	4.1852	-	7471B	PPB	20 May 2025 14:45:16	U	SMPL
Q2080-01 - 1	PL-HRH-COMP-01	3497	0.6935	-	7471B	PPB	20 May 2025 14:47:35	U	SMPL
Q2080-07 - 1	PL-HRH-COMP-02	4176	0.8282	-	7471B	PPB	20 May 2025 14:49:55	U	SMPL
CCV49 - 1	CCV49	26028	5.1635	-	7471B	PPB	20 May 2025 14:52:13	U	SMPL
CCB49 - 1	CCB49	-364	-0.0725	-	7471B	PPB	20 May 2025 14:54:30	U	SMPL
Q2052-01LX5 - 1		1114	0.2207	-	7471B	PPB	20 May 2025 14:56:50	U	SMPL
Q2052-01A - 1		25803	5.1188	-	7471B	PPB	20 May 2025 14:59:05	U	SMPL
Q2074-08LX5 - 1		25	0.0046	-	7471B	PPB	20 May 2025 15:01:21	U	SMPL
Q2074-08A - 1		21061	4.1780	-	7471B	PPB	20 May 2025 15:03:41	U	SMPL
CCV50 - 1	CCV50	25892	5.1365	-	7471B	PPB	20 May 2025 15:05:57	U	SMPL
CCB50 - 1	CCB50	-301	-0.0600	-	7471B	PPB	20 May 2025 15:08:18	U	SMPL