

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2074
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2074 **SAS No.:** Q2074

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168087BS							
Aluminum	mg/Kg	100	90.2		90	80 - 120	P
Antimony	mg/Kg	40.0	38.5		96	80 - 120	P
Arsenic	mg/Kg	40.0	38.6		96	80 - 120	P
Barium	mg/Kg	10.0	8.52		85	80 - 120	P
Beryllium	mg/Kg	10.0	9.28		93	80 - 120	P
Cadmium	mg/Kg	10.0	9.65		96	80 - 120	P
Calcium	mg/Kg	50.0	45.1	J	90	80 - 120	P
Chromium	mg/Kg	20.0	19.0		95	80 - 120	P
Cobalt	mg/Kg	10.0	9.30		93	80 - 120	P
Copper	mg/Kg	15.0	14.6		97	80 - 120	P
Iron	mg/Kg	150	140		93	80 - 120	P
Lead	mg/Kg	50.0	46.6		93	80 - 120	P
Magnesium	mg/Kg	100	90.0	J	90	80 - 120	P
Manganese	mg/Kg	10.0	9.14		91	80 - 120	P
Nickel	mg/Kg	25.0	23.5		94	80 - 120	P
Potassium	mg/Kg	500	439		88	80 - 120	P
Selenium	mg/Kg	100	97.9		98	80 - 120	P
Silver	mg/Kg	3.8	3.54		93	80 - 120	P
Sodium	mg/Kg	150	137		91	80 - 120	P
Thallium	mg/Kg	100	94.2		94	80 - 120	P
Vanadium	mg/Kg	15.0	13.7		91	80 - 120	P
Zinc	mg/Kg	10.0	9.45		94	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: CDM Smith **SDG No.:** Q2074
Contract: CAMP02 **Lab Code:** CHEM **Case No.:** Q2074 **SAS No.:** Q2074

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168090BS Mercury	mg/Kg	0.26	0.26		100	80 - 120	CV