

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: CAMP02
 ProjectID: South River WM Replacement
 Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG No.: Q2074

Calibration Sequence : FF052125		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	131501360	131501	FF015874.D
500	66823160	133646	FF015875.D
200	25639665	128198	FF015876.D
100	14440283	144403	FF015877.D
50	6491363	129827	FF015878.D
AVG RF : 133515		% RSD : 4.803	AVG RT : 15.0206