

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: CAMP02
ProjectID: South River WM Replacement
Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG No.: Q2074
DataFile: FG015865.D Analyst Name: YP\AJ Analyst Date: 05-21-2025

Conc. (PPM)	Area Count	RF	Average RF	%D
500	70332673	140665	126925	10.825