

SOIL DIESEL RANGE ORGANICS LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RI

Lab Name: Chemtech **Client:** CDM Smith
Lab Code: CHEM **Cas No:** Q2074 **SAS No :** Q2074 **SDG No:** Q2074
Matrix Spike - EPA Sample No : PB168084BS **Datafile:** FG015868.D

COMPOUND	SPIKE ADDED ug/kg	CONCENTRATION ug/kg	LCS/LCSD CONCENTRATION ug/kg	% REC	QC LIMITS
DRO	6660	0	6529	98	68-131