

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB051925

Review By	yogesh	Review On	5/19/2025 11:36:03 AM
Supervise By	mohammad	Supervise On	5/21/2025 5:31:15 AM
SubDirectory	FB051925	HP Acquire Method	HP Processing Method
			FB051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24585,PP24586,PP24587,PP24588,PP24589		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24590		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB031773.D	19 May 2025 9:07		YP/AJ	Ok
2	10 GRO STD		FB031774.D	19 May 2025 9:35		YP/AJ	Ok,M
3	20 GRO STD		FB031775.D	19 May 2025 10:03		YP/AJ	Ok
4	50 GRO STD		FB031776.D	19 May 2025 10:31		YP/AJ	Ok,M
5	100 GRO STD		FB031777.D	19 May 2025 10:58		YP/AJ	Ok,M
6	FB051925GROICV		FB031778.D	19 May 2025 11:41		YP/AJ	Ok,M

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB052025

Review By	yogesh	Review On	5/20/2025 12:45:48 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:39:33 AM
SubDirectory	FB052025	HP Acquire Method	HP Processing Method
			FB051925

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24110,PP24585,PP24586,PP24587,PP24588,PP24589
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24592,PP24593,PP24594 PP24111,PP24590

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031779.D	20 May 2025 9:17		YP/AJ	Ok,M
2	VBF0520S1		FB031780.D	20 May 2025 10:00		YP/AJ	Ok
3	VBF0520S2		FB031781.D	20 May 2025 10:28		YP/AJ	Ok
4	BSF0520S1		FB031782.D	20 May 2025 10:55		YP/AJ	Ok
5	Q2074-01		FB031783.D	20 May 2025 11:23	Vial-A	YP/AJ	Ok
6	Q2074-02		FB031784.D	20 May 2025 11:50	Vial-A	YP/AJ	Ok
7	Q2074-03		FB031785.D	20 May 2025 12:27	Vial-A	YP/AJ	Ok
8	Q2074-04		FB031786.D	20 May 2025 13:11	Vial-A	YP/AJ	Ok
9	Q2074-05		FB031787.D	20 May 2025 13:38	Vial-A, Not purged	YP/AJ	Not Ok
10	Q2074-06		FB031788.D	20 May 2025 14:06	Vial-A	YP/AJ	Ok
11	BSF0520S2		FB031789.D	20 May 2025 14:33		YP/AJ	Ok,M
12	20 PPB GRO STD		FB031790.D	20 May 2025 15:01		YP/AJ	Ok,M
13	Q2074-07		FB031791.D	20 May 2025 15:56	Vial-A End CCC Missing, Instrument Stopped	YP/AJ	Not Ok
14	Q2074-08		FB031792.D	20 May 2025 16:31	Vial-A End CCC Missing, Instrument Stopped	YP/AJ	Not Ok
15	Q2074-05		FB031793.D	20 May 2025 16:59	Vial-B End CCC Missing, Instrument Stopped	YP/AJ	Not Ok
16	Q2074-07		FB031794.D	20 May 2025 17:27	Vial-B End CCC Missing, Instrument Stopped	YP/AJ	Not Ok
17	Q2074-08		FB031795.D	20 May 2025 17:54	Vial-B End CCC Missing, Instrument Stopped	YP/AJ	Not Ok

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB052025

Review By	yogesh	Review On	5/20/2025 12:45:48 PM
Supervise By	mohammad	Supervise On	5/22/2025 6:39:33 AM
SubDirectory	FB052025	HP Acquire Method	HP Processing Method FB051925

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24110,PP24585,PP24586,PP24587,PP24588,PP24589
CCC Internal Standard/PEM	PP24592,PP24593,PP24594
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24590

18	Q2074-05		FB031796.D	20 May 2025 18:22	Vial-C End CCC Missing, Instrument Stopped	YP/AJ	Not Ok
----	----------	--	------------	-------------------	---	-------	--------

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB052125

Review By	yogesh	Review On	5/21/2025 2:47:29 PM
Supervise By	mohammad	Supervise On	5/26/2025 2:21:45 AM
SubDirectory	FB052125	HP Acquire Method	HP Processing Method
			FB051925
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24585,PP24586,PP24587,PP24588,PP24589		
CCC Internal Standard/PEM	PP24598,PP24599,PP24600		
ICV/I.BLK Surrogate Standard	PP24111,PP24590		
MS/MSD Standard			
LCS Standard			

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB031797.D	21 May 2025 9:04		YP/AJ	Ok,M
2	VBF0521S1		FB031798.D	21 May 2025 9:44		YP/AJ	Ok
3	VBF0521S2		FB031799.D	21 May 2025 10:12		YP/AJ	Ok
4	BSF0521S1		FB031800.D	21 May 2025 10:40		YP/AJ	Ok,M
5	Q2074-05		FB031801.D	21 May 2025 11:08	Vial-C	YP/AJ	Ok
6	Q2074-07		FB031802.D	21 May 2025 11:35	Vial-C	YP/AJ	Ok
7	Q2074-08		FB031803.D	21 May 2025 12:12	Vial-C	YP/AJ	Ok
8	BSF0521S2		FB031804.D	21 May 2025 12:39		YP/AJ	Ok,M
9	20 PPB GRO STD		FB031805.D	21 May 2025 13:23		YP/AJ	Ok
10	Q2100-01		FB031806.D	21 May 2025 14:34	Vial-A	YP/AJ	Ok
11	Q2100-02		FB031807.D	21 May 2025 15:01	Vial-A	YP/AJ	Ok
12	Q2100-03		FB031808.D	21 May 2025 15:29	Vial-A	YP/AJ	Ok
13	Q2100-04		FB031809.D	21 May 2025 15:57	Vial-A	YP/AJ	Ok
14	BSF0521S3		FB031810.D	21 May 2025 16:25		YP/AJ	Ok,M
15	20 PPB GRO STD		FB031811.D	21 May 2025 16:52		YP/AJ	Ok,M

M : Manual Integration