

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/21/2025

Lab Sample No.: HSTDCCC750 Data File : PS030314.D Time Analyzed: 15:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.129	9.033	9.233	712.280	712.500	0.0
2,4,5-TP(Silvex)	8.847	8.750	8.950	712.570	712.500	0.0
2,4-D	8.006	7.910	8.110	697.640	705.000	-1.0
2,4-DB	9.687	9.589	9.789	755.280	712.500	6.0
2,4-DCAA	6.929	6.834	7.034	726.170	750.000	-3.2
DICAMBA	7.106	7.010	7.210	703.230	705.000	-0.3
DICHLORPROP	7.786	7.689	7.889	684.540	705.000	-2.9
Dinoseb	10.848	10.751	10.951	695.150	705.000	-1.4

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL01 Date Analyzed: 05/21/2025

Lab Sample No.: HSTDCCC750 Data File : PS030314.D Time Analyzed: 15:25

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.913	9.826	10.026	684.310	712.500	-4.0
2,4,5-TP(Silvex)	9.509	9.423	9.623	690.980	712.500	-3.0
2,4-D	8.632	8.544	8.744	672.740	705.000	-4.6
2,4-DB	10.471	10.385	10.585	624.030	712.500	-12.4
2,4-DCAA	7.442	7.354	7.554	692.280	750.000	-7.7
DICAMBA	7.627	7.539	7.739	678.490	705.000	-3.8
DICHLORPROP	8.319	8.232	8.432	652.460	705.000	-7.5
Dinoseb	10.844	10.758	10.958	672.210	705.000	-4.7