

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/21/2025

Lab Sample No.: HSTDCCC750 Data File : PS030324.D Time Analyzed: 20:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.127	9.033	9.233	711.830	712.500	-0.1
2,4,5-TP(Silvex)	8.845	8.750	8.950	708.190	712.500	-0.6
2,4-D	8.005	7.910	8.110	695.040	705.000	-1.4
2,4-DB	9.684	9.589	9.789	759.510	712.500	6.6
2,4-DCAA	6.929	6.834	7.034	728.350	750.000	-2.9
DICAMBA	7.105	7.010	7.210	701.780	705.000	-0.5
DICHLORPROP	7.785	7.689	7.889	686.830	705.000	-2.6
Dinoseb	10.844	10.751	10.951	690.500	705.000	-2.1

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL02 Date Analyzed: 05/21/2025

Lab Sample No.: HSTDCCC750 Data File : PS030324.D Time Analyzed: 20:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.918	9.826	10.026	697.960	712.500	-2.0
2,4,5-TP(Silvex)	9.513	9.423	9.623	702.560	712.500	-1.4
2,4-D	8.636	8.544	8.744	688.830	705.000	-2.3
2,4-DB	10.474	10.385	10.585	634.390	712.500	-11.0
2,4-DCAA	7.447	7.354	7.554	713.880	750.000	-4.8
DICAMBA	7.633	7.539	7.739	690.690	705.000	-2.0
DICHLORPROP	8.324	8.232	8.432	666.080	705.000	-5.5
Dinoseb	10.848	10.758	10.958	683.340	705.000	-3.1