

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS030334.D Time Analyzed: 01:45

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-T	9.128	9.033	9.233	704.880	712.500	-1.1
2,4,5-TP(Silvex)	8.846	8.750	8.950	700.620	712.500	-1.7
2,4-D	8.005	7.910	8.110	686.810	705.000	-2.6
2,4-DB	9.685	9.589	9.789	746.410	712.500	4.8
2,4-DCAA	6.930	6.834	7.034	721.880	750.000	-3.7
DICAMBA	7.106	7.010	7.210	694.030	705.000	-1.6
DICHLORPROP	7.786	7.689	7.889	681.520	705.000	-3.3
Dinoseb	10.845	10.751	10.951	686.410	705.000	-2.6

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL03 Date Analyzed: 05/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS030334.D Time Analyzed: 01:45

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.918	9.826	10.026	690.390	712.500	-3.1
2,4,5-TP(Silvex)	9.514	9.423	9.623	696.770	712.500	-2.2
2,4-D	8.636	8.544	8.744	681.090	705.000	-3.4
2,4-DB	10.475	10.385	10.585	633.280	712.500	-11.1
2,4-DCAA	7.448	7.354	7.554	710.320	750.000	-5.3
DICAMBA	7.633	7.539	7.739	685.810	705.000	-2.7
DICHLORPROP	8.325	8.232	8.432	662.290	705.000	-6.1
Dinoseb	10.849	10.758	10.958	676.960	705.000	-4.0