

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS030359.D Time Analyzed: 20:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.126	9.033	9.233	725.300	712.500	1.8
2,4,5-TP(Silvex)	8.844	8.750	8.950	719.260	712.500	0.9
2,4-D	8.004	7.910	8.110	706.710	705.000	0.2
2,4-DB	9.683	9.589	9.789	763.910	712.500	7.2
2,4-DCAA	6.929	6.834	7.034	749.220	750.000	-0.1
DICAMBA	7.104	7.010	7.210	713.660	705.000	1.2
DICHLORPROP	7.784	7.689	7.889	706.330	705.000	0.2
Dinoseb	10.843	10.751	10.951	700.210	705.000	-0.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL04 Date Analyzed: 05/22/2025

Lab Sample No.: HSTDCCC750 Data File : PS030359.D Time Analyzed: 20:29

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-T	9.915	9.826	10.026	710.130	712.500	-0.3
2,4,5-TP(Silvex)	9.511	9.423	9.623	717.350	712.500	0.7
2,4-D	8.634	8.544	8.744	701.130	705.000	-0.5
2,4-DB	10.472	10.385	10.585	621.450	712.500	-12.8
2,4-DCAA	7.446	7.354	7.554	730.330	750.000	-2.6
DICAMBA	7.631	7.539	7.739	704.850	705.000	0.0
DICHLORPROP	8.322	8.232	8.432	681.030	705.000	-3.4
Dinoseb	10.846	10.758	10.958	679.350	705.000	-3.6