

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL05 Date Analyzed: 05/23/2025

Lab Sample No.: HSTDCCC750 Data File : PS030369.D Time Analyzed: 00:54

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D  |
|------------------|--------|-----------|--------|--------------------|-------------------|-----|
|                  |        | FROM      | TO     |                    |                   |     |
| 2,4,5-T          | 9.126  | 9.033     | 9.233  | 734.260            | 712.500           | 3.1 |
| 2,4,5-TP(Silvex) | 8.844  | 8.750     | 8.950  | 724.900            | 712.500           | 1.7 |
| 2,4-D            | 8.004  | 7.910     | 8.110  | 714.890            | 705.000           | 1.4 |
| 2,4-DB           | 9.683  | 9.589     | 9.789  | 778.330            | 712.500           | 9.2 |
| 2,4-DCAA         | 6.928  | 6.834     | 7.034  | 757.680            | 750.000           | 1.0 |
| DICAMBA          | 7.104  | 7.010     | 7.210  | 721.180            | 705.000           | 2.3 |
| DICHLORPROP      | 7.784  | 7.689     | 7.889  | 710.980            | 705.000           | 0.8 |
| Dinoseb          | 10.843 | 10.751    | 10.951 | 706.700            | 705.000           | 0.2 |

### CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/12/2025 05/12/2025

Client Sample No.: CCAL05 Date Analyzed: 05/23/2025

Lab Sample No.: HSTDCCC750 Data File : PS030369.D Time Analyzed: 00:54

| COMPOUND         | RT     | RT WINDOW |        | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|------------------|--------|-----------|--------|--------------------|-------------------|-------|
|                  |        | FROM      | TO     |                    |                   |       |
| 2,4,5-T          | 9.915  | 9.826     | 10.026 | 724.630            | 712.500           | 1.7   |
| 2,4,5-TP(Silvex) | 9.512  | 9.423     | 9.623  | 729.650            | 712.500           | 2.4   |
| 2,4-D            | 8.634  | 8.544     | 8.744  | 712.970            | 705.000           | 1.1   |
| 2,4-DB           | 10.473 | 10.385    | 10.585 | 623.400            | 712.500           | -12.5 |
| 2,4-DCAA         | 7.446  | 7.354     | 7.554  | 738.500            | 750.000           | -1.5  |
| DICAMBA          | 7.631  | 7.539     | 7.739  | 716.590            | 705.000           | 1.6   |
| DICHLORPROP      | 8.323  | 8.232     | 8.432  | 691.600            | 705.000           | -1.9  |
| Dinoseb          | 10.846 | 10.758    | 10.958 | 703.860            | 705.000           | -0.2  |