

**SW-846**

Datafile : PS030316.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits	
								Qual	Low	High
PB168097BS (Column 1)	DICAMBA	166.6	169	ug/Kg	101				72	129
	DICHLORPROP	166.6	167	ug/Kg	100				77	135
	2,4-D	166.6	168	ug/Kg	101				65	144
	2,4,5-TP(Silvex)	166.6	171	ug/Kg	103				74	146
	2,4,5-T	166.6	172	ug/Kg	103				77	134
	2,4-DB	166.6	178	ug/Kg	107				72	122
	Dinoseb	166.6	169	ug/Kg	101				74	132
PB168097BS (Column 2)	DICAMBA	166.6	159	ug/Kg	95				72	129
	DICHLORPROP	166.6	157	ug/Kg	94				77	135
	2,4-D	166.6	161	ug/Kg	97				65	144
	2,4,5-TP(Silvex)	166.6	164	ug/Kg	98				74	146
	2,4,5-T	166.6	163	ug/Kg	98				77	134
	2,4-DB	166.6	140	ug/Kg	84				72	122
	Dinoseb	166.6	160	ug/Kg	96				74	132