

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP072203.D Time Analyzed: 09:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.654	5.552	5.752	507.610	500.000	1.5
Aroclor-1016-2	5.676	5.574	5.774	481.910	500.000	-3.6
Aroclor-1016-3	5.738	5.636	5.836	466.920	500.000	-6.6
Aroclor-1016-4	5.835	5.734	5.934	484.600	500.000	-3.1
Aroclor-1016-5	6.128	6.027	6.227	488.260	500.000	-2.3
Aroclor-1260-1	7.247	7.147	7.347	486.230	500.000	-2.8
Aroclor-1260-2	7.501	7.401	7.601	484.730	500.000	-3.1
Aroclor-1260-3	7.860	7.760	7.960	496.750	500.000	-0.7
Aroclor-1260-4	8.084	7.985	8.185	497.860	500.000	-0.4
Aroclor-1260-5	8.403	8.303	8.503	493.650	500.000	-1.3
Decachlorobiphenyl	10.212	10.113	10.313	49.190	50.000	-1.6
Tetrachloro-m-xylene	4.500	4.396	4.596	49.670	50.000	-0.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP072203.D Time Analyzed: 09:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.880	4.777	4.977	478.800	500.000	-4.2
Aroclor-1016-2	4.897	4.795	4.995	485.950	500.000	-2.8
Aroclor-1016-3	5.075	4.973	5.173	496.170	500.000	-0.8
Aroclor-1016-4	5.116	5.015	5.215	500.020	500.000	0.0
Aroclor-1016-5	5.331	5.229	5.429	499.920	500.000	0.0
Aroclor-1260-1	6.364	6.264	6.464	498.530	500.000	-0.3
Aroclor-1260-2	6.553	6.453	6.653	476.460	500.000	-4.7
Aroclor-1260-3	6.706	6.606	6.806	514.160	500.000	2.8
Aroclor-1260-4	7.176	7.077	7.277	508.110	500.000	1.6
Aroclor-1260-5	7.418	7.319	7.519	499.410	500.000	-0.1
Decachlorobiphenyl	8.822	8.723	8.923	49.120	50.000	-1.8
Tetrachloro-m-xylene	3.796	3.691	3.891	48.100	50.000	-3.8