

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP072218.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.655	5.552	5.752	459.820	500.000	-8.0
Aroclor-1016-2	5.676	5.574	5.774	496.290	500.000	-0.7
Aroclor-1016-3	5.739	5.636	5.836	491.340	500.000	-1.7
Aroclor-1016-4	5.836	5.734	5.934	502.260	500.000	0.5
Aroclor-1016-5	6.129	6.027	6.227	507.950	500.000	1.6
Aroclor-1260-1	7.248	7.147	7.347	508.540	500.000	1.7
Aroclor-1260-2	7.502	7.401	7.601	495.550	500.000	-0.9
Aroclor-1260-3	7.860	7.760	7.960	511.200	500.000	2.2
Aroclor-1260-4	8.084	7.985	8.185	514.370	500.000	2.9
Aroclor-1260-5	8.403	8.303	8.503	512.480	500.000	2.5
Decachlorobiphenyl	10.212	10.113	10.313	50.530	50.000	1.1
Tetrachloro-m-xylene	4.502	4.396	4.596	50.020	50.000	0.0

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP072218.D Time Analyzed: 17:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.881	4.777	4.977	491.790	500.000	-1.6
Aroclor-1016-2	4.899	4.795	4.995	499.780	500.000	0.0
Aroclor-1016-3	5.076	4.973	5.173	505.010	500.000	1.0
Aroclor-1016-4	5.118	5.015	5.215	506.270	500.000	1.3
Aroclor-1016-5	5.332	5.229	5.429	519.600	500.000	3.9
Aroclor-1260-1	6.365	6.264	6.464	502.580	500.000	0.5
Aroclor-1260-2	6.554	6.453	6.653	473.990	500.000	-5.2
Aroclor-1260-3	6.707	6.606	6.806	508.080	500.000	1.6
Aroclor-1260-4	7.178	7.077	7.277	505.110	500.000	1.0
Aroclor-1260-5	7.419	7.319	7.519	497.310	500.000	-0.5
Decachlorobiphenyl	8.823	8.723	8.923	49.180	50.000	-1.6
Tetrachloro-m-xylene	3.798	3.691	3.891	50.050	50.000	0.1