

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP072233.D Time Analyzed: 22:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.655	5.552	5.752	456.700	500.000	-8.7
Aroclor-1016-2	5.676	5.574	5.774	491.930	500.000	-1.6
Aroclor-1016-3	5.739	5.636	5.836	480.640	500.000	-3.9
Aroclor-1016-4	5.836	5.734	5.934	496.140	500.000	-0.8
Aroclor-1016-5	6.129	6.027	6.227	500.200	500.000	0.0
Aroclor-1260-1	7.247	7.147	7.347	500.850	500.000	0.2
Aroclor-1260-2	7.501	7.401	7.601	487.730	500.000	-2.5
Aroclor-1260-3	7.859	7.760	7.960	505.400	500.000	1.1
Aroclor-1260-4	8.084	7.985	8.185	510.290	500.000	2.1
Aroclor-1260-5	8.403	8.303	8.503	496.080	500.000	-0.8
Decachlorobiphenyl	10.209	10.113	10.313	49.820	50.000	-0.4
Tetrachloro-m-xylene	4.501	4.396	4.596	49.740	50.000	-0.5

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL03 Date Analyzed: 05/20/2025

Lab Sample No.: AR1660CCC500 Data File : PP072233.D Time Analyzed: 22:48

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.880	4.777	4.977	494.880	500.000	-1.0
Aroclor-1016-2	4.898	4.795	4.995	502.110	500.000	0.4
Aroclor-1016-3	5.075	4.973	5.173	512.910	500.000	2.6
Aroclor-1016-4	5.116	5.015	5.215	513.720	500.000	2.7
Aroclor-1016-5	5.331	5.229	5.429	539.830	500.000	8.0
Aroclor-1260-1	6.364	6.264	6.464	518.550	500.000	3.7
Aroclor-1260-2	6.553	6.453	6.653	487.150	500.000	-2.6
Aroclor-1260-3	6.705	6.606	6.806	519.560	500.000	3.9
Aroclor-1260-4	7.176	7.077	7.277	514.650	500.000	2.9
Aroclor-1260-5	7.417	7.319	7.519	507.790	500.000	1.6
Decachlorobiphenyl	8.820	8.723	8.923	51.090	50.000	2.2
Tetrachloro-m-xylene	3.796	3.691	3.891	49.960	50.000	-0.1