

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/20/2025

Lab Sample No.: PSTDCCC050 Data File : PD088658.D Time Analyzed: 16:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.707	6.608	6.808	53.280	50.000	6.6
4,4'-DDE	6.197	6.098	6.298	50.230	50.000	0.5
4,4'-DDT	7.022	6.923	7.123	41.180	50.000	-17.6
Aldrin	5.273	5.173	5.373	51.710	50.000	3.4
alpha-BHC	4.001	3.900	4.100	52.930	50.000	5.9
alpha-Chlordane	6.029	5.930	6.130	50.580	50.000	1.2
beta-BHC	4.517	4.417	4.617	50.950	50.000	1.9
Decachlorobiphenyl	9.077	8.977	9.177	52.040	50.000	4.1
delta-BHC	4.765	4.666	4.866	54.910	50.000	9.8
Dieldrin	6.349	6.250	6.450	50.690	50.000	1.4
Endosulfan I	6.076	5.977	6.177	50.220	50.000	0.4
Endosulfan II	6.788	6.688	6.888	48.520	50.000	-3.0
Endosulfan sulfate	7.151	7.052	7.252	49.860	50.000	-0.3
Endrin	6.576	6.477	6.677	50.030	50.000	0.1
Endrin aldehyde	6.917	6.817	7.017	46.700	50.000	-6.6
Endrin ketone	7.633	7.533	7.733	49.030	50.000	-1.9
gamma-BHC (Lindane)	4.332	4.232	4.432	52.290	50.000	4.6
gamma-Chlordane	5.948	5.848	6.048	50.900	50.000	1.8
Heptachlor	4.931	4.832	5.032	50.650	50.000	1.3
Heptachlor epoxide	5.693	5.593	5.793	50.950	50.000	1.9
Methoxychlor	7.495	7.395	7.595	41.930	50.000	-16.1
Tetrachloro-m-xylene	3.551	3.451	3.651	56.870	50.000	13.7

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL01 Date Analyzed: 05/20/2025

Lab Sample No.: PSTDCCC050 Data File : PD088658.D Time Analyzed: 16:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.932	5.833	6.033	50.370	50.000	0.7
4,4'-DDE	5.375	5.278	5.478	48.790	50.000	-2.4
4,4'-DDT	6.185	6.087	6.287	43.570	50.000	-12.9
Aldrin	4.370	4.271	4.471	51.230	50.000	2.5
alpha-BHC	3.394	3.294	3.494	51.270	50.000	2.5
alpha-Chlordane	5.192	5.094	5.294	49.000	50.000	-2.0
beta-BHC	4.027	3.927	4.127	50.530	50.000	1.1
Decachlorobiphenyl	8.075	7.977	8.177	48.390	50.000	-3.2
delta-BHC	4.264	4.164	4.364	50.860	50.000	1.7
Dieldrin	5.514	5.416	5.616	48.540	50.000	-2.9
Endosulfan I	5.248	5.150	5.350	46.230	50.000	-7.5
Endosulfan II	6.082	5.984	6.184	48.690	50.000	-2.6
Endosulfan sulfate	6.484	6.386	6.586	47.050	50.000	-5.9
Endrin	5.791	5.692	5.892	48.650	50.000	-2.7
Endrin aldehyde	6.260	6.162	6.362	46.280	50.000	-7.4
Endrin ketone	6.993	6.895	7.095	47.210	50.000	-5.6
gamma-BHC (Lindane)	3.731	3.631	3.831	50.790	50.000	1.6
gamma-Chlordane	5.127	5.029	5.229	49.950	50.000	-0.1
Heptachlor	4.085	3.985	4.185	48.750	50.000	-2.5
Heptachlor epoxide	4.874	4.775	4.975	49.450	50.000	-1.1
Methoxychlor	6.756	6.658	6.858	44.530	50.000	-10.9
Tetrachloro-m-xylene	2.882	2.782	2.982	56.190	50.000	12.4