

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/20/2025

Lab Sample No.: PSTDCCC050 Data File : PD088674.D Time Analyzed: 20:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.706	6.608	6.808	54.060	50.000	8.1
4,4'-DDE	6.197	6.098	6.298	50.980	50.000	2.0
4,4'-DDT	7.021	6.923	7.123	42.100	50.000	-15.8
Aldrin	5.273	5.173	5.373	52.660	50.000	5.3
alpha-BHC	4.001	3.900	4.100	53.540	50.000	7.1
alpha-Chlordane	6.028	5.930	6.130	51.370	50.000	2.7
beta-BHC	4.516	4.417	4.617	51.570	50.000	3.1
Decachlorobiphenyl	9.075	8.977	9.177	53.190	50.000	6.4
delta-BHC	4.765	4.666	4.866	55.920	50.000	11.8
Dieldrin	6.348	6.250	6.450	51.610	50.000	3.2
Endosulfan I	6.076	5.977	6.177	51.080	50.000	2.2
Endosulfan II	6.787	6.688	6.888	49.400	50.000	-1.2
Endosulfan sulfate	7.151	7.052	7.252	50.980	50.000	2.0
Endrin	6.576	6.477	6.677	50.860	50.000	1.7
Endrin aldehyde	6.916	6.817	7.017	48.290	50.000	-3.4
Endrin ketone	7.631	7.533	7.733	50.150	50.000	0.3
gamma-BHC (Lindane)	4.332	4.232	4.432	52.830	50.000	5.7
gamma-Chlordane	5.947	5.848	6.048	51.770	50.000	3.5
Heptachlor	4.931	4.832	5.032	51.470	50.000	2.9
Heptachlor epoxide	5.692	5.593	5.793	51.690	50.000	3.4
Methoxychlor	7.494	7.395	7.595	43.420	50.000	-13.2
Tetrachloro-m-xylene	3.551	3.451	3.651	57.440	50.000	14.9

CALIBRATION VERIFICATION SUMMARY

Contract: CAMP02

Lab Code: CHEM Case No.: Q2074 SAS No.: Q2074 SDG NO.: Q2074

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 05/19/2025 05/19/2025

Client Sample No.: CCAL02 Date Analyzed: 05/20/2025

Lab Sample No.: PSTDCCC050 Data File : PD088674.D Time Analyzed: 20:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.931	5.833	6.033	51.370	50.000	2.7
4,4'-DDE	5.375	5.278	5.478	49.730	50.000	-0.5
4,4'-DDT	6.185	6.087	6.287	44.760	50.000	-10.5
Aldrin	4.370	4.271	4.471	51.920	50.000	3.8
alpha-BHC	3.394	3.294	3.494	51.880	50.000	3.8
alpha-Chlordane	5.192	5.094	5.294	49.900	50.000	-0.2
beta-BHC	4.026	3.927	4.127	51.430	50.000	2.9
Decachlorobiphenyl	8.074	7.977	8.177	50.710	50.000	1.4
delta-BHC	4.263	4.164	4.364	51.610	50.000	3.2
Dieldrin	5.514	5.416	5.616	49.340	50.000	-1.3
Endosulfan I	5.248	5.150	5.350	46.770	50.000	-6.5
Endosulfan II	6.082	5.984	6.184	49.660	50.000	-0.7
Endosulfan sulfate	6.483	6.386	6.586	48.120	50.000	-3.8
Endrin	5.791	5.692	5.892	49.400	50.000	-1.2
Endrin aldehyde	6.260	6.162	6.362	47.470	50.000	-5.1
Endrin ketone	6.993	6.895	7.095	48.220	50.000	-3.6
gamma-BHC (Lindane)	3.731	3.631	3.831	51.590	50.000	3.2
gamma-Chlordane	5.127	5.029	5.229	50.730	50.000	1.5
Heptachlor	4.084	3.985	4.185	49.920	50.000	-0.2
Heptachlor epoxide	4.874	4.775	4.975	50.390	50.000	0.8
Methoxychlor	6.756	6.658	6.858	45.850	50.000	-8.3
Tetrachloro-m-xylene	2.882	2.782	2.982	56.890	50.000	13.8