

Surrogate Summary

SDG No.: Q2074

Client: CDM Smith

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PD088583.D	PIBLK-PD088583.D	Decachlorobiphenyl	1	20	18.4	92		43	140
		Tetrachloro-m-xylene	1	20	16.5	83		77	126
		Decachlorobiphenyl	2	20	18.0	90		43	140
		Tetrachloro-m-xylene	2	20	16.9	84		77	126
I.BLK-PD088656.D	PIBLK-PD088656.D	Decachlorobiphenyl	1	20	18.2	91		43	140
		Tetrachloro-m-xylene	1	20	18.5	92		77	126
		Decachlorobiphenyl	2	20	17.1	85		43	140
		Tetrachloro-m-xylene	2	20	18.8	94		77	126
PB168075BL	PB168075BL	Decachlorobiphenyl	1	20	18.7	93		20	144
		Tetrachloro-m-xylene	1	20	17.4	87		19	148
		Decachlorobiphenyl	2	20	17.4	87		20	144
		Tetrachloro-m-xylene	2	20	19.1	95		19	148
PB168075BS	PB168075BS	Decachlorobiphenyl	1	20	24.6	123		20	144
		Tetrachloro-m-xylene	1	20	23.5	117		19	148
		Decachlorobiphenyl	2	20	23.8	119		20	144
		Tetrachloro-m-xylene	2	20	25.2	126		19	148
Q2074-01	TP-12	Decachlorobiphenyl	1	20	18.1	90		20	144
		Tetrachloro-m-xylene	1	20	17.4	87		19	148
		Decachlorobiphenyl	2	20	15.5	78		20	144
		Tetrachloro-m-xylene	2	20	18.4	92		19	148
Q2074-02	TP-7	Decachlorobiphenyl	1	20	20.0	100		20	144
		Tetrachloro-m-xylene	1	20	20.7	104		19	148
		Decachlorobiphenyl	2	20	17.7	88		20	144
		Tetrachloro-m-xylene	2	20	21.8	109		19	148
Q2074-03	TP-15	Decachlorobiphenyl	1	20	13.3	67		20	144
		Tetrachloro-m-xylene	1	20	17.7	88		19	148
		Decachlorobiphenyl	2	20	13.2	66		20	144
		Tetrachloro-m-xylene	2	20	18.6	93		19	148
Q2074-04	TP-20	Decachlorobiphenyl	1	20	16.6	83		20	144
		Tetrachloro-m-xylene	1	20	17.9	89		19	148
		Decachlorobiphenyl	2	20	15.4	77		20	144
		Tetrachloro-m-xylene	2	20	19.0	95		19	148
Q2074-04MS	TP-20MS	Decachlorobiphenyl	1	20	17.3	87		20	144
		Tetrachloro-m-xylene	1	20	17.8	89		19	148
		Decachlorobiphenyl	2	20	16.7	83		20	144
		Tetrachloro-m-xylene	2	20	19.6	98		19	148
Q2074-04MSD	TP-20MSD	Decachlorobiphenyl	1	20	17.6	88		20	144
		Tetrachloro-m-xylene	1	20	17.8	89		19	148
		Decachlorobiphenyl	2	20	16.5	83		20	144
		Tetrachloro-m-xylene	2	20	19.7	98		19	148
Q2074-05	TP-38	Decachlorobiphenyl	1	20	14.1	70		20	144

Surrogate Summary

SDG No.: Q2074

Client: CDM Smith

Analytical Method: 8081B

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q2074-05	TP-38	Tetrachloro-m-xylene	1	20	16.7	83		19	148
		Decachlorobiphenyl	2	20	13.3	67		20	144
Q2074-06	TP-19	Tetrachloro-m-xylene	2	20	17.6	88		19	148
		Decachlorobiphenyl	1	20	15.7	78		20	144
		Tetrachloro-m-xylene	1	20	17.6	88		19	148
		Decachlorobiphenyl	2	20	14.6	73		20	144
		Tetrachloro-m-xylene	2	20	18.5	92		19	148
Q2074-07	TP-40	Decachlorobiphenyl	1	20	15.5	77		20	144
		Tetrachloro-m-xylene	1	20	18.5	93		19	148
		Decachlorobiphenyl	2	20	15.3	76		20	144
		Tetrachloro-m-xylene	2	20	19.2	96		19	148
Q2074-08	TP-18	Decachlorobiphenyl	1	20	14.8	74		20	144
		Tetrachloro-m-xylene	1	20	17.4	87		19	148
		Decachlorobiphenyl	2	20	14.1	70		20	144
		Tetrachloro-m-xylene	2	20	18.3	91		19	148
I.BLK-PD088673.D	PIBLK-PD088673.D	Decachlorobiphenyl	1	20	18.8	94		43	140
		Tetrachloro-m-xylene	1	20	18.4	92		77	126
		Decachlorobiphenyl	2	20	18.5	93		43	140
		Tetrachloro-m-xylene	2	20	19.1	96		77	126