

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: bf052125

SequenceID : bf052125

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓
5. GC/MS Calibration Met:	_____	_____	_____✓
a. Initial calibration Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
b. Continuous Calibration(CCC) Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓	_____
a. B/N Fraction			
d. Acid Fraction			

7. Surrogate Recoveries Meet Criteria	_____	_____	✓
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
a. B/N Fraction			
d. Acid Fraction			
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria	_____	_____	✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
a. B/N Fraction			
d. Acid Fraction			
9. Internal Standard Area/Retention Time Shift Meet Criteria	_____	✓	_____
Comments:			
Internal standard failed in both Q2075-04MS/MSD due to matrix interference. hence no corrective action is required. and Internal standards failed in the sample Q2075-06 due to the presence of non-targeted hydrocarbons which can be observed by the abnormal chromatogram. Hence this analysis will be final.			
10. Extraction Holding Time Met	_____	_____	✓
If not met, list number of days exceeded for each sample:			
11. Analysis Holding Time Met	_____	_____	✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The sample Q2068-01 had to be analyzed with 10X dilution due to oily matrix, Internal standard failed even in this diluted sample proving matrix interference which can also be witnessed by the abnormal chromatogram. Therefore this analysis will be final.



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,  
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

**GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**

CHEMTECH PROJECT NUMBER: bf052125

SequenceID : bf052125 NA NO YES