

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2074

SDG NO.: Q2074

Lab File ID: BF142477.D

DFTPP Injection Date: 05/21/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:41

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	33.5
68	Less than 2.0% of mass 69	0.6 (2) 1
69	Mass 69 relative abundance	29.3
70	Less than 2.0% of mass 69	0.2 (0.7) 1
127	10.0 - 80.0% of mass 198	39.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	5.7
275	10.0 - 60.0% of mass 198	24.4
365	Greater than 1% of mass 198	3.2
441	Present, but less than mass 443	15.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142478.D	05/21/2025	10:40
PB168083BL	PB168083BL	BF142479.D	05/21/2025	11:09
PB168083BS	PB168083BS	BF142480.D	05/21/2025	11:37
TP-7	Q2074-02	BF142481.D	05/21/2025	12:14
TP-20	Q2074-04	BF142482.D	05/21/2025	12:43
TP-19	Q2074-06	BF142484.D	05/21/2025	13:40
SS-11MS	Q2075-04MS	BF142487.D	05/21/2025	15:06
SS-11MSD	Q2075-05MSD	BF142488.D	05/21/2025	15:34