

Surrogate Summary

SDG No.: Q2074

Client: CDM Smith

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q2074-01	TP-12	1,2-Dichloroethane-d4	50	48.7	97	63	155
		Dibromofluoromethane	50	50.7	101	70	134
		Toluene-d8	50	48.7	97	74	123
		4-Bromofluorobenzene	50	40.3	81	38	136
Q2074-02	TP-7	1,2-Dichloroethane-d4	50	47.6	95	63	155
		Dibromofluoromethane	50	50.0	100	70	134
		Toluene-d8	50	48.5	97	74	123
		4-Bromofluorobenzene	50	39.7	79	38	136
Q2074-03	TP-15	1,2-Dichloroethane-d4	50	45.2	90	63	155
		Dibromofluoromethane	50	49.5	99	70	134
		Toluene-d8	50	48.2	96	74	123
		4-Bromofluorobenzene	50	37.1	74	38	136
Q2074-04	TP-20	1,2-Dichloroethane-d4	50	49.9	100	63	155
		Dibromofluoromethane	50	50.7	101	70	134
		Toluene-d8	50	49.3	99	74	123
		4-Bromofluorobenzene	50	40.1	80	38	136
Q2074-05	TP-38	1,2-Dichloroethane-d4	50	47.5	95	63	155
		Dibromofluoromethane	50	50.7	101	70	134
		Toluene-d8	50	49.0	98	74	123
		4-Bromofluorobenzene	50	39.7	79	38	136
Q2074-06	TP-19	1,2-Dichloroethane-d4	50	46.6	93	63	155
		Dibromofluoromethane	50	49.7	99	70	134
		Toluene-d8	50	48.0	96	74	123
		4-Bromofluorobenzene	50	33.7	67	38	136
Q2074-07	TP-40	1,2-Dichloroethane-d4	50	46.7	93	63	155
		Dibromofluoromethane	50	49.0	98	70	134
		Toluene-d8	50	47.7	95	74	123
		4-Bromofluorobenzene	50	36.5	73	38	136
Q2074-08	TP-18	1,2-Dichloroethane-d4	50	46.9	94	63	155
		Dibromofluoromethane	50	48.5	97	70	134
		Toluene-d8	50	47.9	96	74	123
		4-Bromofluorobenzene	50	37.3	75	38	136
VY0522SBL01	VY0522SBL01	1,2-Dichloroethane-d4	50	48.9	98	63	155
		Dibromofluoromethane	50	49.8	100	70	134
		Toluene-d8	50	48.7	97	74	123
		4-Bromofluorobenzene	50	39.5	79	38	136
VY0522SBS01	VY0522SBS01	1,2-Dichloroethane-d4	50	48.4	97	63	155
		Dibromofluoromethane	50	50.3	101	70	134
		Toluene-d8	50	50.3	101	74	123
		4-Bromofluorobenzene	50	47.7	95	38	136
VY0522SBSD01	VY0522SBSD01	1,2-Dichloroethane-d4	50	50.4	101	63	155
		Dibromofluoromethane	50	51.0	102	70	134
		Toluene-d8	50	49.9	100	74	123
		4-Bromofluorobenzene	50	47.1	94	38	136