

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168098BSD	SDG No.:	Q2075
Lab Sample ID:	PB168098BSD	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024756.D	1	05/21/25 08:40	05/22/25 12:58	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
91-20-3	Naphthalene	43.4		0.50	5.00	ug/L
208-96-8	Acenaphthylene	43.6		0.75	5.00	ug/L
83-32-9	Acenaphthene	43.5		0.55	5.00	ug/L
86-73-7	Fluorene	44.9		0.63	5.00	ug/L
85-01-8	Phenanthrene	44.0		0.50	5.00	ug/L
120-12-7	Anthracene	45.1		0.61	5.00	ug/L
206-44-0	Fluoranthene	45.8		0.82	5.00	ug/L
129-00-0	Pyrene	45.6		0.50	5.00	ug/L
56-55-3	Benzo(a)anthracene	45.5		0.45	5.00	ug/L
218-01-9	Chrysene	44.4		0.44	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	46.7		0.49	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	45.6		0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	46.1		0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	44.8		0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	45.2		0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	43.9		0.69	5.00	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	73.9		49 - 133	74%	SPK: 100
321-60-8	2-Fluorobiphenyl	76.8		52 - 132	77%	SPK: 100
1718-51-0	Terphenyl-d14	81.5		48 - 125	81%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	126000		7.652		
1146-65-2	Naphthalene-d8	523000		10.416		
15067-26-2	Acenaphthene-d10	339000		14.287		
1517-22-2	Phenanthrene-d10	687000		17.092		
1719-03-5	Chrysene-d12	769000		21.51		
1520-96-3	Perylene-d12	836000		24.786		

Report of Analysis

Client:	CDM Smith		Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY		Date Received:	
Client Sample ID:	PB168098BSD		SDG No.:	Q2075
Lab Sample ID:	PB168098BSD		Matrix:	Water
Analytical Method:	8270E		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOCMS Group3
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024756.D	1	05/21/25 08:40	05/22/25 12:58	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products