

LAB CHRONICLE

OrderID:	Q2075	OrderDate:	5/16/2025 4:10:00 PM
Client:	CDM Smith	Project:	Con Ed UTEN Mount Vernon, NY
Contact:	Marcie Ann Encinas	Location:	L41,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q2075-01	SS-10	SOIL			05/15/25			05/16/25
			SVOCMS Group3	8270E		05/20/25	05/22/25	
Q2075-02	SS-910	SOIL			05/15/25			05/16/25
			SVOCMS Group3	8270E		05/20/25	05/22/25	
Q2075-03	SS-11	SOIL			05/15/25			05/16/25
			SVOCMS Group3	8270E		05/20/25	05/21/25	
Q2075-06	SS-MW1-11.5	SOIL			05/16/25			05/16/25
			SVOCMS Group3	8270E		05/20/25	05/21/25	
Q2075-07	FB-05152025	Water			05/15/25			05/16/25
			SVOCMS Group3	8270E		05/21/25	05/22/25	
Q2075-08	FB-05162025	Water			05/16/25			05/16/25
			SVOCMS Group3	8270E		05/21/25	05/21/25	

Hit Summary Sheet SW-846

SDG No.: Q2075
Client: CDM Smith

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : SS-10								
Q2075-01	SS-10	SOIL	Phenanthrene	830.000		26.2	210	ug/Kg
Q2075-01	SS-10	SOIL	Anthracene	110.000	J	41.7	210	ug/Kg
Q2075-01	SS-10	SOIL	Fluoranthene	1,200.000		37.6	210	ug/Kg
Q2075-01	SS-10	SOIL	Pyrene	660.000		45.1	210	ug/Kg
Q2075-01	SS-10	SOIL	Benzo(a)anthracene	420.000		28.8	210	ug/Kg
Q2075-01	SS-10	SOIL	Chrysene	490.000		24.9	210	ug/Kg
Q2075-01	SS-10	SOIL	Benzo(b)fluoranthene	640.000		23.8	210	ug/Kg
Q2075-01	SS-10	SOIL	Benzo(k)fluoranthene	220.000		28.1	210	ug/Kg
Q2075-01	SS-10	SOIL	Benzo(a)pyrene	350.000		37	210	ug/Kg
Q2075-01	SS-10	SOIL	Indeno(1,2,3-cd)pyrene	160.000	J	36.5	210	ug/Kg
Q2075-01	SS-10	SOIL	Benzo(g,h,i)perylene	170.000	J	32.2	210	ug/Kg
Total Svoc :				5,250.00				
Total Concentration:				5,250.00				
Client ID : SS-910								
Q2075-02	SS-910	SOIL	Phenanthrene	440.000		25.7	210	ug/Kg
Q2075-02	SS-910	SOIL	Fluoranthene	940.000		36.8	210	ug/Kg
Q2075-02	SS-910	SOIL	Pyrene	490.000		44.2	210	ug/Kg
Q2075-02	SS-910	SOIL	Benzo(a)anthracene	290.000		28.2	210	ug/Kg
Q2075-02	SS-910	SOIL	Chrysene	320.000		24.4	210	ug/Kg
Q2075-02	SS-910	SOIL	Benzo(b)fluoranthene	520.000		23.3	210	ug/Kg
Q2075-02	SS-910	SOIL	Benzo(k)fluoranthene	180.000	J	27.5	210	ug/Kg
Q2075-02	SS-910	SOIL	Benzo(a)pyrene	310.000		36.2	210	ug/Kg
Q2075-02	SS-910	SOIL	Indeno(1,2,3-cd)pyrene	150.000	J	35.7	210	ug/Kg
Q2075-02	SS-910	SOIL	Benzo(g,h,i)perylene	170.000	J	31.5	210	ug/Kg
Total Svoc :				3,810.00				
Total Concentration:				3,810.00				
Client ID : SS-11								
Q2075-03	SS-11	SOIL	Fluoranthene	120.000	J	34.3	190	ug/Kg
Total Svoc :				120.00				
Total Concentration:				120.00				



SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-10	SDG No.:	Q2075
Lab Sample ID:	Q2075-01	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	79.8
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Level :	LOW
		Final Vol:	1000 uL
		Test:	SVOCMS Group3
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142510.D	1	05/20/25 09:45	05/22/25 15:07	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	210	U	28.4	210	ug/Kg
208-96-8	Acenaphthylene	210	U	36.2	210	ug/Kg
83-32-9	Acenaphthene	210	U	26.7	210	ug/Kg
86-73-7	Fluorene	210	U	31.7	210	ug/Kg
85-01-8	Phenanthrene	830		26.2	210	ug/Kg
120-12-7	Anthracene	110	J	41.7	210	ug/Kg
206-44-0	Fluoranthene	1200		37.6	210	ug/Kg
129-00-0	Pyrene	660		45.1	210	ug/Kg
56-55-3	Benzo(a)anthracene	420		28.8	210	ug/Kg
218-01-9	Chrysene	490		24.9	210	ug/Kg
205-99-2	Benzo(b)fluoranthene	640		23.8	210	ug/Kg
207-08-9	Benzo(k)fluoranthene	220		28.1	210	ug/Kg
50-32-8	Benzo(a)pyrene	350		37.0	210	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	160	J	36.5	210	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	210	U	34.3	210	ug/Kg
191-24-2	Benzo(g,h,i)perylene	170	J	32.2	210	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	43.6		18 - 107	44%	SPK: 100
321-60-8	2-Fluorobiphenyl	37.5		20 - 109	38%	SPK: 100
1718-51-0	Terphenyl-d14	29.7		10 - 105	30%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	120000		6.898		
1146-65-2	Naphthalene-d8	453000		8.18		
15067-26-2	Acenaphthene-d10	223000		9.933		
1517-22-2	Phenanthrene-d10	308000		11.421		
1719-03-5	Chrysene-d12	229000		14.062		
1520-96-3	Perylene-d12	198000		15.556		

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-10	SDG No.:	Q2075
Lab Sample ID:	Q2075-01	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	79.8
Sample Wt/Vol:	30.01	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	SVOCMS Group3
	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	1.0	GPC Cleanup :
		N	PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142510.D	1	05/20/25 09:45	05/22/25 15:07	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-910	SDG No.:	Q2075
Lab Sample ID:	Q2075-02	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	81.3
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Final Vol:	1000
		Test:	SVOCMS Group3
		Level :	LOW
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142511.D	1	05/20/25 09:45	05/22/25 15:36	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	210	U	27.9	210	ug/Kg
208-96-8	Acenaphthylene	210	U	35.5	210	ug/Kg
83-32-9	Acenaphthene	210	U	26.1	210	ug/Kg
86-73-7	Fluorene	210	U	31.1	210	ug/Kg
85-01-8	Phenanthrene	440		25.7	210	ug/Kg
120-12-7	Anthracene	210	U	40.9	210	ug/Kg
206-44-0	Fluoranthene	940		36.8	210	ug/Kg
129-00-0	Pyrene	490		44.2	210	ug/Kg
56-55-3	Benzo(a)anthracene	290		28.2	210	ug/Kg
218-01-9	Chrysene	320		24.4	210	ug/Kg
205-99-2	Benzo(b)fluoranthene	520		23.3	210	ug/Kg
207-08-9	Benzo(k)fluoranthene	180	J	27.5	210	ug/Kg
50-32-8	Benzo(a)pyrene	310		36.2	210	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	150	J	35.7	210	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	210	U	33.6	210	ug/Kg
191-24-2	Benzo(g,h,i)perylene	170	J	31.5	210	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	49.5		18 - 107	49%	SPK: 100
321-60-8	2-Fluorobiphenyl	43.1		20 - 109	43%	SPK: 100
1718-51-0	Terphenyl-d14	35.4		10 - 105	35%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	117000		6.898		
1146-65-2	Naphthalene-d8	418000		8.181		
15067-26-2	Acenaphthene-d10	191000		9.934		
1517-22-2	Phenanthrene-d10	271000		11.422		
1719-03-5	Chrysene-d12	225000		14.063		
1520-96-3	Perylene-d12	177000		15.557		

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-910	SDG No.:	Q2075
Lab Sample ID:	Q2075-02	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	81.3
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	SVOCMS Group3
	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	1.0	GPC Cleanup :
		N	PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142511.D	1	05/20/25 09:45	05/22/25 15:36	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-11	SDG No.:	Q2075
Lab Sample ID:	Q2075-03	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	87.4
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Final Vol:	1000
		Test:	SVOCMS Group3
		Level :	LOW
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142486.D	1	05/20/25 09:45	05/21/25 14:37	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	190	U	25.9	190	ug/Kg
208-96-8	Acenaphthylene	190	U	33.0	190	ug/Kg
83-32-9	Acenaphthene	190	U	24.3	190	ug/Kg
86-73-7	Fluorene	190	U	28.9	190	ug/Kg
85-01-8	Phenanthrene	190	U	23.9	190	ug/Kg
120-12-7	Anthracene	190	U	38.1	190	ug/Kg
206-44-0	Fluoranthene	120	J	34.3	190	ug/Kg
129-00-0	Pyrene	190	U	41.1	190	ug/Kg
56-55-3	Benzo(a)anthracene	190	U	26.3	190	ug/Kg
218-01-9	Chrysene	190	U	22.7	190	ug/Kg
205-99-2	Benzo(b)fluoranthene	190	U	21.7	190	ug/Kg
207-08-9	Benzo(k)fluoranthene	190	U	25.6	190	ug/Kg
50-32-8	Benzo(a)pyrene	190	U	33.7	190	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	190	U	33.3	190	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	190	U	31.3	190	ug/Kg
191-24-2	Benzo(g,h,i)perylene	190	U	29.4	190	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	64.7		18 - 107	65%	SPK: 100
321-60-8	2-Fluorobiphenyl	56.2		20 - 109	56%	SPK: 100
1718-51-0	Terphenyl-d14	42.2		10 - 105	42%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	63600		6.898		
1146-65-2	Naphthalene-d8	239000		8.181		
15067-26-2	Acenaphthene-d10	132000		9.939		
1517-22-2	Phenanthrene-d10	251000		11.427		
1719-03-5	Chrysene-d12	246000		14.068		
1520-96-3	Perylene-d12	209000		15.557		

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-11	SDG No.:	Q2075
Lab Sample ID:	Q2075-03	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	87.4
Sample Wt/Vol:	30.04	Units:	g
Soil Aliquot Vol:		Final Vol:	1000 uL
Extraction Type :		Test:	SVOCMS Group3
	Decanted :	Level :	LOW
Injection Volume :	GPC Factor :	1.0	GPC Cleanup :
		N	PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142486.D	1	05/20/25 09:45	05/21/25 14:37	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	05/16/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-MW1-11.5	SDG No.:	Q2075
Lab Sample ID:	Q2075-06	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	91
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Final Vol:	1000
		Test:	SVOCMS Group3
		Level :	LOW
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142489.D	1	05/20/25 09:45	05/21/25 16:03	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	190	U	24.9	190	ug/Kg
208-96-8	Acenaphthylene	190	U	31.7	190	ug/Kg
83-32-9	Acenaphthene	190	U	23.4	190	ug/Kg
86-73-7	Fluorene	190	U	27.7	190	ug/Kg
85-01-8	Phenanthrene	190	U	22.9	190	ug/Kg
120-12-7	Anthracene	190	U	36.5	190	ug/Kg
206-44-0	Fluoranthene	190	U	32.9	190	ug/Kg
129-00-0	Pyrene	190	U	39.5	190	ug/Kg
56-55-3	Benzo(a)anthracene	190	U	25.2	190	ug/Kg
218-01-9	Chrysene	190	U	21.8	190	ug/Kg
205-99-2	Benzo(b)fluoranthene	190	U	20.8	190	ug/Kg
207-08-9	Benzo(k)fluoranthene	190	U	24.6	190	ug/Kg
50-32-8	Benzo(a)pyrene	190	U	32.4	190	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	190	U	31.9	190	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	190	U	30.0	190	ug/Kg
191-24-2	Benzo(g,h,i)perylene	190	U	28.2	190	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	53.3		18 - 107	53%	SPK: 100
321-60-8	2-Fluorobiphenyl	48.6		20 - 109	49%	SPK: 100
1718-51-0	Terphenyl-d14	38.9		10 - 105	39%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	59300	6.898			
1146-65-2	Naphthalene-d8	207000	8.181			
15067-26-2	Acenaphthene-d10	104000	9.939			
1517-22-2	Phenanthrene-d10	179000	11.422			
1719-03-5	Chrysene-d12	153000	14.063			
1520-96-3	Perylene-d12	126000	15.557			

Report of Analysis

Client:	CDM Smith	Date Collected:	05/16/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-MW1-11.5	SDG No.:	Q2075
Lab Sample ID:	Q2075-06	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	91
Sample Wt/Vol:	30.06	Units:	g
Soil Aliquot Vol:			uL
Extraction Type :		Decanted :	N
Injection Volume :		GPC Factor :	1.0
Prep Method :	SW3541	Level :	LOW
		Final Vol:	1000
		Test:	SVOCMS Group3
		GPC Cleanup :	N
		PH :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142489.D	1	05/20/25 09:45	05/21/25 16:03	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	FB-05152025	SDG No.:	Q2075
Lab Sample ID:	Q2075-07	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	810 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024758.D	1	05/21/25 08:40	05/22/25 14:19	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
91-20-3	Naphthalene	6.20	U	0.62	6.20	ug/L
208-96-8	Acenaphthylene	6.20	U	0.93	6.20	ug/L
83-32-9	Acenaphthene	6.20	U	0.68	6.20	ug/L
86-73-7	Fluorene	6.20	U	0.78	6.20	ug/L
85-01-8	Phenanthrene	6.20	U	0.62	6.20	ug/L
120-12-7	Anthracene	6.20	U	0.75	6.20	ug/L
206-44-0	Fluoranthene	6.20	U	1.00	6.20	ug/L
129-00-0	Pyrene	6.20	U	0.62	6.20	ug/L
56-55-3	Benzo(a)anthracene	6.20	U	0.56	6.20	ug/L
218-01-9	Chrysene	6.20	U	0.54	6.20	ug/L
205-99-2	Benzo(b)fluoranthene	6.20	U	0.60	6.20	ug/L
207-08-9	Benzo(k)fluoranthene	6.20	U	0.59	6.20	ug/L
50-32-8	Benzo(a)pyrene	6.20	U	0.68	6.20	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	6.20	U	0.73	6.20	ug/L
53-70-3	Dibenzo(a,h)anthracene	6.20	U	0.83	6.20	ug/L
191-24-2	Benzo(g,h,i)perylene	6.20	U	0.85	6.20	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	66.0		49 - 133	66%	SPK: 100
321-60-8	2-Fluorobiphenyl	63.0		52 - 132	63%	SPK: 100
1718-51-0	Terphenyl-d14	83.4		48 - 125	83%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	104000		7.652		
1146-65-2	Naphthalene-d8	395000		10.422		
15067-26-2	Acenaphthene-d10	245000		14.287		
1517-22-2	Phenanthrene-d10	487000		17.086		
1719-03-5	Chrysene-d12	600000		21.522		
1520-96-3	Perylene-d12	736000		24.815		

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	FB-05152025	SDG No.:	Q2075
Lab Sample ID:	Q2075-07	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	810 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024758.D	1	05/21/25 08:40	05/22/25 14:19	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	05/16/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	FB-05162025	SDG No.:	Q2075
Lab Sample ID:	Q2075-08	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	940 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024748.D	1	05/21/25 08:40	05/21/25 22:18	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
91-20-3	Naphthalene	5.30	U	0.53	5.30	ug/L
208-96-8	Acenaphthylene	5.30	U	0.80	5.30	ug/L
83-32-9	Acenaphthene	5.30	U	0.59	5.30	ug/L
86-73-7	Fluorene	5.30	U	0.67	5.30	ug/L
85-01-8	Phenanthrene	5.30	U	0.53	5.30	ug/L
120-12-7	Anthracene	5.30	U	0.65	5.30	ug/L
206-44-0	Fluoranthene	5.30	U	0.87	5.30	ug/L
129-00-0	Pyrene	5.30	U	0.53	5.30	ug/L
56-55-3	Benzo(a)anthracene	5.30	U	0.48	5.30	ug/L
218-01-9	Chrysene	5.30	U	0.47	5.30	ug/L
205-99-2	Benzo(b)fluoranthene	5.30	U	0.52	5.30	ug/L
207-08-9	Benzo(k)fluoranthene	5.30	U	0.51	5.30	ug/L
50-32-8	Benzo(a)pyrene	5.30	U	0.59	5.30	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.30	U	0.63	5.30	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.30	U	0.71	5.30	ug/L
191-24-2	Benzo(g,h,i)perylene	5.30	U	0.73	5.30	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	73.7		49 - 133	74%	SPK: 100
321-60-8	2-Fluorobiphenyl	74.4		52 - 132	74%	SPK: 100
1718-51-0	Terphenyl-d14	89.5		48 - 125	90%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	87700		7.652		
1146-65-2	Naphthalene-d8	326000		10.422		
15067-26-2	Acenaphthene-d10	217000		14.286		
1517-22-2	Phenanthrene-d10	458000		17.092		
1719-03-5	Chrysene-d12	547000		21.51		
1520-96-3	Perylene-d12	672000		24.798		

Report of Analysis

Client:	CDM Smith	Date Collected:	05/16/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	FB-05162025	SDG No.:	Q2075
Lab Sample ID:	Q2075-08	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	940 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024748.D	1	05/21/25 08:40	05/21/25 22:18	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168083BL	PB168083BL	Nitrobenzene-d5	100	67.4	67		18	107
		2-Fluorobiphenyl	100	69.4	69		20	109
		Terphenyl-d14	100	62.4	62		10	105
PB168083BS	PB168083BS	Nitrobenzene-d5	100	69.4	69		18	107
		2-Fluorobiphenyl	100	66.0	66		20	109
		Terphenyl-d14	100	81.6	82		10	105
Q2075-01	SS-10	Nitrobenzene-d5	100	43.6	44		18	107
		2-Fluorobiphenyl	100	37.5	38		20	109
		Terphenyl-d14	100	29.7	30		10	105
Q2075-02	SS-910	Nitrobenzene-d5	100	49.5	49		18	107
		2-Fluorobiphenyl	100	43.1	43		20	109
		Terphenyl-d14	100	35.4	35		10	105
Q2075-03	SS-11	Nitrobenzene-d5	100	64.7	65		18	107
		2-Fluorobiphenyl	100	56.2	56		20	109
		Terphenyl-d14	100	42.2	42		10	105
Q2075-04MS	SS-11MS	Nitrobenzene-d5	100	60.4	60		18	107
		2-Fluorobiphenyl	100	51.6	52		20	109
		Terphenyl-d14	100	41.5	41		10	105
Q2075-05MSD	SS-11MSD	Nitrobenzene-d5	100	59.6	60		18	107
		2-Fluorobiphenyl	100	54.4	54		20	109
		Terphenyl-d14	100	40.2	40		10	105
Q2075-06	SS-MW1-11.5	Nitrobenzene-d5	100	53.3	53		18	107
		2-Fluorobiphenyl	100	48.6	49		20	109
		Terphenyl-d14	100	38.9	39		10	105

Surrogate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB168098BL	PB168098BL	Nitrobenzene-d5	100	75.0	75		49	133
		2-Fluorobiphenyl	100	77.3	77		52	132
		Terphenyl-d14	100	83.0	83		48	125
PB168098BS	PB168098BS	Nitrobenzene-d5	100	72.8	73		49	133
		2-Fluorobiphenyl	100	76.3	76		52	132
		Terphenyl-d14	100	77.9	78		48	125
PB168098BSD	PB168098BSD	Nitrobenzene-d5	100	73.9	74		49	133
		2-Fluorobiphenyl	100	76.8	77		52	132
		Terphenyl-d14	100	81.5	81		48	125
Q2075-07	FB-05152025	Nitrobenzene-d5	100	66.0	66		49	133
		2-Fluorobiphenyl	100	63.0	63		52	132
		Terphenyl-d14	100	83.4	83		48	125
Q2075-08	FB-05162025	Nitrobenzene-d5	100	73.7	74		49	133
		2-Fluorobiphenyl	100	74.4	74		52	132
		Terphenyl-d14	100	89.5	90		48	125

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2075-04MS	Client Sample ID:	SS-11MS					DataFile:	BF142487.D		
Naphthalene	1900	0	1600	ug/Kg	84				72	110	
Acenaphthylene	1900	0	1600	ug/Kg	84				79	118	
Acenaphthene	1900	0	1800	ug/Kg	95				70	121	
Fluorene	1900	0	1700	ug/Kg	89				68	116	
Phenanthrene	1900	0	1600	ug/Kg	84				52	128	
Anthracene	1900	0	1700	ug/Kg	89				62	124	
Fluoranthene	1900	120	2200	ug/Kg	109				44	125	
Pyrene	1900	0	1300	ug/Kg	68				26	142	
Benzo(a)anthracene	1900	0	1700	ug/Kg	89				71	114	
Chrysene	1900	0	1600	ug/Kg	84				57	121	
Benzo(b)fluoranthene	1900	0	1900	ug/Kg	100				67	121	
Benzo(k)fluoranthene	1900	0	1600	ug/Kg	84				57	134	
Benzo(a)pyrene	1900	0	1700	ug/Kg	89				70	142	
Indeno(1,2,3-cd)pyrene	1900	0	1400	ug/Kg	74				40	129	
Dibenz(a,h)anthracene	1900	0	1400	ug/Kg	74				43	123	
Benzo(g,h,i)perylene	1900	0	1300	ug/Kg	68				24	125	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q2075-05MSD	Client Sample ID:	SS-11MSD					DataFile:	BF142488.D		
Naphthalene	1900	0	1600	ug/Kg	84	0			72	110	20
Acenaphthylene	1900	0	1600	ug/Kg	84	0			79	118	20
Acenaphthene	1900	0	1800	ug/Kg	95	0			70	121	20
Fluorene	1900	0	1600	ug/Kg	84	6			68	116	20
Phenanthrene	1900	0	1700	ug/Kg	89	6			52	128	20
Anthracene	1900	0	1700	ug/Kg	89	0			62	124	20
Fluoranthene	1900	120	2300	ug/Kg	115	5			44	125	20
Pyrene	1900	0	1300	ug/Kg	68	0			26	142	20
Benzo(a)anthracene	1900	0	1600	ug/Kg	84	6			71	114	20
Chrysene	1900	0	1600	ug/Kg	84	0			57	121	20
Benzo(b)fluoranthene	1900	0	1700	ug/Kg	89	12			67	121	20
Benzo(k)fluoranthene	1900	0	1700	ug/Kg	89	6			57	134	20
Benzo(a)pyrene	1900	0	1700	ug/Kg	89	0			70	142	20
Indeno(1,2,3-cd)pyrene	1900	0	1400	ug/Kg	74	0			40	129	20
Dibenz(a,h)anthracene	1900	0	1300	ug/Kg	68	8			43	123	20
Benzo(g,h,i)perylene	1900	0	1300	ug/Kg	68	0			24	125	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: 8270E DataFile: BF142480.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits		RPD
								Qual	Low	High		
PB168083BS	Naphthalene	1700	1400	ug/Kg	82				62	100		
	Acenaphthylene	1700	1300	ug/Kg	76				63	101		
	Acenaphthene	1700	1500	ug/Kg	88				57	104		
	Fluorene	1700	1300	ug/Kg	76				61	101		
	Phenanthrene	1700	1400	ug/Kg	82				59	103		
	Anthracene	1700	1300	ug/Kg	76				61	105		
	Fluoranthene	1700	1500	ug/Kg	88				57	107		
	Pyrene	1700	1600	ug/Kg	94				59	103		
	Benzo(a)anthracene	1700	1400	ug/Kg	82				60	102		
	Chrysene	1700	1400	ug/Kg	82				59	101		
	Benzo(b)fluoranthene	1700	1500	ug/Kg	88				62	109		
	Benzo(k)fluoranthene	1700	1400	ug/Kg	82				62	109		
	Benzo(a)pyrene	1700	1500	ug/Kg	88				63	103		
	Indeno(1,2,3-cd)pyrene	1700	1600	ug/Kg	94				63	101		
	Dibenz(a,h)anthracene	1700	1600	ug/Kg	94				61	112		
	Benzo(g,h,i)perylene	1700	1600	ug/Kg	94				70	108		

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: 8270E DataFile: BP024755.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168098BS	Naphthalene	50	43.0	ug/L	86				64	107	
	Acenaphthylene	50	42.4	ug/L	85				79	103	
	Acenaphthene	50	42.1	ug/L	84				59	113	
	Fluorene	50	44.1	ug/L	88				64	107	
	Phenanthrene	50	43.2	ug/L	86				62	109	
	Anthracene	50	44.5	ug/L	89				65	110	
	Fluoranthene	50	46.0	ug/L	92				64	110	
	Pyrene	50	44.5	ug/L	89				71	103	
	Benzo(a)anthracene	50	43.9	ug/L	88				62	107	
	Chrysene	50	44.1	ug/L	88				61	108	
	Benzo(b)fluoranthene	50	44.2	ug/L	88				77	113	
	Benzo(k)fluoranthene	50	44.0	ug/L	88				77	105	
	Benzo(a)pyrene	50	44.8	ug/L	90				72	131	
	Indeno(1,2,3-cd)pyrene	50	45.7	ug/L	91				72	105	
	Dibenz(a,h)anthracene	50	46.0	ug/L	92				78	115	
	Benzo(g,h,i)perylene	50	44.9	ug/L	90				75	118	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: 8270E DataFile: BP024756.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	RPD		Limits		RPD
							Qual	Qual	Low	High	
PB168098BSD	Naphthalene	50	43.4	ug/L	87	1			64	107	20
	Acenaphthylene	50	43.6	ug/L	87	3			79	103	20
	Acenaphthene	50	43.5	ug/L	87	3			59	113	20
	Fluorene	50	44.9	ug/L	90	2			64	107	20
	Phenanthrene	50	44.0	ug/L	88	2			62	109	20
	Anthracene	50	45.1	ug/L	90	1			65	110	20
	Fluoranthene	50	45.8	ug/L	92	0			64	110	20
	Pyrene	50	45.6	ug/L	91	2			71	103	20
	Benzo(a)anthracene	50	45.5	ug/L	91	4			62	107	20
	Chrysene	50	44.4	ug/L	89	1			61	108	20
	Benzo(b)fluoranthene	50	46.7	ug/L	93	6			77	113	20
	Benzo(k)fluoranthene	50	45.6	ug/L	91	4			77	105	20
	Benzo(a)pyrene	50	46.1	ug/L	92	3			72	131	20
	Indeno(1,2,3-cd)pyrene	50	44.8	ug/L	90	2			72	105	20
	Dibenz(a,h)anthracene	50	45.2	ug/L	90	2			78	115	20
	Benzo(g,h,i)perylene	50	43.9	ug/L	88	2			75	118	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168083BL

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q2075

SAS No.: Q2075 SDG NO.: Q2075

Lab File ID: BF142479.D

Lab Sample ID: PB168083BL

Instrument ID: BNA_F

Date Extracted: 05/20/2025

Matrix: (soil/water) SOIL

Date Analyzed: 05/21/2025

Level: (low/med) LOW

Time Analyzed: 11:09

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB168083BS	PB168083BS	BF142480.D	05/21/2025
SS-11	Q2075-03	BF142486.D	05/21/2025
SS-11MS	Q2075-04MS	BF142487.D	05/21/2025
SS-11MSD	Q2075-05MSD	BF142488.D	05/21/2025
SS-MW1-11.5	Q2075-06	BF142489.D	05/21/2025
SS-10	Q2075-01	BF142510.D	05/22/2025
SS-910	Q2075-02	BF142511.D	05/22/2025

COMMENTS: _____

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB168098BL

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM Case No.: Q2075

SAS No.: Q2075 SDG NO.: Q2075

Lab File ID: BP024754.D

Lab Sample ID: PB168098BL

Instrument ID: BNA_P

Date Extracted: 05/21/2025

Matrix: (soil/water) Water

Date Analyzed: 05/22/2025

Level: (low/med) LOW

Time Analyzed: 11:37

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB168098BS	PB168098BS	BP024755.D	05/22/2025
PB168098BSD	PB168098BSD	BP024756.D	05/22/2025
FB-05152025	Q2075-07	BP024758.D	05/22/2025
FB-05162025	Q2075-08	BP024748.D	05/21/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2075

SDG NO.: Q2075

Lab File ID: BF142465.D

DFTPP Injection Date: 05/20/2025

Instrument ID: BNA_F

DFTPP Injection Time: 11:13

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	36.8
68	Less than 2.0% of mass 69	0.5 (1.9) 1
69	Mass 69 relative abundance	33.1
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	44.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	29.9
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	19.9
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19 (19) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BF142467.D	05/20/2025	12:10
SSTDICC005	SSTDICC005	BF142468.D	05/20/2025	12:38
SSTDICC010	SSTDICC010	BF142469.D	05/20/2025	13:07
SSTDICC020	SSTDICC020	BF142470.D	05/20/2025	13:36
SSTDICCC040	SSTDICCC040	BF142471.D	05/20/2025	14:05
SSTDICC050	SSTDICC050	BF142472.D	05/20/2025	14:34
SSTDICC060	SSTDICC060	BF142473.D	05/20/2025	15:03
SSTDICC080	SSTDICC080	BF142474.D	05/20/2025	15:31

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2075

SDG NO.: Q2075

Lab File ID: BF142477.D

DFTPP Injection Date: 05/21/2025

Instrument ID: BNA_F

DFTPP Injection Time: 09:41

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	40.0
68	Less than 2.0% of mass 69	0.6 (2) 1
69	Mass 69 relative abundance	34.9
70	Less than 2.0% of mass 69	0.2 (0.7) 1
127	10.0 - 80.0% of mass 198	46.9
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	29.1
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	18.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142478.D	05/21/2025	10:40
PB168083BL	PB168083BL	BF142479.D	05/21/2025	11:09
PB168083BS	PB168083BS	BF142480.D	05/21/2025	11:37
SS-11	Q2075-03	BF142486.D	05/21/2025	14:37
SS-11MS	Q2075-04MS	BF142487.D	05/21/2025	15:06
SS-11MSD	Q2075-05MSD	BF142488.D	05/21/2025	15:34
SS-MW1-11.5	Q2075-06	BF142489.D	05/21/2025	16:03

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2075 SDG NO.: Q2075

Lab File ID: BF142501.D

DFTPP Injection Date: 05/22/2025

Instrument ID: BNA_F

DFTPP Injection Time: 10:43

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	37.5
68	Less than 2.0% of mass 69	0.5 (1.8) 1
69	Mass 69 relative abundance	33.6
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	46.8
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	30.3
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	20.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.7 (18.7) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142502.D	05/22/2025	11:11
SS-10	Q2075-01	BF142510.D	05/22/2025	15:07
SS-910	Q2075-02	BF142511.D	05/22/2025	15:36

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2075

SDG NO.: Q2075

Lab File ID: BP024613.D

DFTPP Injection Date: 05/13/2025

Instrument ID: BNA_P

DFTPP Injection Time: 10:00

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	33.5
68	Less than 2.0% of mass 69	0.2 (0.6) 1
69	Mass 69 relative abundance	36.2
70	Less than 2.0% of mass 69	0.2 (0.6) 1
127	10.0 - 80.0% of mass 198	49.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.6
275	10.0 - 60.0% of mass 198	33.1
365	Greater than 1% of mass 198	5.4
441	Present, but less than mass 443	19.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.4 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP024614.D	05/13/2025	10:41
SSTDICC005	SSTDICC005	BP024615.D	05/13/2025	11:22
SSTDICC010	SSTDICC010	BP024616.D	05/13/2025	12:03
SSTDICC020	SSTDICC020	BP024617.D	05/13/2025	12:43
SSTDICCC040	SSTDICCC040	BP024618.D	05/13/2025	13:24
SSTDICC050	SSTDICC050	BP024619.D	05/13/2025	14:05
SSTDICC060	SSTDICC060	BP024620.D	05/13/2025	14:45
SSTDICC080	SSTDICC080	BP024621.D	05/13/2025	15:26

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2075

SDG NO.: Q2075

Lab File ID: BP024736.D

DFTPP Injection Date: 05/21/2025

Instrument ID: BNA_P

DFTPP Injection Time: 12:54

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	31.7
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	33.9
70	Less than 2.0% of mass 69	0.1 (0.4) 1
127	10.0 - 80.0% of mass 198	47.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	33.3
365	Greater than 1% of mass 198	5.6
441	Present, but less than mass 443	22.0
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.9 (19.9) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024737.D	05/21/2025	14:48
FB-05162025	Q2075-08	BP024748.D	05/21/2025	22:18

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2075 SDG NO.: Q2075

Lab File ID: BP024752.D

DFTPP Injection Date: 05/22/2025

Instrument ID: BNA_P

DFTPP Injection Time: 09:35

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	30.1
68	Less than 2.0% of mass 69	0.3 (1.3) 1
69	Mass 69 relative abundance	32.0
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	46.0
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7.0
275	10.0 - 60.0% of mass 198	33.7
365	Greater than 1% of mass 198	5.4
441	Present, but less than mass 443	22.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.6 (19.6) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024753.D	05/22/2025	10:56
PB168098BL	PB168098BL	BP024754.D	05/22/2025	11:37
PB168098BS	PB168098BS	BP024755.D	05/22/2025	12:18
PB168098BSD	PB168098BSD	BP024756.D	05/22/2025	12:58
FB-05152025	Q2075-07	BP024758.D	05/22/2025	14:19

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/21/2025

Lab File ID: BF142478.D Time Analyzed: 10:40

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	124102	6.898	475212	8.19	258765	9.94
UPPER LIMIT	248204	7.398	950424	8.686	517530	10.439
LOWER LIMIT	62051	6.398	237606	7.686	129383	9.439
EPA SAMPLE NO.						
01 PB168083BL	127003	6.90	492428	8.18	264579	9.94
02 PB168083BS	124824	6.90	495438	8.19	277044	9.94
03 SS-11	63588	6.90	238505	8.18	131818	9.94
04 SS-11MS	54835 *	6.90	200654 *	8.18	113830 *	9.94
05 SS-11MSD	53473 *	6.90	192497 *	8.18	95696 *	9.94
06 SS-MW1-11.5	59263 *	6.90	206974 *	8.18	103588 *	9.94

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/21/2025

Lab File ID: BF142478.D Time Analyzed: 10:40

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	454448	11.427	231233	14.068	233183	15.562
UPPER LIMIT	908896	11.927	462466	14.568	466366	16.062
LOWER LIMIT	227224	10.927	115617	13.568	116592	15.062
EPA SAMPLE NO.						
01 PB168083BL	511963	11.43	324877	14.07	234653	15.56
02 PB168083BS	468851	11.43	214548	14.07	229705	15.56
03 SS-11	250730	11.43	246437	14.07	209013	15.56
04 SS-11MS	229012	11.43	210926	14.07	176162	15.56
05 SS-11MSD	179294 *	11.43	174160	14.07	147670	15.56
06 SS-MW1-11.5	179364 *	11.42	153166	14.06	125915	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/22/2025

Lab File ID: BF142502.D Time Analyzed: 11:11

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	96659	6.898	372701	8.19	197309	9.94
UPPER LIMIT	193318	7.398	745402	8.687	394618	10.439
LOWER LIMIT	48329.5	6.398	186351	7.687	98654.5	9.439
EPA SAMPLE NO.						
01 SS-10	120361	6.90	452635	8.18	223139	9.93
02 SS-910	116551	6.90	418491	8.18	190753	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/22/2025

Lab File ID: BF142502.D Time Analyzed: 11:11

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	315055	11.428	157895	14.069	195527	15.557
UPPER LIMIT	630110	11.928	315790	14.569	391054	16.057
LOWER LIMIT	157528	10.928	78947.5	13.569	97763.5	15.057
EPA SAMPLE NO.						
01 SS-10	308296	11.42	228905	14.06	198076	15.56
02 SS-910	271357	11.42	225312	14.06	176642	15.56

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/21/2025

Lab File ID: BP024737.D Time Analyzed: 14:48

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	128557	7.646	504282	10.42	338214	14.29
UPPER LIMIT	257114	8.146	1008560	10.916	676428	14.787
LOWER LIMIT	64278.5	7.146	252141	9.916	169107	13.787
EPA SAMPLE NO.						
01 FB-05162025	87679	7.65	325514	10.42	217493	14.29

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/21/2025

Lab File ID: BP024737.D Time Analyzed: 14:48

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	643332	17.081	728456	21.51	831576	24.786
UPPER LIMIT	1286660	17.581	1456910	22.01	1663150	25.286
LOWER LIMIT	321666	16.581	364228	21.01	415788	24.286
EPA SAMPLE NO.						
01 FB-05162025	457611	17.09	547413	21.51	671507	24.80

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/22/2025

Lab File ID: BP024753.D Time Analyzed: 10:56

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	103714	7.652	424752	10.42	258861	14.28
UPPER LIMIT	207428	8.152	849504	10.922	517722	14.781
LOWER LIMIT	51857	7.152	212376	9.922	129431	13.781
EPA SAMPLE NO.						
01 PB168098BL	115351	7.65	453620	10.42	285428	14.29
02 PB168098BS	114889	7.65	451979	10.42	283120	14.29
03 PB168098BSD	126327	7.65	523024	10.42	338733	14.29
04 FB-05152025	103638	7.65	395386	10.42	244728	14.29

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075

EPA Sample No.: SSTDCCC040 Date Analyzed: 05/22/2025

Lab File ID: BP024753.D Time Analyzed: 10:56

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	515151	17.075	613668	21.516	759397	24.792
UPPER LIMIT	1030300	17.575	1227340	22.016	1518790	25.292
LOWER LIMIT	257576	16.575	306834	21.016	379699	24.292
EPA SAMPLE NO.						
01 PB168098BL	613349	17.09	715536	21.52	855120	24.82
02 PB168098BS	568580	17.09	655197	21.52	754804	24.82
03 PB168098BSD	686645	17.09	768545	21.51	835982	24.79
04 FB-05152025	486633	17.09	599957	21.52	736321	24.82

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168083BL	SDG No.:	Q2075
Lab Sample ID:	PB168083BL	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142479.D	1	05/20/25 09:45	05/21/25 11:09	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	170	U	22.7	170	ug/Kg
208-96-8	Acenaphthylene	170	U	28.9	170	ug/Kg
83-32-9	Acenaphthene	170	U	21.3	170	ug/Kg
86-73-7	Fluorene	170	U	25.3	170	ug/Kg
85-01-8	Phenanthrene	170	U	20.9	170	ug/Kg
120-12-7	Anthracene	170	U	33.3	170	ug/Kg
206-44-0	Fluoranthene	170	U	30.0	170	ug/Kg
129-00-0	Pyrene	170	U	36.0	170	ug/Kg
56-55-3	Benzo(a)anthracene	170	U	23.0	170	ug/Kg
218-01-9	Chrysene	170	U	19.9	170	ug/Kg
205-99-2	Benzo(b)fluoranthene	170	U	19.0	170	ug/Kg
207-08-9	Benzo(k)fluoranthene	170	U	22.4	170	ug/Kg
50-32-8	Benzo(a)pyrene	170	U	29.5	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	170	U	29.1	170	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	170	U	27.4	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	170	U	25.7	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	67.4		18 - 107	67%	SPK: 100
321-60-8	2-Fluorobiphenyl	69.4		20 - 109	69%	SPK: 100
1718-51-0	Terphenyl-d14	62.4		10 - 105	62%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	127000	6.898			
1146-65-2	Naphthalene-d8	492000	8.181			
15067-26-2	Acenaphthene-d10	265000	9.939			
1517-22-2	Phenanthrene-d10	512000	11.428			
1719-03-5	Chrysene-d12	325000	14.069			
1520-96-3	Perylene-d12	235000	15.563			

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168083BL	SDG No.:	Q2075
Lab Sample ID:	PB168083BL	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142479.D	1	05/20/25 09:45	05/21/25 11:09	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168098BL	SDG No.:	Q2075
Lab Sample ID:	PB168098BL	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024754.D	1	05/21/25 08:40	05/22/25 11:37	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
91-20-3	Naphthalene	5.00	U	0.50	5.00	ug/L
208-96-8	Acenaphthylene	5.00	U	0.75	5.00	ug/L
83-32-9	Acenaphthene	5.00	U	0.55	5.00	ug/L
86-73-7	Fluorene	5.00	U	0.63	5.00	ug/L
85-01-8	Phenanthrene	5.00	U	0.50	5.00	ug/L
120-12-7	Anthracene	5.00	U	0.61	5.00	ug/L
206-44-0	Fluoranthene	5.00	U	0.82	5.00	ug/L
129-00-0	Pyrene	5.00	U	0.50	5.00	ug/L
56-55-3	Benzo(a)anthracene	5.00	U	0.45	5.00	ug/L
218-01-9	Chrysene	5.00	U	0.44	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	5.00	U	0.49	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	5.00	U	0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	5.00	U	0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	5.00	U	0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	5.00	U	0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	5.00	U	0.69	5.00	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	75.0		49 - 133	75%	SPK: 100
321-60-8	2-Fluorobiphenyl	77.3		52 - 132	77%	SPK: 100
1718-51-0	Terphenyl-d14	83.0		48 - 125	83%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	115000		7.652		
1146-65-2	Naphthalene-d8	454000		10.422		
15067-26-2	Acenaphthene-d10	285000		14.287		
1517-22-2	Phenanthrene-d10	613000		17.092		
1719-03-5	Chrysene-d12	716000		21.522		
1520-96-3	Perylene-d12	855000		24.821		

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168098BL	SDG No.:	Q2075
Lab Sample ID:	PB168098BL	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024754.D	1	05/21/25 08:40	05/22/25 11:37	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168083BS	SDG No.:	Q2075
Lab Sample ID:	PB168083BS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142480.D	1	05/20/25 09:45	05/21/25 11:37	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	1400		22.7	170	ug/Kg
208-96-8	Acenaphthylene	1300		28.9	170	ug/Kg
83-32-9	Acenaphthene	1500		21.3	170	ug/Kg
86-73-7	Fluorene	1300		25.3	170	ug/Kg
85-01-8	Phenanthrene	1400		20.9	170	ug/Kg
120-12-7	Anthracene	1300		33.3	170	ug/Kg
206-44-0	Fluoranthene	1500		30.0	170	ug/Kg
129-00-0	Pyrene	1600		36.0	170	ug/Kg
56-55-3	Benzo(a)anthracene	1400		23.0	170	ug/Kg
218-01-9	Chrysene	1400		19.9	170	ug/Kg
205-99-2	Benzo(b)fluoranthene	1500		19.0	170	ug/Kg
207-08-9	Benzo(k)fluoranthene	1400		22.4	170	ug/Kg
50-32-8	Benzo(a)pyrene	1500		29.5	170	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1600		29.1	170	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	1600		27.4	170	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1600		25.7	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	69.4		18 - 107	69%	SPK: 100
321-60-8	2-Fluorobiphenyl	66.0		20 - 109	66%	SPK: 100
1718-51-0	Terphenyl-d14	81.6		10 - 105	82%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	125000		6.898		
1146-65-2	Naphthalene-d8	495000		8.186		
15067-26-2	Acenaphthene-d10	277000		9.939		
1517-22-2	Phenanthrene-d10	469000		11.427		
1719-03-5	Chrysene-d12	215000		14.068		
1520-96-3	Perylene-d12	230000		15.562		

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168083BS	SDG No.:	Q2075
Lab Sample ID:	PB168083BS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142480.D	1	05/20/25 09:45	05/21/25 11:37	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168098BS	SDG No.:	Q2075
Lab Sample ID:	PB168098BS	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024755.D	1	05/21/25 08:40	05/22/25 12:18	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
91-20-3	Naphthalene	43.0		0.50	5.00	ug/L
208-96-8	Acenaphthylene	42.4		0.75	5.00	ug/L
83-32-9	Acenaphthene	42.1		0.55	5.00	ug/L
86-73-7	Fluorene	44.1		0.63	5.00	ug/L
85-01-8	Phenanthrene	43.2		0.50	5.00	ug/L
120-12-7	Anthracene	44.5		0.61	5.00	ug/L
206-44-0	Fluoranthene	46.0		0.82	5.00	ug/L
129-00-0	Pyrene	44.5		0.50	5.00	ug/L
56-55-3	Benzo(a)anthracene	43.9		0.45	5.00	ug/L
218-01-9	Chrysene	44.1		0.44	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	44.2		0.49	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	44.0		0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	44.8		0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	45.7		0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	46.0		0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	44.9		0.69	5.00	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	72.8		49 - 133	73%	SPK: 100
321-60-8	2-Fluorobiphenyl	76.3		52 - 132	76%	SPK: 100
1718-51-0	Terphenyl-d14	77.9		48 - 125	78%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	115000		7.652		
1146-65-2	Naphthalene-d8	452000		10.422		
15067-26-2	Acenaphthene-d10	283000		14.287		
1517-22-2	Phenanthrene-d10	569000		17.087		
1719-03-5	Chrysene-d12	655000		21.522		
1520-96-3	Perylene-d12	755000		24.821		

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168098BS	SDG No.:	Q2075
Lab Sample ID:	PB168098BS	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024755.D	1	05/21/25 08:40	05/22/25 12:18	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168098BSD	SDG No.:	Q2075
Lab Sample ID:	PB168098BSD	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024756.D	1	05/21/25 08:40	05/22/25 12:58	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
91-20-3	Naphthalene	43.4		0.50	5.00	ug/L
208-96-8	Acenaphthylene	43.6		0.75	5.00	ug/L
83-32-9	Acenaphthene	43.5		0.55	5.00	ug/L
86-73-7	Fluorene	44.9		0.63	5.00	ug/L
85-01-8	Phenanthrene	44.0		0.50	5.00	ug/L
120-12-7	Anthracene	45.1		0.61	5.00	ug/L
206-44-0	Fluoranthene	45.8		0.82	5.00	ug/L
129-00-0	Pyrene	45.6		0.50	5.00	ug/L
56-55-3	Benzo(a)anthracene	45.5		0.45	5.00	ug/L
218-01-9	Chrysene	44.4		0.44	5.00	ug/L
205-99-2	Benzo(b)fluoranthene	46.7		0.49	5.00	ug/L
207-08-9	Benzo(k)fluoranthene	45.6		0.48	5.00	ug/L
50-32-8	Benzo(a)pyrene	46.1		0.55	5.00	ug/L
193-39-5	Indeno(1,2,3-cd)pyrene	44.8		0.59	5.00	ug/L
53-70-3	Dibenzo(a,h)anthracene	45.2		0.67	5.00	ug/L
191-24-2	Benzo(g,h,i)perylene	43.9		0.69	5.00	ug/L
SURROGATES						
4165-60-0	Nitrobenzene-d5	73.9		49 - 133	74%	SPK: 100
321-60-8	2-Fluorobiphenyl	76.8		52 - 132	77%	SPK: 100
1718-51-0	Terphenyl-d14	81.5		48 - 125	81%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	126000		7.652		
1146-65-2	Naphthalene-d8	523000		10.416		
15067-26-2	Acenaphthene-d10	339000		14.287		
1517-22-2	Phenanthrene-d10	687000		17.092		
1719-03-5	Chrysene-d12	769000		21.51		
1520-96-3	Perylene-d12	836000		24.786		

Report of Analysis

Client:	CDM Smith	Date Collected:	
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	
Client Sample ID:	PB168098BSD	SDG No.:	Q2075
Lab Sample ID:	PB168098BSD	Matrix:	Water
Analytical Method:	8270E	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BP024756.D	1	05/21/25 08:40	05/22/25 12:58	PB168098

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-11MS	SDG No.:	Q2075
Lab Sample ID:	Q2075-04MS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	87.4
Sample Wt/Vol:	30.07 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142487.D	1	05/20/25 09:45	05/21/25 15:06	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	1600		25.9	190	ug/Kg
208-96-8	Acenaphthylene	1600		33.0	190	ug/Kg
83-32-9	Acenaphthene	1800		24.3	190	ug/Kg
86-73-7	Fluorene	1700		28.9	190	ug/Kg
85-01-8	Phenanthrene	1600		23.9	190	ug/Kg
120-12-7	Anthracene	1700		38.0	190	ug/Kg
206-44-0	Fluoranthene	2200		34.2	190	ug/Kg
129-00-0	Pyrene	1300		41.1	190	ug/Kg
56-55-3	Benzo(a)anthracene	1700		26.3	190	ug/Kg
218-01-9	Chrysene	1600		22.7	190	ug/Kg
205-99-2	Benzo(b)fluoranthene	1900		21.7	190	ug/Kg
207-08-9	Benzo(k)fluoranthene	1600		25.6	190	ug/Kg
50-32-8	Benzo(a)pyrene	1700		33.7	190	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1400		33.2	190	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	1400		31.3	190	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1300		29.3	190	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	60.4		18 - 107	60%	SPK: 100
321-60-8	2-Fluorobiphenyl	51.6		20 - 109	52%	SPK: 100
1718-51-0	Terphenyl-d14	41.5		10 - 105	41%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	54800	6.898			
1146-65-2	Naphthalene-d8	201000	8.181			
15067-26-2	Acenaphthene-d10	114000	9.939			
1517-22-2	Phenanthrene-d10	229000	11.427			
1719-03-5	Chrysene-d12	211000	14.068			
1520-96-3	Perylene-d12	176000	15.563			

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-11MS	SDG No.:	Q2075
Lab Sample ID:	Q2075-04MS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	87.4
Sample Wt/Vol:	30.07 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142487.D	1	05/20/25 09:45	05/21/25 15:06	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-11MSD	SDG No.:	Q2075
Lab Sample ID:	Q2075-05MSD	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	87.4
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142488.D	1	05/20/25 09:45	05/21/25 15:34	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
91-20-3	Naphthalene	1600		25.9	190	ug/Kg
208-96-8	Acenaphthylene	1600		33.0	190	ug/Kg
83-32-9	Acenaphthene	1800		24.3	190	ug/Kg
86-73-7	Fluorene	1600		28.9	190	ug/Kg
85-01-8	Phenanthrene	1700		23.9	190	ug/Kg
120-12-7	Anthracene	1700		38.1	190	ug/Kg
206-44-0	Fluoranthene	2300		34.3	190	ug/Kg
129-00-0	Pyrene	1300		41.1	190	ug/Kg
56-55-3	Benzo(a)anthracene	1600		26.3	190	ug/Kg
218-01-9	Chrysene	1600		22.7	190	ug/Kg
205-99-2	Benzo(b)fluoranthene	1700		21.7	190	ug/Kg
207-08-9	Benzo(k)fluoranthene	1700		25.6	190	ug/Kg
50-32-8	Benzo(a)pyrene	1700		33.7	190	ug/Kg
193-39-5	Indeno(1,2,3-cd)pyrene	1400		33.3	190	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	1300		31.3	190	ug/Kg
191-24-2	Benzo(g,h,i)perylene	1300		29.4	190	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	59.6		18 - 107	60%	SPK: 100
321-60-8	2-Fluorobiphenyl	54.4		20 - 109	54%	SPK: 100
1718-51-0	Terphenyl-d14	40.2		10 - 105	40%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	53500		6.898		
1146-65-2	Naphthalene-d8	192000		8.181		
15067-26-2	Acenaphthene-d10	95700		9.939		
1517-22-2	Phenanthrene-d10	179000		11.427		
1719-03-5	Chrysene-d12	174000		14.068		
1520-96-3	Perylene-d12	148000		15.557		

Report of Analysis

Client:	CDM Smith	Date Collected:	05/15/25
Project:	Con Ed UTEN Mount Vernon, NY	Date Received:	05/16/25
Client Sample ID:	SS-11MSD	SDG No.:	Q2075
Lab Sample ID:	Q2075-05MSD	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	87.4
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group3
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF142488.D	1	05/20/25 09:45	05/21/25 15:34	PB168083

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF052025.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 20 16:26:47 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BF142467.D 5 =BF142468.D 10 =BF142469.D 20 =BF142470.D 40 =BF142471.D 50 =BF142472.D 60 =BF142473.D 80 =BF142474.D

Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I 1,4-Dichlorobenzen...	-----ISTD-----									
2) 1,4-Dioxane		0.493	0.460	0.474	0.458	0.500	0.486	0.456	0.475	3.79
3) Pyridine		1.237	1.150	1.212	1.170	1.273	1.234	1.190	1.210	3.52
4) n-Nitrosodimet...		0.612	0.593	0.627	0.619	0.673	0.652	0.621	0.628	4.21
5) S 2-Fluorophenol		1.264	1.214	1.225	1.125	1.220	1.164	1.098	1.187	5.03
6) Aniline		2.044	1.889	1.963	1.844	1.993	1.910	1.808	1.921	4.35
7) S Phenol-d6		1.530	1.449	1.459	1.367	1.467	1.403	1.328	1.429	4.77
8) 2-Chlorophenol		1.345	1.293	1.315	1.252	1.338	1.285	1.223	1.293	3.44
9) Benzaldehyde		1.035	0.975	0.969	0.817	0.872	0.758	0.591	0.859	17.81
10) C Phenol		1.716	1.621	1.646	1.530	1.657	1.597	1.487	1.608	4.85
11) bis(2-Chloroet...		1.202	1.155	1.168	1.108	1.209	1.156	1.105	1.157	3.52
12) 1,3-Dichlorobe...		1.562	1.470	1.473	1.389	1.482	1.407	1.317	1.443	5.48
13) C 1,4-Dichlorobe...		1.540	1.476	1.491	1.407	1.495	1.430	1.335	1.453	4.70
14) 1,2-Dichlorobe...		1.495	1.405	1.436	1.327	1.432	1.357	1.284	1.391	5.20
15) Benzyl Alcohol		1.059	1.024	1.058	1.021	1.131	1.073	1.026	1.056	3.69
16) 2,2'-oxybis(1-...		2.082	1.978	1.983	1.868	2.011	1.898	1.786	1.944	5.11
17) 2-Methylphenol		1.040	0.992	1.026	0.976	1.053	1.015	0.965	1.010	3.27
18) Hexachloroethane		0.533	0.503	0.523	0.489	0.529	0.497	0.477	0.507	4.23
19) P n-Nitroso-di-n...	0.923	0.941	0.880	0.900	0.843	0.912	0.866	0.819	0.886	4.69
20) 3+4-Methylphenols		1.412	1.319	1.337	1.246	1.337	1.250	1.149	1.293	6.59

21) I Naphthalene-d8	-----ISTD-----									
22) Acetophenone		0.480	0.453	0.459	0.429	0.452	0.428	0.399	0.443	5.98
23) S Nitrobenzene-d5		0.376	0.365	0.379	0.356	0.382	0.363	0.347	0.367	3.51
24) Nitrobenzene		0.338	0.328	0.338	0.323	0.343	0.331	0.316	0.331	2.94
25) Isophorone		0.636	0.615	0.620	0.593	0.638	0.607	0.585	0.613	3.26
26) C 2-Nitrophenol		0.167	0.170	0.180	0.175	0.190	0.182	0.173	0.177	4.44
27) 2,4-Dimethylph...		0.315	0.315	0.318	0.303	0.325	0.312	0.295	0.312	3.22
28) bis(2-Chloroet...		0.406	0.394	0.394	0.364	0.391	0.375	0.358	0.383	4.63
29) C 2,4-Dichloroph...		0.288	0.282	0.290	0.276	0.300	0.283	0.266	0.283	3.74
30) 1,2,4-Trichlor...		0.325	0.313	0.317	0.295	0.320	0.300	0.284	0.308	4.89
31) Naphthalene		1.061	1.021	1.020	0.941	1.007	0.953	0.891	0.985	5.94
32) Benzoic acid			0.153	0.176	0.188	0.211	0.209	0.201	0.190	11.73
33) 4-Chloroaniline		0.424	0.409	0.416	0.389	0.415	0.397	0.343	0.399	6.87
34) C Hexachlorobuta...		0.203	0.192	0.197	0.187	0.198	0.192	0.176	0.192	4.52
35) Caprolactam		0.081	0.076	0.083	0.079	0.085	0.078	0.076	0.080	4.29
36) C 4-Chloro-3-met...		0.304	0.290	0.299	0.282	0.301	0.287	0.271	0.291	4.02
37) 2-Methylnaphth...		0.679	0.638	0.646	0.590	0.631	0.598	0.556	0.620	6.62
38) 1-Methylnaphth...		0.703	0.668	0.672	0.615	0.650	0.611	0.566	0.641	7.19

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
 Method File : 8270-BF052025.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.592	0.580	0.588	0.552	0.592	0.573	0.544	0.574	3.39	
41) P	Hexachlorocycl...	0.318	0.350	0.383	0.393	0.430	0.425	0.411	0.387	10.57	
42) S	2,4,6-Tribromo...	0.230	0.225	0.234	0.211	0.233	0.218	0.202	0.222	5.39	
43) C	2,4,6-Trichlor...	0.387	0.373	0.406	0.371	0.406	0.388	0.379	0.387	3.69	
44)	2,4,5-Trichlor...	0.410	0.411	0.416	0.395	0.437	0.408	0.381	0.408	4.27	
45) S	2-Fluorobiphenyl	1.726	1.619	1.558	1.387	1.480	1.396	1.268	1.490	10.46	
46)	1,1'-Biphenyl	1.672	1.605	1.595	1.480	1.595	1.507	1.408	1.552	5.82	
47)	2-Chloronaphth...	1.218	1.170	1.177	1.094	1.183	1.122	1.061	1.146	4.84	
48)	2-Nitroaniline	0.333	0.318	0.338	0.324	0.354	0.332	0.320	0.331	3.72	
49)	Acenaphthylene	2.064	1.982	2.027	1.851	1.998	1.885	1.745	1.936	5.85	
50)	Dimethylphthalate	1.441	1.340	1.367	1.255	1.366	1.256	1.211	1.320	6.16	
51)	2,6-Dinitrotol...	0.290	0.283	0.289	0.280	0.302	0.281	0.268	0.285	3.74	
52) C	Acenaphthene	1.259	1.222	1.227	1.120	1.223	1.146	1.077	1.182	5.72	
53)	3-Nitroaniline	0.320	0.309	0.327	0.299	0.330	0.305	0.287	0.311	5.02	
54) P	2,4-Dinitrophenol		0.110	0.137	0.149	0.169	0.160	0.158	0.147	14.36	
55)	Dibenzofuran	1.886	1.766	1.768	1.606	1.736	1.635	1.509	1.701	7.38	
56) P	4-Nitrophenol	0.221	0.217	0.242	0.227	0.252	0.229	0.220	0.230	5.57	
57)	2,4-Dinitrotol...	0.372	0.367	0.387	0.364	0.398	0.372	0.344	0.372	4.62	
58)	Fluorene	1.477	1.399	1.372	1.231	1.343	1.238	1.140	1.314	8.85	
59)	2,3,4,6-Tetrac...	0.347	0.336	0.360	0.333	0.361	0.333	0.317	0.341	4.71	
60)	Diethylphthalate	1.422	1.310	1.359	1.231	1.343	1.218	1.140	1.289	7.51	
61)	4-Chlorophenyl...	0.724	0.678	0.676	0.609	0.653	0.612	0.566	0.646	8.25	
62)	4-Nitroaniline	0.303	0.283	0.303	0.277	0.294	0.265	0.251	0.282	6.95	
63)	Azobenzene	1.239	1.194	1.188	1.107	1.197	1.123	1.055	1.158	5.55	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...	0.086	0.094	0.110	0.115	0.129	0.125	0.123	0.112	14.72	
66) c	n-Nitrosodiphe...	0.712	0.682	0.696	0.649	0.697	0.694	0.660	0.684	3.30	
67)	4-Bromophenyl-...	0.240	0.235	0.239	0.222	0.246	0.243	0.230	0.236	3.45	
68)	Hexachlorobenzene	0.265	0.267	0.263	0.251	0.273	0.262	0.250	0.262	3.19	
69)	Atrazine	0.184	0.174	0.186	0.178	0.196	0.181	0.174	0.182	4.29	
70) C	Pentachlorophenol	0.130	0.135	0.149	0.147	0.162	0.157	0.154	0.148	7.88	
71)	Phenanthrene	1.196	1.094	1.100	1.012	1.085	1.033	0.964	1.069	7.00	
72)	Anthracene	1.191	1.132	1.124	1.036	1.110	1.048	0.996	1.091	6.15	
73)	Carbazole	1.030	0.968	0.982	0.889	0.950	0.877	0.832	0.933	7.39	
74)	Di-n-butylphth...	1.108	1.039	1.083	0.976	1.059	0.974	0.908	1.021	6.95	
75) C	Fluoranthene	1.177	1.081	1.052	0.938	0.997	0.901	0.846	0.999	11.41	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine	0.670	0.782	0.903	0.797	0.805	0.698	0.556	0.744	15.12	
78)	Pyrene	1.929	1.967	2.066	1.859	1.959	1.773	1.507	1.866	9.79	
79) S	Terphenyl-d14	1.624	1.582	1.660	1.423	1.492	1.337	1.126	1.464	12.80	
80)	Butylbenzylphth...	0.462	0.453	0.525	0.526	0.575	0.550	0.517	0.516	8.54	
81)	Benzo(a)anthra...	1.424	1.287	1.403	1.278	1.391	1.327	1.238	1.336	5.36	
82)	3,3'-Dichlorob...	0.360	0.364	0.402	0.407	0.444	0.442	0.417	0.405	8.30	
83)	Chrysene	1.222	1.204	1.193	1.145	1.255	1.223	1.158	1.200	3.20	
84)	Bis(2-ethylhex...	0.510	0.548	0.618	0.673	0.765	0.778	0.727	0.660	15.91	
85) c	Di-n-octyl pht...		0.958	1.108	1.266	1.432	1.542	1.437	1.290	17.30	

Method Path : Z:\svoasrv\HPCHEM1\BNA_F\Methods\
Method File : 8270-BF052025.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.421	1.495	1.583	1.448	1.583	1.546	1.434	1.501	4.64
88)	Benzo(b)fluora...	1.317	1.105	1.293	1.126	1.209	1.122	1.114	1.184	7.62
89)	Benzo(k)fluora...	1.191	1.166	1.032	1.042	1.178	1.128	1.023	1.109	6.68
90) C	Benzo(a)pyrene	1.154	1.081	1.114	1.081	1.185	1.133	1.062	1.116	4.00
91)	Dibenzo(a,h)an...	1.152	1.228	1.275	1.184	1.290	1.245	1.149	1.218	4.67
92)	Benzo(g,h,i)pe...	1.154	1.220	1.270	1.180	1.299	1.245	1.158	1.218	4.64

(#) = Out of Range

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
 Method File : 8270E-BP051325.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue May 13 16:21:39 2025
 Response Via : Initial Calibration

Calibration Files

2.5 =BP024614.D 5 =BP024615.D 10 =BP024616.D 20 =BP024617.D 40 =BP024618.D 50 =BP024619.D 60 =BP024620.D 80 =BP024621.D

Compound	2.5	5	10	20	40	50	60	80	Avg	%RSD

1) I 1,4-Dichlorobenzen...	-----ISTD-----									
2) 1,4-Dioxane	0.641	0.557	0.535	0.505	0.562	0.530	0.513	0.549		8.31
3) Pyridine	1.054	1.135	1.202	1.198	1.363	1.307	1.263	1.218		8.57
4) n-Nitrosodimet...	0.536	0.507	0.527	0.517	0.580	0.557	0.537	0.538		4.61
5) S 2-Fluorophenol	1.134	1.096	1.162	1.105	1.272	1.223	1.193	1.169		5.50
6) Aniline	1.717	1.703	1.932	1.917	2.130	2.091	1.999	1.927		8.67
7) S Phenol-d6	1.301	1.344	1.486	1.472	1.671	1.620	1.552	1.492		9.09
8) 2-Chlorophenol	1.200	1.204	1.311	1.303	1.465	1.433	1.376	1.327		7.83
9) Benzaldehyde	0.991	1.008	1.046	0.955	1.043	0.954	0.766	0.966		9.91
10) C Phenol	1.373	1.401	1.526	1.509	1.707	1.670	1.597	1.540		8.24
11) bis(2-Chloroet...	1.317	1.170	1.257	1.213	1.357	1.299	1.247	1.266		5.04
12) 1,3-Dichlorobe...	1.596	1.487	1.520	1.440	1.622	1.546	1.480	1.527		4.29
13) C 1,4-Dichlorobe...	1.558	1.505	1.525	1.470	1.637	1.567	1.491	1.536		3.67
14) 1,2-Dichlorobe...	1.530	1.467	1.498	1.426	1.579	1.529	1.449	1.497		3.56
15) Benzyl Alcohol	0.883	0.950	1.128	1.149	1.306	1.288	1.236	1.134		14.43
16) 2,2'-oxybis(1-...	1.543	1.445	1.520	1.467	1.617	1.564	1.454	1.516		4.22
17) 2-Methylphenol	0.966	0.949	1.105	1.099	1.246	1.213	1.154	1.105		10.30
18) Hexachloroethane	0.606	0.591	0.582	0.551	0.623	0.597	0.570	0.589		4.04
19) P n-Nitroso-di-n...	0.880	0.977	0.967	1.086	1.048	1.142	1.132	1.026	1.032	8.63
20) 3+4-Methylphenols	1.221	1.302	1.503	1.515	1.702	1.692	1.584	1.503		12.21
21) I Naphthalene-d8	-----ISTD-----									
22) Acetophenone	0.486	0.481	0.511	0.483	0.534	0.512	0.491	0.500		3.97
23) S Nitrobenzene-d5	0.381	0.376	0.401	0.382	0.429	0.408	0.394	0.396		4.67
24) Nitrobenzene	0.333	0.340	0.355	0.337	0.373	0.356	0.344	0.348		4.03
25) Isophorone	0.644	0.640	0.695	0.673	0.753	0.721	0.682	0.687		5.91
26) C 2-Nitrophenol	0.135	0.143	0.168	0.172	0.196	0.192	0.189	0.171		14.05
27) 2,4-Dimethylph...	0.273	0.279	0.300	0.298	0.329	0.318	0.309	0.301		6.60
28) bis(2-Chloroet...	0.403	0.391	0.411	0.399	0.443	0.419	0.405	0.410		4.15
29) C 2,4-Dichloroph...	0.250	0.258	0.290	0.291	0.327	0.316	0.309	0.292		9.96
30) 1,2,4-Trichlor...	0.346	0.322	0.327	0.306	0.347	0.329	0.323	0.329		4.33
31) Naphthalene	1.064	1.026	1.049	0.980	1.091	1.037	1.007	1.036		3.54
32) Benzoic acid		0.089	0.146	0.191	0.215	0.220	0.225	0.181		29.55
33) 4-Chloroaniline	0.348	0.371	0.420	0.418	0.465	0.462	0.442	0.418		10.62
34) C Hexachlorobuta...	0.231	0.214	0.211	0.198	0.222	0.209	0.206	0.213		5.08
35) Caprolactam	0.082	0.092	0.104	0.109	0.121	0.121	0.110	0.106		13.80
36) C 4-Chloro-3-met...	0.316	0.325	0.358	0.356	0.391	0.384	0.359	0.356		7.78
37) 2-Methylnaphth...	0.673	0.654	0.670	0.640	0.713	0.693	0.654	0.671		3.73
38) 1-Methylnaphth...	0.741	0.700	0.720	0.690	0.751	0.733	0.690	0.718		3.47

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
 Method File : 8270E-BP051325.M

39) I	Acenaphthene-d10	-----ISTD-----									
40)	1,2,4,5-Tetrac...	0.593	0.558	0.574	0.549	0.625	0.580	0.586	0.581	4.26	
41) P	Hexachlorocycl...	0.250	0.276	0.350	0.356	0.420	0.391	0.421	0.352	19.02	
42) S	2,4,6-Tribromo...	0.324	0.321	0.335	0.323	0.369	0.357	0.346	0.339	5.49	
43) C	2,4,6-Trichlor...	0.305	0.341	0.375	0.371	0.428	0.410	0.402	0.376	11.29	
44)	2,4,5-Trichlor...	0.376	0.375	0.418	0.404	0.471	0.447	0.440	0.419	8.68	
45) S	2-Fluorobiphenyl	1.601	1.498	1.553	1.432	1.602	1.515	1.485	1.527	4.10	
46)	1,1'-Biphenyl	1.524	1.434	1.492	1.383	1.586	1.473	1.456	1.478	4.40	
47)	2-Chloronaphth...	1.147	1.106	1.131	1.065	1.196	1.132	1.113	1.127	3.56	
48)	2-Nitroaniline	0.279	0.304	0.331	0.335	0.381	0.364	0.343	0.334	10.28	
49)	Acenaphthylene	1.845	1.818	1.920	1.804	2.029	1.930	1.856	1.886	4.19	
50)	Dimethylphthalate	1.570	1.527	1.518	1.450	1.611	1.535	1.460	1.524	3.73	
51)	2,6-Dinitrotol...	0.299	0.308	0.326	0.312	0.354	0.337	0.319	0.322	5.75	
52) C	Acenaphthene	1.105	1.054	1.094	1.028	1.159	1.099	1.052	1.084	4.02	
53)	3-Nitroaniline	0.258	0.275	0.329	0.340	0.379	0.370	0.350	0.329	14.04	
54) P	2,4-Dinitrophenol		0.098	0.143	0.172	0.201	0.201	0.201	0.169	24.79	
55)	Dibenzofuran	1.810	1.716	1.740	1.633	1.816	1.733	1.658	1.730	4.00	
56) P	4-Nitrophenol		0.159	0.222	0.257	0.299	0.288	0.274	0.250	20.82	
57)	2,4-Dinitrotol...	0.400	0.422	0.456	0.454	0.508	0.483	0.454	0.454	7.87	
58)	Fluorene	1.466	1.415	1.410	1.336	1.494	1.410	1.341	1.410	4.14	
59)	2,3,4,6-Tetrac...	0.367	0.364	0.381	0.379	0.428	0.406	0.394	0.388	5.89	
60)	Diethylphthalate	1.593	1.542	1.537	1.444	1.639	1.563	1.468	1.541	4.40	
61)	4-Chlorophenyl...	0.752	0.697	0.700	0.656	0.741	0.709	0.668	0.703	5.01	
62)	4-Nitroaniline	0.228	0.259	0.326	0.330	0.375	0.363	0.341	0.317	17.06	
63)	Azobenzene	1.327	1.299	1.356	1.250	1.407	1.343	1.243	1.318	4.45	
64) I	Phenanthrene-d10	-----ISTD-----									
65)	4,6-Dinitro-2-...		0.094	0.120	0.129	0.149	0.143	0.141	0.129	15.64	
66) c	n-Nitrosodiphe...	0.603	0.589	0.627	0.593	0.667	0.621	0.600	0.614	4.41	
67)	4-Bromophenyl-...	0.231	0.223	0.237	0.226	0.252	0.242	0.240	0.236	4.31	
68)	Hexachlorobenzene	0.303	0.288	0.298	0.282	0.321	0.304	0.300	0.299	4.25	
69)	Atrazine	0.210	0.209	0.227	0.218	0.248	0.237	0.226	0.225	6.32	
70) C	Pentachlorophenol	0.124	0.137	0.163	0.172	0.198	0.191	0.192	0.168	17.12	
71)	Phenanthrene	1.144	1.085	1.125	1.051	1.170	1.125	1.083	1.112	3.68	
72)	Anthracene	1.097	1.064	1.142	1.075	1.205	1.139	1.108	1.119	4.31	
73)	Carbazole	0.923	0.944	1.039	1.001	1.114	1.053	1.013	1.012	6.43	
74)	Di-n-butylphth...	1.182	1.218	1.332	1.260	1.447	1.363	1.313	1.302	6.97	
75) C	Fluoranthene	1.271	1.254	1.314	1.250	1.399	1.337	1.272	1.300	4.18	
76) I	Chrysene-d12	-----ISTD-----									
77)	Benzidine		0.362	0.543	0.560	0.657	0.599	0.548	0.545	18.22	
78)	Pyrene	1.184	1.137	1.197	1.153	1.273	1.222	1.223	1.198	3.85	
79) S	Terphenyl-d14	1.192	1.113	1.165	1.125	1.226	1.175	1.167	1.166	3.30	
80)	Butylbenzylpht...	0.458	0.465	0.537	0.527	0.604	0.571	0.569	0.533	10.28	
81)	Benzo(a)anthra...	1.251	1.218	1.253	1.208	1.336	1.275	1.250	1.256	3.36	
82)	3,3'-Dichlorob...	0.357	0.392	0.474	0.477	0.539	0.511	0.514	0.466	14.45	
83)	Chrysene	1.214	1.180	1.185	1.131	1.255	1.188	1.180	1.190	3.16	
84)	Bis(2-ethylhex...	0.673	0.683	0.801	0.766	0.884	0.836	0.841	0.783	10.32	
85) c	Di-n-octyl pht...	1.048	1.116	1.272	1.280	1.458	1.376	1.417	1.281	11.95	

Method Path : Z:\svoasrv\HPCHEM1\BNA_P\Methods\
Method File : 8270E-BP051325.M

86) I	Perylene-d12	-----ISTD-----								
87)	Indeno(1,2,3-c...	1.341	1.357	1.447	1.424	1.608	1.549	1.494	1.460	6.70
88)	Benzo(b)fluora...	1.028	1.070	1.152	1.103	1.288	1.184	1.189	1.145	7.58
89)	Benzo(k)fluora...	1.147	1.106	1.182	1.140	1.258	1.238	1.185	1.180	4.59
90) C	Benzo(a)pyrene	0.995	1.001	1.112	1.064	1.221	1.167	1.136	1.099	7.68
91)	Dibenzo(a,h)an...	1.068	1.066	1.186	1.167	1.321	1.271	1.236	1.188	8.18
92)	Benzo(g,h,i)pe...	1.101	1.106	1.167	1.147	1.294	1.242	1.212	1.181	6.07

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG No.: Q2075

Instrument ID: BNA_F Calibration Date/Time: 05/21/2025 10:40

Lab File ID: BF142478.D Init. Calib. Date(s): 05/20/2025 05/20/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:10 15:31

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.187	1.087		-8.4	
Phenol-d6	1.429	1.358		-5.0	
Nitrobenzene-d5	0.367	0.343		-6.5	
Naphthalene	0.985	0.928		-5.8	
2-Fluorobiphenyl	1.490	1.373		-7.9	
Acenaphthylene	1.936	1.795		-7.3	
Acenaphthene	1.182	1.125		-4.8	20.0
Fluorene	1.314	1.229		-6.5	
2,4,6-Tribromophenol	0.222	0.225		1.4	
Phenanthrene	1.069	0.989		-7.5	
Anthracene	1.091	1.022		-6.3	
Fluoranthene	0.999	0.984		-1.5	20.0
Pyrene	1.866	1.932		3.5	
Terphenyl-d14	1.464	1.466		0.1	
Benzo (a) anthracene	1.336	1.289		-3.5	
Chrysene	1.200	1.116		-7.0	
Benzo (b) fluoranthene	1.184	1.216		2.7	
Benzo (k) fluoranthene	1.109	0.996		-10.2	
Benzo (a) pyrene	1.116	1.061		-4.9	20.0
Indeno (1,2,3-cd) pyrene	1.501	1.409		-6.1	
Dibenzo (a,h) anthracene	1.218	1.156		-5.1	
Benzo (g,h,i) perylene	1.218	1.202		-1.3	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG No.: Q2075

Instrument ID: BNA_F Calibration Date/Time: 05/22/2025 11:11

Lab File ID: BF142502.D Init. Calib. Date(s): 05/20/2025 05/20/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 12:10 15:31

GC Column: DB-UI ID: 0.18 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.187	1.158		-2.4	
Phenol-d6	1.429	1.379		-3.5	
Nitrobenzene-d5	0.367	0.356		-3.0	
Naphthalene	0.985	0.939		-4.7	
2-Fluorobiphenyl	1.490	1.389		-6.8	
Acenaphthylene	1.936	1.864		-3.7	
Acenaphthene	1.182	1.118		-5.4	20.0
Fluorene	1.314	1.233		-6.2	
2,4,6-Tribromophenol	0.222	0.208		-6.3	
Phenanthrene	1.069	1.009		-5.6	
Anthracene	1.091	1.031		-5.5	
Fluoranthene	0.999	0.878		-12.1	20.0
Pyrene	1.866	1.705		-8.6	
Terphenyl-d14	1.464	1.259		-14.0	
Benzo (a) anthracene	1.336	1.258		-5.8	
Chrysene	1.200	1.136		-5.3	
Benzo (b) fluoranthene	1.184	1.087		-8.2	
Benzo (k) fluoranthene	1.109	1.044		-5.9	
Benzo (a) pyrene	1.116	1.078		-3.4	20.0
Indeno (1,2,3-cd) pyrene	1.501	1.416		-5.7	
Dibenzo (a,h) anthracene	1.218	1.143		-6.2	
Benzo (g,h,i) perylene	1.218	1.144		-6.1	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG No.: Q2075

Instrument ID: BNA_P Calibration Date/Time: 05/21/2025 14:48

Lab File ID: BP024737.D Init. Calib. Date(s): 05/13/2025 05/13/2025

EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:41 15:26

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.169	1.095		-6.3	
Phenol-d6	1.492	1.334		-10.6	
Nitrobenzene-d5	0.396	0.360		-9.1	
Naphthalene	1.036	1.000		-3.5	
2-Fluorobiphenyl	1.527	1.480		-3.1	
Acenaphthylene	1.886	1.807		-4.2	
Acenaphthene	1.084	1.033		-4.7	20.0
Fluorene	1.410	1.271		-9.9	
2,4,6-Tribromophenol	0.339	0.356		5.0	
Phenanthrene	1.112	1.063		-4.4	
Anthracene	1.119	1.090		-2.6	
Fluoranthene	1.300	1.252		-3.7	20.0
Pyrene	1.198	1.164		-2.8	
Terphenyl-d14	1.166	1.170		0.3	
Benzo (a) anthracene	1.256	1.204		-4.1	
Chrysene	1.190	1.132		-4.9	
Benzo (b) fluoranthene	1.145	1.108		-3.2	
Benzo (k) fluoranthene	1.180	1.122		-4.9	
Benzo (a) pyrene	1.099	1.059		-3.6	20.0
Indeno (1,2,3-cd) pyrene	1.460	1.437		-1.6	
Dibenzo (a,h) anthracene	1.188	1.179		-0.8	
Benzo (g,h,i) perylene	1.181	1.143		-3.2	

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG No.: Q2075
Instrument ID: BNA_P Calibration Date/Time: 05/22/2025 10:56
Lab File ID: BP024753.D Init. Calib. Date(s): 05/13/2025 05/13/2025
EPA Sample No.: SSTDCCC040 Init. Calib. Time(s): 10:41 15:26
GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF040	MIN RRF	%D	MAX%D
2-Fluorophenol	1.169	1.135		-2.9	
Phenol-d6	1.492	1.407		-5.7	
Nitrobenzene-d5	0.396	0.368		-7.1	
Naphthalene	1.036	0.988		-4.6	
2-Fluorobiphenyl	1.527	1.502		-1.6	
Acenaphthylene	1.886	1.818		-3.6	
Acenaphthene	1.084	1.055		-2.7	20.0
Fluorene	1.410	1.365		-3.2	
2,4,6-Tribromophenol	0.339	0.364		7.4	
Phenanthrene	1.112	1.056		-5.0	
Anthracene	1.119	1.093		-2.3	
Fluoranthene	1.300	1.270		-2.3	20.0
Pyrene	1.198	1.114		-7.0	
Terphenyl-d14	1.166	1.110		-4.8	
Benzo (a) anthracene	1.256	1.196		-4.8	
Chrysene	1.190	1.132		-4.9	
Benzo (b) fluoranthene	1.145	1.090		-4.8	
Benzo (k) fluoranthene	1.180	1.078		-8.6	
Benzo (a) pyrene	1.099	1.059		-3.6	20.0
Indeno (1,2,3-cd) pyrene	1.460	1.467		0.5	
Dibenzo (a,h) anthracene	1.188	1.197		0.8	
Benzo (g,h,i) perylene	1.181	1.175		-0.5	

All other compounds must meet a minimum RRF of 0.010.