

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: 8270E DataFile: BF142480.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB168083BS	Naphthalene	1700	1400	ug/Kg	82				62	100	
	Acenaphthylene	1700	1300	ug/Kg	76				63	101	
	Acenaphthene	1700	1500	ug/Kg	88				57	104	
	Fluorene	1700	1300	ug/Kg	76				61	101	
	Phenanthrene	1700	1400	ug/Kg	82				59	103	
	Anthracene	1700	1300	ug/Kg	76				61	105	
	Fluoranthene	1700	1500	ug/Kg	88				57	107	
	Pyrene	1700	1600	ug/Kg	94				59	103	
	Benzo(a)anthracene	1700	1400	ug/Kg	82				60	102	
	Chrysene	1700	1400	ug/Kg	82				59	101	
	Benzo(b)fluoranthene	1700	1500	ug/Kg	88				62	109	
	Benzo(k)fluoranthene	1700	1400	ug/Kg	82				62	109	
	Benzo(a)pyrene	1700	1500	ug/Kg	88				63	103	
	Indeno(1,2,3-cd)pyrene	1700	1600	ug/Kg	94				63	101	
	Dibenz(a,h)anthracene	1700	1600	ug/Kg	94				61	112	
	Benzo(g,h,i)perylene	1700	1600	ug/Kg	94				70	108	