

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: 8270E DataFile: BP024755.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB168098BS	Naphthalene	50	43.0	ug/L	86				64	107	
	Acenaphthylene	50	42.4	ug/L	85				79	103	
	Acenaphthene	50	42.1	ug/L	84				59	113	
	Fluorene	50	44.1	ug/L	88				64	107	
	Phenanthrene	50	43.2	ug/L	86				62	109	
	Anthracene	50	44.5	ug/L	89				65	110	
	Fluoranthene	50	46.0	ug/L	92				64	110	
	Pyrene	50	44.5	ug/L	89				71	103	
	Benzo(a)anthracene	50	43.9	ug/L	88				62	107	
	Chrysene	50	44.1	ug/L	88				61	108	
	Benzo(b)fluoranthene	50	44.2	ug/L	88				77	113	
	Benzo(k)fluoranthene	50	44.0	ug/L	88				77	105	
	Benzo(a)pyrene	50	44.8	ug/L	90				72	131	
	Indeno(1,2,3-cd)pyrene	50	45.7	ug/L	91				72	105	
	Dibenz(a,h)anthracene	50	46.0	ug/L	92				78	115	
Benzo(g,h,i)perylene	50	44.9	ug/L	90				75	118		