

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK  
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: CAMP02

Lab Code: CHEM

SAS No.: Q2075

SDG NO.: Q2075

Lab File ID: BF142501.D

DFTPP Injection Date: 05/22/2025

Instrument ID: BNA\_F

DFTPP Injection Time: 10:43

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	37.5
68	Less than 2.0% of mass 69	0.5 ( 1.8 ) 1
69	Mass 69 relative abundance	33.6
70	Less than 2.0% of mass 69	0.2 ( 0.6 ) 1
127	10.0 - 80.0% of mass 198	46.8
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	30.3
365	Greater than 1% of mass 198	4.1
441	Present, but less than mass 443	20.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	18.7 ( 18.7 ) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF142502.D	05/22/2025	11:11
SS-10	Q2075-01	BF142510.D	05/22/2025	15:07
SS-910	Q2075-02	BF142511.D	05/22/2025	15:36