

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG No.: Q2075

Instrument ID: MSVOA_X Calibration Date/Time: 05/23/2025 09:24

Lab File ID: VX046331.D Init. Calib. Date(s): 05/05/2025 05/05/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:35 16:27

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.417	1.421		0.28	20
Toluene	0.869	0.878		1.04	20
Ethyl Benzene	1.929	1.985		2.9	20
m/p-Xylenes	0.706	0.719		1.84	20
o-Xylene	0.688	0.714		3.78	20
Isopropylbenzene	3.893	3.986		2.39	20
n-propylbenzene	4.526	4.678		3.36	20
1,3,5-Trimethylbenzene	3.252	3.374		3.75	20
tert-Butylbenzene	3.276	3.334		1.77	20
1,2,4-Trimethylbenzene	3.293	3.388		2.88	20
sec-Butylbenzene	4.022	4.204		4.53	20
p-Isopropyltoluene	3.320	3.484		4.94	20
n-Butylbenzene	2.912	3.156		8.38	20
1,2-Dichloroethane-d4	0.932	0.966		3.65	20
Dibromofluoromethane	0.360	0.384		6.67	20
Toluene-d8	1.246	1.307		4.9	20
4-Bromofluorobenzene	0.478	0.514		7.53	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.