

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG No.: Q2075

Instrument ID: MSVOA_Y Calibration Date/Time: 05/20/2025 08:40

Lab File ID: VY022329.D Init. Calib. Date(s): 05/15/2025 05/15/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 11:47

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.457	1.473		1.1	20
Toluene	0.923	0.912		-1.19	20
Ethyl Benzene	2.113	2.120		0.33	20
m/p-Xylenes	0.796	0.803		0.88	20
o-Xylene	0.753	0.738		-1.99	20
Isopropylbenzene	4.142	4.147		0.12	20
n-propylbenzene	4.996	5.062		1.32	20
1,3,5-Trimethylbenzene	3.344	3.297		-1.41	20
tert-Butylbenzene	2.972	2.924		-1.62	20
1,2,4-Trimethylbenzene	3.319	3.232		-2.62	20
sec-Butylbenzene	4.412	4.438		0.59	20
p-Isopropyltoluene	3.710	3.717		0.19	20
n-Butylbenzene	3.510	3.562		1.48	20
1,2-Dichloroethane-d4	0.546	0.533		-2.38	20
Dibromofluoromethane	0.298	0.305		2.35	20
Toluene-d8	1.222	1.259		3.03	20
4-Bromofluorobenzene	0.387	0.375		-3.1	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.