

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CAMP02

Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG No.: Q2075

Instrument ID: MSVOA_Y Calibration Date/Time: 05/22/2025 08:35

Lab File ID: VY022380.D Init. Calib. Date(s): 05/15/2025 05/15/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:46 11:47

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.457	1.548		6.25	20
Toluene	0.923	0.978		5.96	20
Ethyl Benzene	2.113	2.234		5.73	20
m/p-Xylenes	0.796	0.849		6.66	20
o-Xylene	0.753	0.791		5.05	20
Isopropylbenzene	4.142	4.432		7	20
n-propylbenzene	4.996	5.385		7.79	20
1,3,5-Trimethylbenzene	3.344	3.533		5.65	20
tert-Butylbenzene	2.972	3.200		7.67	20
1,2,4-Trimethylbenzene	3.319	3.526		6.24	20
sec-Butylbenzene	4.412	4.835		9.59	20
p-Isopropyltoluene	3.710	4.077		9.89	20
n-Butylbenzene	3.510	3.887		10.74	20
1,2-Dichloroethane-d4	0.546	0.561		2.75	20
Dibromofluoromethane	0.298	0.318		6.71	20
Toluene-d8	1.222	1.298		6.22	20
4-Bromofluorobenzene	0.387	0.387		0	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.