

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075
Lab File ID: VY022380.D Date Analyzed: 05/22/2025
Instrument ID: MSVOA_Y Time Analyzed: 08:35
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	161495	7.72	274168	8.62	235238	11.42
UPPER LIMIT	322990	8.22	548336	9.122	470476	11.92
LOWER LIMIT	80747.5	7.22	137084	8.122	117619	10.92
EPA SAMPLE NO.						
SS-MW1-11.5	234597	7.72	407953	8.62	330547	11.42
VY0522SBL01	263712	7.72	465826	8.62	376365	11.42
VY0522SBS01	174604	7.72	293942	8.62	250842	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.