

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075
 Lab File ID: VY022329.D Date Analyzed: 05/20/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:40
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	120747	13.346				
UPPER LIMIT	241494	13.846				
LOWER LIMIT	60373.5	12.846				
EPA SAMPLE NO.						
SS-10	138928	13.35				
SS-910	126954	13.35				
SS-11	130644	13.35				
SS-11MS	91777	13.35				
SS-11MSD	91227	13.35				
VY0520SBL01	124571	13.35				
VY0520SBS01	111469	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.