

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075
 Lab File ID: VY022380.D Date Analyzed: 05/22/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:35
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	110586	13.353				
UPPER LIMIT	221172	13.853				
LOWER LIMIT	55293	12.853				
EPA SAMPLE NO.						
SS-MW1-11.5	155760	13.35				
VY0522SBL01	137350	13.35				
VY0522SBS01	118480	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.