

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\X050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	50.000	55.594	-11.2	94	0.00
3 P	Chloromethane	50.000	52.592	-5.2	96	0.00
4 C	Vinyl Chloride	50.000	51.598	-3.2#	96	0.00
5 T	Bromomethane	50.000	50.627	-1.3	95	0.00
6 T	Chloroethane	50.000	52.271	-4.5	97	0.00
7 T	Trichlorofluoromethane	50.000	52.560	-5.1	96	0.00
8 T	Diethyl Ether	50.000	49.536	0.9	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	50.822	-1.6	96	0.00
10 T	Methyl Iodide	50.000	53.419	-6.8	94	0.00
11 T	Tert butyl alcohol	250.000	252.025	-0.8	98	0.00
12 CM	1,1-Dichloroethene	50.000	51.097	-2.2#	96	0.00
13 T	Acrolein	250.000	232.505	7.0	87	0.00
14 T	Allyl chloride	50.000	51.833	-3.7	95	0.00
15 T	Acrylonitrile	250.000	262.412	-5.0	97	0.00
16 T	Acetone	250.000	245.518	1.8	97	0.00
17 T	Carbon Disulfide	50.000	50.981	-2.0	94	0.00
18 T	Methyl Acetate	50.000	49.968	0.1	97	0.00
19 T	Methyl tert-butyl Ether	50.000	50.777	-1.6	93	0.00
20 T	Methylene Chloride	50.000	47.779	4.4	95	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.513	-1.0	94	0.00
22 T	Diisopropyl ether	50.000	51.800	-3.6	95	0.00
23 T	Vinyl Acetate	250.000	265.585	-6.2	95	0.00
24 P	1,1-Dichloroethane	50.000	50.825	-1.7	94	0.00
25 T	2-Butanone	250.000	257.110	-2.8	96	0.00
26 T	2,2-Dichloropropane	50.000	49.934	0.1	95	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.956	-1.9	95	0.00
28 T	Bromochloromethane	50.000	50.975	-2.0	99	0.00
29 T	Tetrahydrofuran	250.000	259.958	-4.0	96	0.00
30 C	Chloroform	50.000	51.332	-2.7#	96	0.00
31 T	Cyclohexane	50.000	50.147	-0.3	94	0.00
32 T	1,1,1-Trichloroethane	50.000	51.108	-2.2	95	0.00
33 S	1,2-Dichloroethane-d4	50.000	48.509	3.0	95	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	49.683	0.6	98	0.00
36 T	1,1-Dichloropropene	50.000	49.521	1.0	94	0.00
37 T	Ethyl Acetate	50.000	50.714	-1.4	96	0.00
38 T	Carbon Tetrachloride	50.000	49.825	0.3	94	0.00
39 T	Methylcyclohexane	50.000	49.757	0.5	94	0.00
40 TM	Benzene	50.000	50.759	-1.5	95	0.00
41 T	Methacrylonitrile	50.000	55.076	-10.2	96	0.00
42 TM	1,2-Dichloroethane	50.000	50.861	-1.7	96	0.00
43 T	Isopropyl Acetate	50.000	52.195	-4.4	96	0.00
44 TM	Trichloroethene	50.000	50.018	-0.0	93	0.00
45 C	1,2-Dichloropropane	50.000	51.983	-4.0#	96	0.00
46 T	Dibromomethane	50.000	49.978	0.0	94	0.00
47 T	Bromodichloromethane	50.000	51.889	-3.8	95	0.00
48 T	Methyl methacrylate	50.000	54.226	-8.5	97	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046048.D
 Acq On : 05 May 2025 16:50
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ICVVX050525

Quant Time: May 06 07:17:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	1,4-Dioxane	1000.000	1059.599	-6.0	100	0.00
50 S	Toluene-d8	50.000	48.984	2.0	97	0.00
51 T	4-Methyl-2-Pentanone	250.000	259.133	-3.7	96	0.00
52 CM	Toluene	50.000	51.150	-2.3#	96	0.00
53 T	t-1,3-Dichloropropene	50.000	52.920	-5.8	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	52.892	-5.8	95	0.00
55 T	1,1,2-Trichloroethane	50.000	50.887	-1.8	95	0.00
56 T	Ethyl methacrylate	50.000	53.918	-7.8	96	0.00
57 T	1,3-Dichloropropane	50.000	49.793	0.4	95	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	265.780	-6.3	93	0.00
59 T	2-Hexanone	250.000	268.218	-7.3	98	0.00
60 T	Dibromochloromethane	50.000	52.075	-4.2	95	0.00
61 T	1,2-Dibromoethane	50.000	51.050	-2.1	94	0.00
62 S	4-Bromofluorobenzene	50.000	48.596	2.8	96	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	99	0.00
64 T	Tetrachloroethene	50.000	46.795	6.4	88	0.00
65 PM	Chlorobenzene	50.000	48.699	2.6	96	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	49.557	0.9	94	0.00
67 C	Ethyl Benzene	50.000	50.245	-0.5#	95	0.00
68 T	m/p-Xylenes	100.000	101.022	-1.0	95	0.00
69 T	o-Xylene	50.000	50.841	-1.7	95	0.00
70 T	Styrene	50.000	51.907	-3.8	95	0.00
71 P	Bromoform	50.000	49.583	0.8	91	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	50.733	-1.5	96	0.00
74 T	N-ethyl acetate	50.000	52.407	-4.8	98	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.827	4.3	98	0.00
76 T	1,2,3-Trichloropropane	50.000	48.568	2.9	99	0.00
77 T	Bromobenzene	50.000	48.549	2.9	95	0.00
78 T	n-propylbenzene	50.000	50.562	-1.1	95	0.00
79 T	2-Chlorotoluene	50.000	49.107	1.8	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.500	-3.0	96	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.276	1.4	95	0.00
82 T	4-Chlorotoluene	50.000	50.661	-1.3	96	0.00
83 T	tert-Butylbenzene	50.000	50.511	-1.0	97	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.485	-3.0	97	0.00
85 T	sec-Butylbenzene	50.000	51.122	-2.2	96	0.00
86 T	p-Isopropyltoluene	50.000	51.198	-2.4	96	0.00
87 T	1,3-Dichlorobenzene	50.000	50.272	-0.5	98	0.00
88 T	1,4-Dichlorobenzene	50.000	48.368	3.3	97	0.00
89 T	n-Butylbenzene	50.000	51.804	-3.6	96	0.00
90 T	Hexachloroethane	50.000	49.895	0.2	94	0.00
91 T	1,2-Dichlorobenzene	50.000	49.361	1.3	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	51.458	-2.9	97	0.00
93 T	1,2,4-Trichlorobenzene	50.000	49.662	0.7	97	0.00
94 T	Hexachlorobutadiene	50.000	48.408	3.2	95	0.00
95 T	Naphthalene	50.000	50.404	-0.8	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046048.D
Acq On : 05 May 2025 16:50
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ICVVX050525

Quant Time: May 06 07:17:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	48.525	3.0	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6