

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	97076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	167947	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	146257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	67976	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	88371	30.601	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	61.200%#		
35) Dibromofluoromethane	5.379	113	59550	31.633	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	63.260%#		
50) Toluene-d8	8.647	98	205479	32.395	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	64.780%#		
62) 4-Bromofluorobenzene	11.079	95	79012	34.229	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	68.460%#		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	83826	39.120	ug/l	100
3) Chloromethane	1.307	50	75238	35.658	ug/l	100
4) Vinyl Chloride	1.374	62	68901	37.273	ug/l	100
5) Bromomethane	1.593	94	31648	33.765	ug/l	100
6) Chloroethane	1.666	64	36722	38.558	ug/l	100
7) Trichlorofluoromethane	1.874	101	103693	36.824	ug/l	100
8) Diethyl Ether	2.136	74	32724	35.350	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.319	101	62251	36.997	ug/l	100
10) Methyl Iodide	2.447	142	78277	39.622	ug/l	100
11) Tert butyl alcohol	2.977	59	62557	181.745	ug/l	100
12) 1,1-Dichloroethene	2.313	96	58334	35.610	ug/l	100
13) Acrolein	2.233	56	73927	184.907	ug/l	100
14) Allyl chloride	2.654	41	114445	36.971	ug/l	100
15) Acrylonitrile	3.062	53	188304	182.874	ug/l	100
16) Acetone	2.386	43	175777	177.910	ug/l	100
17) Carbon Disulfide	2.502	76	141281	37.223	ug/l	100
18) Methyl Acetate	2.703	43	82347	34.915	ug/l	100
19) Methyl tert-butyl Ether	3.111	73	209665	36.695	ug/l	100
20) Methylene Chloride	2.782	84	66412	33.182	ug/l	100
21) trans-1,2-Dichloroethene	3.087	96	59234	35.421	ug/l	100
22) Diisopropyl ether	3.757	45	221172	38.241	ug/l	100
23) Vinyl Acetate	3.721	43	994236	194.528	ug/l	100
24) 1,1-Dichloroethane	3.605	63	122601	36.067	ug/l	100
25) 2-Butanone	4.556	43	269224	188.783	ug/l	100
26) 2,2-Dichloropropane	4.465	77	92857	36.059	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	71518	35.466	ug/l	100
28) Bromochloromethane	4.898	49	56091	31.031	ug/l	100
29) Tetrahydrofuran	5.001	42	170124	183.679	ug/l	100
30) Chloroform	5.087	83	125850	35.626	ug/l	100
31) Cyclohexane	5.464	56	109459	38.205	ug/l	100
32) 1,1,1-Trichloroethane	5.373	97	109781	36.034	ug/l	100
36) 1,1-Dichloropropene	5.684	75	83215	38.073	ug/l	100
37) Ethyl Acetate	4.715	43	102537	37.394	ug/l	100
38) Carbon Tetrachloride	5.672	117	93712	37.320	ug/l	100
39) Methylcyclohexane	7.373	83	107651	39.307	ug/l	100
40) Benzene	6.031	78	247476	36.530	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
 Data File : VX046042.D
 Acq On : 05 May 2025 11:58
 Operator : JC/MD
 Sample : VSTDICCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/06/2025
 Supervised By : Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 06:04:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	58082	38.123	ug/l	100
42) 1,2-Dichloroethane	6.080	62	105332	37.605	ug/l	100
43) Isopropyl Acetate	6.336	43	161787	38.598	ug/l	100
44) Trichloroethene	7.123	130	59623	37.230	ug/l	100
45) 1,2-Dichloropropane	7.428	63	62387	37.064	ug/l	100
46) Dibromomethane	7.574	93	48201	36.326	ug/l	100
47) Bromodichloromethane	7.818	83	96916	37.685	ug/l	100
48) Methyl methacrylate	7.690	41	84277	39.157	ug/l	100
49) 1,4-Dioxane	7.659	88	30529	713.477	ug/l	100
51) 4-Methyl-2-Pentanone	8.574	43	532388	193.653	ug/l	100
52) Toluene	8.714	92	150757	37.530	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	88655	41.056	ug/l	100
54) cis-1,3-Dichloropropene	8.360	75	97031	38.521	ug/l	100
55) 1,1,2-Trichloroethane	9.147	97	59499	36.648	ug/l	100
56) Ethyl methacrylate	9.116	69	99947	39.726	ug/l	100
57) 1,3-Dichloropropane	9.305	76	104606	36.347	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	257678	227.902	ug/l	100
59) 2-Hexanone	9.427	43	396784	190.168	ug/l	100
60) Dibromochloromethane	9.519	129	67198	38.027	ug/l	100
61) 1,2-Dibromoethane	9.604	107	62633	37.567	ug/l	100
64) Tetrachloroethene	9.269	164	54853	36.726	ug/l	100
65) Chlorobenzene	10.079	112	160600	35.686	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	57066	38.017	ug/l	100
67) Ethyl Benzene	10.189	91	295712	38.543	ug/l	100
68) m/p-Xylenes	10.299	106	216578	78.142	ug/l	100
69) o-Xylene	10.640	106	106335	38.064	ug/l	100
70) Styrene	10.653	104	178313	39.888	ug/l	100
71) Bromoform	10.799	173	44486	39.457	ug/l #	100
73) Isopropylbenzene	10.957	105	280719	37.865	ug/l	100
74) N-amyl acetate	10.842	43	140509	38.012	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	90975	34.495	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	80678m	28.468	ug/l	
77) Bromobenzene	11.195	156	63053	36.763	ug/l	100
78) n-propylbenzene	11.299	91	329981	39.707	ug/l	100
79) 2-Chlorotoluene	11.360	91	203509	36.336	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	237066	38.516	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	26179	38.445	ug/l	100
82) 4-Chlorotoluene	11.451	91	233172	37.880	ug/l	100
83) tert-Butylbenzene	11.713	119	233468	38.622	ug/l	100
84) 1,2,4-Trimethylbenzene	11.750	105	239433	38.946	ug/l	100
85) sec-Butylbenzene	11.890	105	291074	38.786	ug/l	100
86) p-Isopropyltoluene	12.006	119	241658	39.879	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	115651	36.199	ug/l	100
88) 1,4-Dichlorobenzene	12.036	146	115097	36.763	ug/l	100
89) n-Butylbenzene	12.329	91	213899	40.603	ug/l	100
90) Hexachloroethane	12.536	117	42314	37.534	ug/l	100
91) 1,2-Dichlorobenzene	12.335	146	115286	37.047	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	12.939	75	21909	36.496	ug/l	100
93) 1,2,4-Trichlorobenzene	13.585	180	66655	39.117	ug/l	100
94) Hexachlorobutadiene	13.725	225	29007	36.395	ug/l	100
95) Naphthalene	13.774	128	245568	39.575	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	69262	38.765	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX050525\
Data File : VX046042.D
Acq On : 05 May 2025 11:58
Operator : JC/MD
Sample : VSTDICCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025

Quant Time: May 06 06:09:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 06:04:56 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:168 Resp: 97070
Ion Ratio Lower Upper
168 100
99 68.6 54.9 82.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#2
Dichlorodifluoromethane
Concen: 39.120 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 85 Resp: 83826
Ion Ratio Lower Upper
85 100
87 31.4 15.7 47.1





#3

Chloromethane

Concen: 35.658 ug/l

RT: 1.307 min Scan# 30

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 50 Resp: 75238

Ion Ratio Lower Upper

50 100





52 31.8 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#4

Vinyl Chloride

Concen: 37.273 ug/l

RT: 1.374 min Scan# 47

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58

Tgt Ion: 62 Resp: 68901

Ion Ratio Lower Upper

62 100


64 31.5 25.2 37.8





#5
Bromomethane
Concen: 33.765 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 94 Resp: 31648
Ion Ratio Lower Upper
94 100
96 91.0 72.8 109.2




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#6
Chloroethane
Concen: 38.558 ug/l
RT: 1.666 min Scan# 95
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 64 Resp: 36722
Ion Ratio Lower Upper
64 100
66 30.4 24.3 36.5





#7
 Trichlorofluoromethane
 Concen: 36.824 ug/l
 RT: 1.874 min Scan# 129
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050



Tgt Ion:101 Resp: 10369
 Ion Ratio Lower Upper
 101 100
 103 63.7 51.0 76.4




Manual Integrations
 APPROVED

Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025



#8
 Diethyl Ether
 Concen: 35.350 ug/l
 RT: 2.136 min Scan# 172
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion: 74 Resp: 32724
 Ion Ratio Lower Upper
 74 100
 45 109.9 54.9 164.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 36.997 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:101 Resp: 6225
Ion Ratio Lower Upper
101 100
85 48.3 38.6 58.0
151 69.0 55.2 82.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#10
Methyl Iodide
Concen: 39.622 ug/l
RT: 2.447 min Scan# 223
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion:142 Resp: 78277
Ion Ratio Lower Upper
142 100
127 52.1 41.7 62.5
141 14.4 11.5 17.3





#11
Tert butyl alcohol
Concen: 181.745 ug/l
RT: 2.977 min Scan# 310
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 59 Resp: 6255
Ion Ratio Lower Upper
59 100
57 10.7 8.6 12.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#12
1,1-Dichloroethene
Concen: 35.610 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Tgt Ion: 96 Resp: 58334
Ion Ratio Lower Upper
96 100
61 182.7 146.2 219.2
98 63.8 51.0 76.6





#13
Acrolein
Concen: 184.907 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 56 Resp: 7392
Ion Ratio Lower Upper
56 100
55 70.3 56.2 84.4




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#14
Allyl chloride
Concen: 36.971 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 41 Resp: 114445
Ion Ratio Lower Upper
41 100
39 75.7 60.6 90.8
76 31.1 24.9 37.3





#15
Acrylonitrile
Concen: 182.874 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 53 Resp: 188304
Ion Ratio Lower Upper
53 100
52 81.6 65.3 97.9
51 37.3 29.8 44.8




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#16
Acetone
Concen: 177.910 ug/l
RT: 2.386 min Scan# 213
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 175777
Ion Ratio Lower Upper
43 100
58 26.5 21.2 31.8





#17

Carbon Disulfide

Concen: 37.223 ug/l

RT: 2.502 min Scan# 21

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 14128

Ion Ratio Lower Upper

76 100





78 8.7 7.0 10.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#18

Methyl Acetate

Concen: 34.915 ug/l

RT: 2.703 min Scan# 265

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 43 Resp: 82347

Ion Ratio Lower Upper

43 100

74 20.9 16.7 25.1









#19
Methyl tert-butyl Ether
Concen: 36.695 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 73 Resp: 20966
Ion Ratio Lower Upper
73 100
57 22.1 17.7 26.5




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#20
Methylene Chloride
Concen: 33.182 ug/l
RT: 2.782 min Scan# 278
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 84 Resp: 66412
Ion Ratio Lower Upper
84 100
49 142.4 113.9 170.9
51 41.9 33.5 50.3
86 67.3 53.8 80.8





#21
trans-1,2-Dichloroethene
Concen: 35.421 ug/l
RT: 3.087 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 96 Resp: 59234
Ion Ratio Lower Upper
96 100
61 149.4 119.5 179.3
98 62.5 50.0 75.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#22
Diisopropyl ether
Concen: 38.241 ug/l
RT: 3.757 min Scan# 438
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 45 Resp: 221172
Ion Ratio Lower Upper
45 100
43 83.3 66.6 100.0
87 24.7 19.8 29.6
59 10.7 8.6 12.8





#23
 Vinyl Acetate
 Concen: 194.528 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58



Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050




Tgt Ion: 43 Resp: 994230
 Ion Ratio Lower Upper
 43 100
 86 9.4 7.5 11.3

Manual Integrations
 APPROVED



Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025




#24
 1,1-Dichloroethane
 Concen: 36.067 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58



Tgt Ion: 63 Resp: 122601
 Ion Ratio Lower Upper
 63 100
 98 7.2 3.6 10.8
 100 4.2 2.1 6.3








#25
2-Butanone
Concen: 188.783 ug/l
RT: 4.556 min Scan# 569
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 269224
Ion Ratio Lower Upper
43 100
72 23.0 18.4 27.6




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#26
2,2-Dichloropropane
Concen: 36.059 ug/l
RT: 4.465 min Scan# 554
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 77 Resp: 92857
Ion Ratio Lower Upper
77 100
97 21.0 10.5 31.5





#27
 cis-1,2-Dichloroethene
 Concen: 35.466 ug/l
 RT: 4.483 min Scan# 557
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58



Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050



Tgt Ion: 96 Resp: 71518
 Ion Ratio Lower Upper
 96 100
 61 161.4 0.0 322.8
 98 64.5 0.0 129.0




Manual Integrations
 APPROVED


Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025



#28
 Bromochloromethane
 Concen: 31.031 ug/l
 RT: 4.898 min Scan# 625
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58

Tgt Ion: 49 Resp: 56091
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 4.0
 130 70.2 56.2 84.2





#29

Tetrahydrofuran

Concen: 183.679 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 42 Resp: 170124

Ion Ratio Lower Upper

42 100

72 42.1 33.7 50.5





71 39.3 31.4 47.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#30

Chloroform

Concen: 35.626 ug/l

RT: 5.087 min Scan# 656

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 83 Resp: 125850

Ion Ratio Lower Upper

83 100


85 61.6 49.3 73.9





#31
Cyclohexane
Concen: 38.205 ug/l
RT: 5.464 min Scan# 718
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 56 Resp: 109459
Ion Ratio Lower Upper
56 100
69 30.5 24.4 36.6
84 83.6 66.9 100.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#32
1,1,1-Trichloroethane
Concen: 36.034 ug/l
RT: 5.373 min Scan# 703
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Tgt Ion: 97 Resp: 109781
Ion Ratio Lower Upper
97 100
99 64.7 51.8 77.6
61 50.1 40.1 60.1





#33

1,2-Dichloroethane-d4

Concen: 30.601 ug/l

RT: 5.952 min Scan# 798

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 8837

Ion Ratio Lower Upper

65 100





67 49.5 0.0 99.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:114 Resp: 167947


Ion Ratio Lower Upper

114 100

63 24.6 0.0 49.2


88 16.8 0.0 33.6





#35
Dibromofluoromethane
Concen: 31.633 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 113 Resp: 59550
Ion Ratio Lower Upper
113 100
111 103.9 83.1 124.7
192 16.6 13.3 19.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#36
1,1-Dichloropropene
Concen: 38.073 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 75 Resp: 83215
Ion Ratio Lower Upper
75 100
110 32.6 16.3 48.9
77 30.4 24.3 36.5





#37
Ethyl Acetate
Concen: 37.394 ug/l
RT: 4.715 min Scan# 595
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 10253
Ion Ratio Lower Upper
43 100
61 12.9 10.3 15.5
70 9.9 7.9 11.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#38
Carbon Tetrachloride
Concen: 37.320 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:117 Resp: 93712
Ion Ratio Lower Upper
117 100
119 94.0 75.2 112.8
121 30.3 24.2 36.4





#39

Methylcyclohexane

Concen: 39.307 ug/l

RT: 7.373 min Scan# 1031

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 83 Resp: 10765

Ion Ratio Lower Upper

83 100

55 80.9 64.7 97.1





98 46.8 37.4 56.2

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#40

Benzene

Concen: 36.530 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58

Tgt Ion: 78 Resp: 247476

Ion Ratio Lower Upper

78 100

77 23.7 19.0 28.4









#41
Methacrylonitrile
Concen: 38.123 ug/l
RT: 4.916 min Scan# 612
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 41 Resp: 5808
Ion Ratio Lower Upper
41 100
39 59.0 47.2 70.8
67 63.4 50.7 76.1
52 30.8 24.6 37.0




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#42
1,2-Dichloroethane
Concen: 37.605 ug/l
RT: 6.080 min Scan# 819
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 62 Resp: 105332
Ion Ratio Lower Upper
62 100
98 7.6 0.0 15.2





#43
Isopropyl Acetate
Concen: 38.598 ug/l
RT: 6.336 min Scan# 861
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 43 Resp: 16178
Ion Ratio Lower Upper
43 100
61 17.9 14.3 21.5
87 11.9 9.5 14.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#44
Trichloroethene
Concen: 37.230 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Tgt Ion:130 Resp: 59623
Ion Ratio Lower Upper
130 100
95 102.1 0.0 204.2









#45
1,2-Dichloropropane
Concen: 37.064 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 63 Resp: 6238
Ion Ratio Lower Upper
63 100
65 30.1 24.1 36.1




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#46
Dibromomethane
Concen: 36.326 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 93 Resp: 48201
Ion Ratio Lower Upper
93 100
95 82.0 65.6 98.4
174 85.2 68.2 102.2








#47
Bromodichloromethane
Concen: 37.685 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 83 Resp: 96910
Ion Ratio Lower Upper
83 100
85 63.1 50.5 75.7
127 8.1 6.5 9.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#48
Methyl methacrylate
Concen: 39.157 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 41 Resp: 84277
Ion Ratio Lower Upper
41 100
69 73.1 58.5 87.7
39 64.6 51.7 77.5








#49
1,4-Dioxane
Concen: 713.477 ug/l
RT: 7.659 min Scan# 1078
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 88 Resp: 30529
Ion Ratio Lower Upper
88 100
43 41.8 33.4 50.2
58 73.3 58.6 88.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#50
Toluene-d8
Concen: 32.395 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 98 Resp: 205479
Ion Ratio Lower Upper
98 100
100 66.9 53.5 80.3








#51
4-Methyl-2-Pentanone
Concen: 193.653 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 43 Resp: 53238
Ion Ratio Lower Upper
43 100
58 36.1 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#52
Toluene
Concen: 37.530 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 92 Resp: 150757
Ion Ratio Lower Upper
92 100
91 170.8 136.6 205.0





#53

t-1,3-Dichloropropene

Concen: 41.056 ug/l

RT: 8.976 min Scan# 1294

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 75 Resp: 8865

Ion Ratio Lower Upper

75 100





77 31.3 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#54

cis-1,3-Dichloropropene

Concen: 38.521 ug/l

RT: 8.360 min Scan# 1193

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion: 75 Resp: 97031

Ion Ratio Lower Upper

75 100

77 31.7 25.4 38.0


39 65.3 52.2 78.4





#55
1,1,2-Trichloroethane
Concen: 36.648 ug/l
RT: 9.147 min Scan# 1322
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion	Ratio	Lower	Upper
97	100		
83	86.3	69.0	103.6
85	55.4	44.3	66.5
99	59.7	47.8	71.6

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#56
Ethyl methacrylate
Concen: 39.726 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Tgt Ion	Ratio	Lower	Upper
69	100		
41	76.0	60.8	91.2
39	48.8	39.0	58.6





#57

1,3-Dichloropropane

Concen: 36.347 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 104600

Ion Ratio Lower Upper

76 100





78 32.9 26.3 39.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#58

2-Chloroethyl Vinyl ether

Concen: 227.902 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58


Tgt Ion: 63 Resp: 257678

Ion Ratio Lower Upper

63 100


106 26.9 21.5 32.3





#59
2-Hexanone
Concen: 190.168 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 396784
Ion Ratio Lower Upper
43 100
58 49.7 24.9 74.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#60
Dibromochloromethane
Concen: 38.027 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:129 Resp: 67198
Ion Ratio Lower Upper
129 100
127 78.5 39.3 117.8









#61
1,2-Dibromoethane
Concen: 37.567 ug/l
RT: 9.604 min Scan# 1397
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:107 Resp: 6263
Ion Ratio Lower Upper
107 100
109 91.6 73.3 109.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#62
4-Bromofluorobenzene
Concen: 34.229 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion: 95 Resp: 79012
Ion Ratio Lower Upper
95 100
174 67.9 0.0 135.8
176 65.7 0.0 131.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.049 min Scan# 1470
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion:117 Resp: 14625
Ion Ratio Lower Upper
117 100
82 63.3 50.6 76.0
119 32.2 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#64
Tetrachloroethene
Concen: 36.726 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:164 Resp: 54853
Ion Ratio Lower Upper
164 100
166 131.3 105.0 157.6
129 104.4 83.5 125.3
131 95.6 76.5 114.7









#65
Chlorobenzene
Concen: 35.686 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:112 Resp: 160600
Ion Ratio Lower Upper
112 100
114 31.8 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 38.017 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion:131 Resp: 57066
Ion Ratio Lower Upper
131 100
133 94.6 47.3 141.9
119 63.3 31.6 95.0





#67
Ethyl Benzene
Concen: 38.543 ug/l
RT: 10.189 min Scan# 1493
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion: 91 Resp: 295712
Ion Ratio Lower Upper
91 100
106 29.3 23.4 35.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025




#68
m/p-Xylenes
Concen: 78.142 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Tgt Ion:106 Resp: 216578
Ion Ratio Lower Upper
106 100
91 214.0 171.2 256.8









#69
o-Xylene
Concen: 38.064 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:106 Resp: 106333
Ion Ratio Lower Upper
106 100
91 225.4 112.7 338.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#70
Styrene
Concen: 39.888 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion:104 Resp: 178313
Ion Ratio Lower Upper
104 100
78 57.1 45.7 68.5
103 54.6 43.7 65.5





#71

Bromoform

Concen: 39.457 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:173 Resp: 44480
 Ion Ratio Lower Upper
 173 100
 175 46.7 23.4 70.0
 254 0.0 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.018 min Scan# 1793

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:152 Resp: 67976
 Ion Ratio Lower Upper
 152 100
 115 93.8 46.9 140.7
 150 175.5 0.0 351.0









#73
Isopropylbenzene
Concen: 37.865 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion:105 Resp: 280719
Ion Ratio Lower Upper
105 100
120 25.6 12.8 38.4




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#74
N-ethyl acetate
Concen: 38.012 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 43 Resp: 140509
Ion Ratio Lower Upper
43 100
70 38.6 30.9 46.3
55 23.4 18.7 28.1
61 21.4 17.1 25.7









#75
1,1,2,2-Tetrachloroethane
Concen: 34.495 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 83 Resp: 90975
Ion Ratio Lower Upper
83 100
131 9.9 5.0 14.9
85 63.8 31.9 95.7




Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#76
1,2,3-Trichloropropane
Concen: 28.468 ug/l m
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 75 Resp: 80678
Ion Ratio Lower Upper
75 100
77 41.0 20.5 61.5





#77

Bromobenzene

Concen: 36.763 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050





Tgt Ion:	156	Resp:	6305
Ion Ratio	Lower	Upper	
156	100		
77	185.5	92.8	278.3
158	97.5	48.8	146.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#78

n-propylbenzene

Concen: 39.707 ug/l




RT: 11.299 min Scan# 1675


Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Tgt Ion:	91	Resp:	329981
Ion Ratio	Lower	Upper	
91	100		
120	21.6	10.8	32.4









#79
2-Chlorotoluene
Concen: 36.336 ug/l
RT: 11.360 min Scan# 1685
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 203509
Ion Ratio Lower Upper
91 100
126 31.1 15.6 46.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#80
1,3,5-Trimethylbenzene
Concen: 38.516 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58


Tgt Ion:105 Resp: 237066
Ion Ratio Lower Upper
105 100
120 46.1 23.1 69.2





#81
trans-1,4-Dichloro-2-butene
Concen: 38.445 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050



Tgt Ion: 75 Resp: 26179
Ion Ratio Lower Upper
75 100
53 125.6 100.5 150.7
89 46.3 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#82
4-Chlorotoluene
Concen: 37.880 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58

Tgt Ion: 91 Resp: 233172
Ion Ratio Lower Upper
91 100
126 26.5 13.3 39.8





#83

tert-Butylbenzene

Concen: 38.622 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:119 Resp: 233468

Ion Ratio Lower Upper

119 100

91 65.8 32.9 98.7





134 22.7 11.4 34.1

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#84

1,2,4-Trimethylbenzene

Concen: 38.946 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58

Tgt Ion:105 Resp: 239433

Ion Ratio Lower Upper

105 100


120 42.4 21.2 63.6





#85
 sec-Butylbenzene
 Concen: 38.786 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58



Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDICCC050




Tgt Ion:105 Resp: 291074
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.7 29.1

Manual Integrations
 APPROVED



Reviewed By :John Carlone 05/06/2025
 Supervised By :Mahesh Dadoda 05/06/2025



#86
 p-Isopropyltoluene
 Concen: 39.879 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX046042.D
 Acq: 05 May 2025 11:58



Tgt Ion:119 Resp: 241658
 Ion Ratio Lower Upper
 119 100
 134 25.0 12.5 37.5
 91 27.6 13.8 41.4





#87

1,3-Dichlorobenzene

Concen: 36.199 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion:146 Resp: 11565

Ion Ratio Lower Upper

146 100

111 44.2 22.1 66.3





148 64.3 32.1 96.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#88

1,4-Dichlorobenzene

Concen: 36.763 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58




Tgt Ion:146 Resp: 115097


Ion Ratio Lower Upper

146 100

111 42.6 21.3 63.9

148 63.7 31.9 95.5





#89

n-Butylbenzene

Concen: 40.603 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min

Lab File: VX046042.D

Acq: 05 May 2025 11:58

Instrument :

MSVOA_X

ClientSampleId :

VSTDICCC050

Tgt Ion: 91 Resp: 21389

Ion Ratio Lower Upper

91 100

92 53.8 26.9 80.7





134 23.5 11.8 35.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 05/06/2025

Supervised By :Mahesh Dadoda 05/06/2025



#90

Hexachloroethane

Concen: 37.534 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX046042.D




Acq: 05 May 2025 11:58

Tgt Ion:117 Resp: 42314

Ion Ratio Lower Upper

117 100


201 63.1 31.6 94.7





#91
1,2-Dichlorobenzene
Concen: 37.047 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:146 Resp: 115280
Ion Ratio Lower Upper
146 100
111 44.9 22.4 67.3
148 63.2 31.6 94.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 36.496 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Tgt Ion: 75 Resp: 21909
Ion Ratio Lower Upper
75 100
155 69.9 34.9 104.8
157 87.6 43.8 131.4





#93
1,2,4-Trichlorobenzene
Concen: 39.117 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:180 Resp: 6665
Ion Ratio Lower Upper
180 100
182 95.9 47.9 143.8
145 34.1 17.1 51.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#94
Hexachlorobutadiene
Concen: 36.395 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Tgt Ion:225 Resp: 29007
Ion Ratio Lower Upper
225 100
223 61.6 30.8 92.4
227 61.8 30.9 92.7





#95
Naphthalene
Concen: 39.575 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58




Instrument :
MSVOA_X
ClientSampleId :
VSTDICCC050




Tgt Ion:128 Resp: 245568
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.8 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 05/06/2025
Supervised By :Mahesh Dadoda 05/06/2025



#96
1,2,3-Trichlorobenzene
Concen: 38.765 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. 0.000 min
Lab File: VX046042.D
Acq: 05 May 2025 11:58



Tgt Ion:180 Resp: 69262
Ion Ratio Lower Upper
180 100
182 95.5 47.8 143.3
145 36.2 18.1 54.3

