

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
 Data File : VY022254.D
 Acq On : 15 May 2025 10:16
 Operator : SY/MD
 Sample : VSTDIC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC010

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025

Quant Time: May 16 01:21:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:19:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	265034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	446109	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	377211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	171972	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	29608	10.222	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	20.440%#		
35) Dibromofluoromethane	7.634	113	26242	9.856	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	19.720%#		
50) Toluene-d8	10.109	98	107048	9.816	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	19.640%#		
62) 4-Bromofluorobenzene	12.408	95	34423	9.959	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	19.920%#		
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.867	85	31097	11.245	ug/l	99
3) Chloromethane	2.068	50	65589	11.275	ug/l	98
4) Vinyl Chloride	2.202	62	69800	10.739	ug/l	95
5) Bromomethane	2.592	94	54773	10.352	ug/l	99
6) Chloroethane	2.732	64	43122	10.665	ug/l	99
7) Trichlorofluoromethane	3.056	101	68501	10.410	ug/l	99
8) Diethyl Ether	3.458	74	16945	10.730	ug/l	96
9) 1,1,2-Trichlorotrifluo...	3.812	101	30591	10.708	ug/l	99
10) Methyl Iodide	4.007	142	31088	9.402	ug/l	97
11) Tert butyl alcohol	4.878	59	12043	54.446	ug/l #	100
12) 1,1-Dichloroethene	3.793	96	28831	10.170	ug/l	95
13) Acrolein	3.659	56	17927	55.439	ug/l	96
14) Allyl chloride	4.378	41	51275	10.315	ug/l	99
15) Acrylonitrile	5.061	53	33277	51.139	ug/l	100
16) Acetone	3.879	43	30229	55.657	ug/l	100
17) Carbon Disulfide	4.104	76	96591	10.448	ug/l	97
18) Methyl Acetate	4.385	43	16338	9.886	ug/l	94
19) Methyl tert-butyl Ether	5.116	73	80714	10.326	ug/l	96
20) Methylene Chloride	4.622	84	42991	12.250	ug/l	95
21) trans-1,2-Dichloroethene	5.116	96	32258	10.309	ug/l	98
22) Diisopropyl ether	6.018	45	106493	10.145	ug/l	94
23) Vinyl Acetate	5.964	43	324651	51.192	ug/l	100
24) 1,1-Dichloroethane	5.915	63	60198	10.265	ug/l	97
25) 2-Butanone	6.896	43	48022	52.859	ug/l	98
26) 2,2-Dichloropropane	6.890	77	54264	10.403	ug/l	99
27) cis-1,2-Dichloroethene	6.890	96	36403	10.081	ug/l	98
28) Bromochloromethane	7.244	49	25664	10.259	ug/l	98
29) Tetrahydrofuran	7.274	42	30684	52.498	ug/l	98
30) Chloroform	7.421	83	57172	10.163	ug/l	100
31) Cyclohexane	7.701	56	63155	10.778	ug/l	95
32) 1,1,1-Trichloroethane	7.622	97	52856	10.242	ug/l	99
36) 1,1-Dichloropropene	7.835	75	44817	10.219	ug/l	97
37) Ethyl Acetate	6.994	43	20624	9.774	ug/l	97
38) Carbon Tetrachloride	7.817	117	46772	10.303	ug/l	97
39) Methylcyclohexane	9.109	83	60489	10.266	ug/l	98
40) Benzene	8.079	78	130423	10.032	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
 Data File : VY022254.D
 Acq On : 15 May 2025 10:16
 Operator : SY/MD
 Sample : VSTDICC010
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDICC010

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025

Quant Time: May 16 01:21:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
 Quant Title : SW846 8260
 QLast Update : Fri May 16 01:19:39 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.250	41	16228m	10.510	ug/l	
42) 1,2-Dichloroethane	8.158	62	35565	10.115	ug/l	100
43) Isopropyl Acetate	8.201	43	47020	10.217	ug/l	98
44) Trichloroethene	8.865	130	32655	10.101	ug/l	99
45) 1,2-Dichloropropane	9.140	63	32172	10.264	ug/l	93
46) Dibromomethane	9.231	93	17134	10.135	ug/l	99
47) Bromodichloromethane	9.426	83	44045	10.042	ug/l	96
48) Methyl methacrylate	9.225	41	21202	9.851	ug/l	98
49) 1,4-Dioxane	9.231	88	3930	205.769	ug/l #	75
51) 4-Methyl-2-Pentanone	9.999	43	110396	48.889	ug/l	100
52) Toluene	10.170	92	82453	10.015	ug/l	96
53) t-1,3-Dichloropropene	10.396	75	42158	9.849	ug/l	96
54) cis-1,3-Dichloropropene	9.859	75	48923	9.976	ug/l	96
55) 1,1,2-Trichloroethane	10.572	97	21720	10.033	ug/l	97
56) Ethyl methacrylate	10.438	69	32286	9.574	ug/l	98
57) 1,3-Dichloropropane	10.719	76	38908	10.115	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.713	63	66700	44.066	ug/l	100
59) 2-Hexanone	10.761	43	74432	49.162	ug/l	99
60) Dibromochloromethane	10.914	129	28002	9.965	ug/l	99
61) 1,2-Dibromoethane	11.018	107	20378	10.099	ug/l	99
64) Tetrachloroethene	10.646	164	36274	10.404	ug/l	99
65) Chlorobenzene	11.444	112	83730	9.839	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.517	131	28441	9.721	ug/l	97
67) Ethyl Benzene	11.517	91	154155	9.672	ug/l	97
68) m/p-Xylenes	11.627	106	116113	19.334	ug/l	99
69) o-Xylene	11.956	106	55896	9.837	ug/l	96
70) Styrene	11.969	104	91569	9.553	ug/l	99
71) Bromoform	12.133	173	16151	10.228	ug/l #	95
73) Isopropylbenzene	12.255	105	144186	10.120	ug/l	99
74) N-amyl acetate	12.072	43	40345	9.901	ug/l	99
75) 1,1,1,2,2-Tetrachloroethane	12.505	83	22243	10.152	ug/l	96
76) 1,2,3-Trichloropropane	12.560	75	16056m	5.297	ug/l	
77) Bromobenzene	12.536	156	30445	10.005	ug/l	99
78) n-propylbenzene	12.597	91	173575	10.101	ug/l	100
79) 2-Chlorotoluene	12.682	91	95815	10.146	ug/l	100
80) 1,3,5-Trimethylbenzene	12.737	105	116461	10.126	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.304	75	8934	10.388	ug/l	94
82) 4-Chlorotoluene	12.779	91	97309	10.018	ug/l	99
83) tert-Butylbenzene	12.999	119	102274	10.005	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	115175	10.088	ug/l	99
85) sec-Butylbenzene	13.176	105	153443	10.112	ug/l	100
86) p-Isopropyltoluene	13.291	119	124147	9.728	ug/l	100
87) 1,3-Dichlorobenzene	13.285	146	61011	9.822	ug/l	98
88) 1,4-Dichlorobenzene	13.365	146	60723	10.163	ug/l	97
89) n-Butylbenzene	13.621	91	120371	9.970	ug/l	99
90) Hexachloroethane	13.883	117	26089	10.127	ug/l	97
91) 1,2-Dichlorobenzene	13.663	146	52282	10.020	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	3588	10.155	ug/l	99
93) 1,2,4-Trichlorobenzene	14.925	180	29073	9.646	ug/l	99
94) Hexachlorobutadiene	15.023	225	16686	9.927	ug/l	98
95) Naphthalene	15.145	128	54307	9.495	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	24790	9.696	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051525\
Data File : VY022254.D
Acq On : 15 May 2025 10:16
Operator : SY/MD
Sample : VSTDICC010
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025

Quant Time: May 16 01:21:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:19:39 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Manual Integrations


Quant Time: May 16 01:21:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y051525S.M
Quant Title : SW846 8260
QLast Update : Fri May 16 01:19:39 2025
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010



Tgt Ion:168 Resp: 265034
Ion Ratio Lower Upper
168 100
99 56.1 44.2 66.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#2
Dichlorodifluoromethane
Concen: 11.245 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16


Tgt Ion: 85 Resp: 31097
Ion Ratio Lower Upper
85 100
87 32.2 16.5 49.5





#3
Chloromethane
Concen: 11.275 ug/l
RT: 2.068 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010




Tgt Ion: 50 Resp: 65589
Ion Ratio Lower Upper
50 100
52 34.1 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#4
Vinyl Chloride
Concen: 10.739 ug/l
RT: 2.202 min Scan# 79
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Tgt Ion: 62 Resp: 69800
Ion Ratio Lower Upper
62 100
64 33.2 24.5 36.7





#5
Bromomethane
Concen: 10.352 ug/l
RT: 2.592 min Scan# 143
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion: 94 Resp: 5477
Ion Ratio Lower Upper
94 100
96 90.9 73.4 110.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#6
Chloroethane
Concen: 10.665 ug/l
RT: 2.732 min Scan# 166
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Tgt Ion: 64 Resp: 43122
Ion Ratio Lower Upper
64 100
66 31.6 24.8 37.2





#7
Trichlorofluoromethane
Concen: 10.410 ug/l
RT: 3.056 min Scan# 218
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010




Tgt Ion:101 Resp: 6850
Ion Ratio Lower Upper
101 100
103 63.8 51.7 77.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#8
Diethyl Ether
Concen: 10.730 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16



Tgt Ion: 74 Resp: 16945
Ion Ratio Lower Upper
74 100
45 93.2 48.6 145.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 10.708 ug/l
RT: 3.812 min Scan# 344
Delta R.T. -0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Tgt Ion:101 Resp: 3059
Ion Ratio Lower Upper
101 100
85 47.5 38.6 58.0
151 75.6 59.7 89.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#10
Methyl Iodide
Concen: 9.402 ug/l
RT: 4.007 min Scan# 375
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Tgt Ion:142 Resp: 31088
Ion Ratio Lower Upper
142 100
127 46.2 35.7 53.5
141 15.1 11.1 16.7









#11
Tert butyl alcohol
Concen: 54.446 ug/l
RT: 4.878 min Scan# 517
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Tgt Ion: 59 Resp: 1204
Ion Ratio Lower Upper
59 100
57 5.7 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#12
1,1-Dichloroethene
Concen: 10.170 ug/l
RT: 3.793 min Scan# 340
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Tgt Ion: 96 Resp: 28831
Ion Ratio Lower Upper
96 100
61 176.2 134.9 202.3
98 65.1 49.9 74.9





#13

Acrolein

Concen: 55.439 ug/l

RT: 3.659 min Scan# 317

Delta R.T. 0.006 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion: 56 Resp: 1792

Ion Ratio Lower Upper

56 100





55 73.0 55.6 83.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025

Supervised By :Semsettin Yesilyurt 05/16/2025



#14

Allyl chloride

Concen: 10.315 ug/l

RT: 4.378 min Scan# 436

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16




Tgt Ion: 41 Resp: 51275


Ion Ratio Lower Upper

41 100

39 79.2 63.4 95.0


76 37.8 29.5 44.3





#15
Acrylonitrile
Concen: 51.139 ug/l
RT: 5.061 min Scan# 548
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010




Tgt Ion: 53 Resp: 3327
Ion Ratio Lower Upper
53 100
52 82.7 66.1 99.1
51 38.2 30.0 45.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#16
Acetone
Concen: 55.657 ug/l
RT: 3.879 min Scan# 354
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Tgt Ion: 43 Resp: 30229
Ion Ratio Lower Upper
43 100
58 29.5 23.6 35.4





#17
Carbon Disulfide
Concen: 10.448 ug/l
RT: 4.104 min Scan# 391
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010




Tgt Ion: 76 Resp: 9659
Ion Ratio Lower Upper
76 100
78 9.9 7.2 10.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#18
Methyl Acetate
Concen: 9.886 ug/l
RT: 4.385 min Scan# 437
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16



Tgt Ion: 43 Resp: 16338
Ion Ratio Lower Upper
43 100
74 25.6 18.2 27.4





#19

Methyl tert-butyl Ether

Concen: 10.326 ug/l

RT: 5.116 min Scan# 51

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion: 73 Resp: 80714

Ion Ratio Lower Upper

73 100





57 21.3 18.6 28.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025

Supervised By :Semsettin Yesilyurt 05/16/2025



#20

Methylene Chloride

Concen: 12.250 ug/l

RT: 4.622 min Scan# 476

Delta R.T. 0.006 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Tgt Ion: 84 Resp: 42991




Ion Ratio Lower Upper


84 100

49 129.4 109.6 164.4

51 42.0 32.4 48.6


86 61.5 52.3 78.5





#21
trans-1,2-Dichloroethene
Concen: 10.309 ug/l
RT: 5.116 min Scan# 511
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion: 96 Resp: 32258
Ion Ratio Lower Upper
96 100
61 139.6 113.4 170.0
98 62.0 50.9 76.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#22
Diisopropyl ether
Concen: 10.145 ug/l
RT: 6.018 min Scan# 705
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Tgt Ion: 45 Resp: 106493
Ion Ratio Lower Upper
45 100
43 67.7 48.2 72.2
87 26.8 21.4 32.0
59 10.9 9.0 13.4





#23
 Vinyl Acetate
 Concen: 51.192 ug/l
 RT: 5.964 min Scan# 695
 Delta R.T. 0.006 min
 Lab File: VY022254.D
 Acq: 15 May 2025 10:16




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC010




Tgt Ion: 43 Resp: 32465
 Ion Ratio Lower Upper
 43 100
 86 9.7 7.8 11.8

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025




#24
 1,1-Dichloroethane
 Concen: 10.265 ug/l
 RT: 5.915 min Scan# 688
 Delta R.T. -0.000 min
 Lab File: VY022254.D
 Acq: 15 May 2025 10:16



Tgt Ion: 63 Resp: 60198
 Ion Ratio Lower Upper
 63 100
 98 5.6 3.5 10.5
 100 4.6 2.4 7.2









#25
2-Butanone
Concen: 52.859 ug/l
RT: 6.896 min Scan# 849
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010

Tgt Ion: 43 Resp: 4802
Ion Ratio Lower Upper
43 100
72 26.7 22.1 33.1




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#26
2,2-Dichloropropane
Concen: 10.403 ug/l
RT: 6.890 min Scan# 848
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Tgt Ion: 77 Resp: 54264
Ion Ratio Lower Upper
77 100
97 22.9 11.2 33.5





#27
 cis-1,2-Dichloroethene
 Concen: 10.081 ug/l
 RT: 6.890 min Scan# 848
 Delta R.T. -0.000 min
 Lab File: VY022254.D
 Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y





ClientSampleId :

VSTDICC010

Tgt Ion: 96 Resp: 3640
 Ion Ratio Lower Upper
 96 100
 61 152.8 0.0 302.6
 98 67.2 0.0 128.8




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025



#28
 Bromochloromethane
 Concen: 10.259 ug/l
 RT: 7.244 min Scan# 906
 Delta R.T. -0.000 min
 Lab File: VY022254.D
 Acq: 15 May 2025 10:16

Tgt Ion: 49 Resp: 25664
 Ion Ratio Lower Upper
 49 100
 129 2.6 0.0 3.6
 130 67.2 54.7 82.1





#29

Tetrahydrofuran

Concen: 52.498 ug/l

RT: 7.274 min Scan# 911

Delta R.T. 0.012 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC010

Tgt Ion:	42	Resp:	30684
Ion	Ratio	Lower	Upper
42	100		
72	39.9	33.8	50.6
71	39.0	31.5	47.3

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#30

Chloroform

Concen: 10.163 ug/l




RT: 7.421 min Scan# 935


Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16


Tgt Ion:	83	Resp:	57172
Ion	Ratio	Lower	Upper
83	100		
85	65.2	51.8	77.8





#31
Cyclohexane
Concen: 10.778 ug/l
RT: 7.701 min Scan# 981
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion: 56 Resp: 6315
Ion Ratio Lower Upper
56 100
69 39.9 27.4 41.2
84 84.7 70.0 105.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#32
1,1,1-Trichloroethane
Concen: 10.242 ug/l
RT: 7.622 min Scan# 968
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16



Tgt Ion: 97 Resp: 52856
Ion Ratio Lower Upper
97 100
99 64.5 51.4 77.0
61 46.0 38.1 57.1









#33
1,2-Dichloroethane-d4
Concen: 10.222 ug/l
RT: 8.067 min Scan# 1040
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010

Tgt Ion: 65 Resp: 29608
Ion Ratio Lower Upper
65 100
67 52.9 0.0 104.6




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.616 min Scan# 1131
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Tgt Ion:114 Resp: 446109
Ion Ratio Lower Upper
114 100
63 20.7 0.0 41.0
88 15.4 0.0 29.4





#35

Dibromofluoromethane

Concen: 9.856 ug/l

RT: 7.634 min Scan# 91

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010





Tgt Ion: 113 Resp: 2624

Ion	Ratio	Lower	Upper
113	100		
111	103.6	82.6	123.8
192	19.3	15.2	22.8

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#36

1,1-Dichloropropene

Concen: 10.219 ug/l

RT: 7.835 min Scan# 1003




Delta R.T. -0.000 min


Lab File: VY022254.D

Acq: 15 May 2025 10:16

Tgt Ion: 75 Resp: 44817


Ion	Ratio	Lower	Upper
75	100		
110	35.1	16.6	49.8
77	29.8	24.6	37.0





#37
Ethyl Acetate
Concen: 9.774 ug/l
RT: 6.994 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion: 43 Resp: 20624
Ion Ratio Lower Upper
43 100
61 14.2 10.2 15.2
70 10.6 7.6 11.4

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#38
Carbon Tetrachloride
Concen: 10.303 ug/l
RT: 7.817 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Tgt Ion:117 Resp: 46772
Ion Ratio Lower Upper
117 100
119 91.4 75.3 112.9
121 28.4 24.7 37.1





#39
Methylcyclohexane
Concen: 10.266 ug/l
RT: 9.109 min Scan# 1212
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion: 83 Resp: 60489
Ion Ratio Lower Upper
83 100
55 77.9 60.6 90.8
98 51.1 40.0 60.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#40
Benzene
Concen: 10.032 ug/l
RT: 8.079 min Scan# 1043
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Tgt Ion: 78 Resp: 130423
Ion Ratio Lower Upper
78 100
77 23.7 18.9 28.3





#41
Methacrylonitrile
Concen: 10.510 ug/l m
RT: 7.250 min Scan# 902
Delta R.T. 0.030 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010



Tgt Ion: 41 Resp: 16228
Ion Ratio Lower Upper
41 100
39 38.5 39.1 58.7
67 54.4 47.1 70.7
52 27.4 21.4 32.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#42
1,2-Dichloroethane
Concen: 10.115 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16


Tgt Ion: 62 Resp: 35565
Ion Ratio Lower Upper
62 100
98 9.2 0.0 18.4





#43
Isopropyl Acetate
Concen: 10.217 ug/l
RT: 8.201 min Scan# 1063
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010



Tgt Ion: 43 Resp: 47020
Ion Ratio Lower Upper
43 100
61 19.6 15.1 22.7
87 13.1 9.7 14.5




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#44
Trichloroethene
Concen: 10.101 ug/l
RT: 8.865 min Scan# 1172
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Tgt Ion:130 Resp: 32655
Ion Ratio Lower Upper
130 100
95 101.4 0.0 204.8





#45

1,2-Dichloropropane

Concen: 10.264 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion: 63 Resp: 3217

Ion Ratio Lower Upper





63 100

65 28.0 25.7 38.5

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#46

Dibromomethane

Concen: 10.135 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16




Tgt Ion: 93 Resp: 17134


Ion Ratio Lower Upper

93 100

95 82.1 65.2 97.8

174 93.7 74.3 111.5









#47
Bromodichloromethane
Concen: 10.042 ug/l
RT: 9.426 min Scan# 1263
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010

Tgt Ion: 83 Resp: 4404
Ion Ratio Lower Upper
83 100
85 60.3 51.1 76.7
127 8.7 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#48
Methyl methacrylate
Concen: 9.851 ug/l
RT: 9.225 min Scan# 1231
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Tgt Ion: 41 Resp: 21202
Ion Ratio Lower Upper
41 100
69 80.5 63.5 95.3
39 62.0 50.8 76.2





#49

1,4-Dioxane

Concen: 205.769 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC010

Tgt Ion: 88 Resp: 3930
 Ion Ratio Lower Upper
 88 100
 43 0.0 28.5 42.7
 58 84.5 61.4 92.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025



#50

Toluene-d8

Concen: 9.816 ug/l




RT: 10.109 min Scan# 1376


Delta R.T. 0.006 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16


Tgt Ion: 98 Resp: 107048
 Ion Ratio Lower Upper
 98 100
 100 63.7 51.8 77.8





#51
4-Methyl-2-Pentanone
Concen: 48.889 ug/l
RT: 9.999 min Scan# 1358
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010



Tgt Ion: 43 Resp: 110390
Ion Ratio Lower Upper
43 100
58 36.7 29.4 44.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#52
Toluene
Concen: 10.015 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16


Tgt Ion: 92 Resp: 82453
Ion Ratio Lower Upper
92 100
91 168.8 139.1 208.7





#53
t-1,3-Dichloropropene
Concen: 9.849 ug/l
RT: 10.396 min Scan# 1422
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion: 75 Resp: 42158
Ion Ratio Lower Upper
75 100
77 33.5 25.0 37.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#54
cis-1,3-Dichloropropene
Concen: 9.976 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Tgt Ion: 75 Resp: 48923
Ion Ratio Lower Upper
75 100
77 34.1 25.1 37.7
39 58.7 44.5 66.7





#55
1,1,2-Trichloroethane
Concen: 10.033 ug/l
RT: 10.572 min Scan# 1452
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010




Tgt Ion: 97 Resp: 21720
Ion Ratio Lower Upper
97 100
83 94.5 70.3 105.5
85 57.6 46.2 69.4
99 63.1 50.3 75.5

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#56
Ethyl methacrylate
Concen: 9.574 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Tgt Ion: 69 Resp: 32286
Ion Ratio Lower Upper
69 100
41 72.5 56.5 84.7
39 45.6 37.1 55.7





#57
1,3-Dichloropropane
Concen: 10.115 ug/l
RT: 10.719 min Scan# 1476
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion: 76 Resp: 38908
Ion Ratio Lower Upper
76 100
78 31.3 25.8 38.8

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#58
2-Chloroethyl Vinyl ether
Concen: 44.066 ug/l
RT: 9.713 min Scan# 1311
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Tgt Ion: 63 Resp: 66700
Ion Ratio Lower Upper
63 100
106 26.3 21.0 31.4





#59
2-Hexanone
Concen: 49.162 ug/l
RT: 10.761 min Scan# 1483
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion: 43 Resp: 7443
Ion Ratio Lower Upper
43 100
58 53.2 26.1 78.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#60
Dibromochloromethane
Concen: 9.965 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Tgt Ion:129 Resp: 28002
Ion Ratio Lower Upper
129 100
127 77.8 38.6 115.7





#61
1,2-Dibromoethane
Concen: 10.099 ug/l
RT: 11.018 min Scan# 1524
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010



Tgt Ion:107 Resp: 20378
Ion Ratio Lower Upper
107 100
109 92.6 74.6 111.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#62
4-Bromofluorobenzene
Concen: 9.959 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16


Tgt Ion: 95 Resp: 34423
Ion Ratio Lower Upper
95 100
174 81.3 0.0 166.8
176 79.7 0.0 160.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010




Tgt Ion:117 Resp: 37721
Ion Ratio Lower Upper
117 100
82 57.0 46.6 70.0
119 32.7 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#64
Tetrachloroethene
Concen: 10.404 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16



Tgt Ion:164 Resp: 36274
Ion Ratio Lower Upper
164 100
166 125.3 100.4 150.6
129 96.3 76.4 114.6
131 93.9 73.1 109.7





#65

Chlorobenzene

Concen: 9.839 ug/l

RT: 11.444 min Scan# 11

Delta R.T. 0.006 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010

Tgt Ion:112 Resp: 83730

Ion Ratio Lower Upper

112 100





114 30.3 25.8 38.6

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025

Supervised By :Semsettin Yesilyurt 05/16/2025



#66

1,1,1,2-Tetrachloroethane

Concen: 9.721 ug/l

RT: 11.517 min Scan# 1607

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16




Tgt Ion:131 Resp: 28441


Ion Ratio Lower Upper

131 100

133 92.3 47.3 141.9


119 69.0 33.1 99.2





#67
Ethyl Benzene
Concen: 9.672 ug/l
RT: 11.517 min Scan# 1607
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion: 91 Resp: 154158
Ion Ratio Lower Upper
91 100
106 29.8 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#68
m/p-Xylenes
Concen: 19.334 ug/l
RT: 11.627 min Scan# 1625
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16



Tgt Ion:106 Resp: 116113
Ion Ratio Lower Upper
106 100
91 202.4 163.0 244.4









#69
o-Xylene
Concen: 9.837 ug/l
RT: 11.956 min Scan# 1679
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010

Tgt Ion:106 Resp: 55890
Ion Ratio Lower Upper
106 100
91 210.6 108.3 324.8




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#70
Styrene
Concen: 9.553 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Tgt Ion:104 Resp: 91569
Ion Ratio Lower Upper
104 100
78 50.4 40.9 61.3
103 54.7 43.7 65.5





#71

Bromoform

Concen: 10.228 ug/l

RT: 12.133 min Scan# 1708

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :

VSTDICC010





Tgt Ion:173 Resp: 1615

Ion	Ratio	Lower	Upper
173	100		
175	45.9	24.7	74.1
254	0.0	0.0	0.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907




Delta R.T. -0.000 min


Lab File: VY022254.D

Acq: 15 May 2025 10:16

Tgt Ion:152 Resp: 171972


Ion	Ratio	Lower	Upper
152	100		
115	58.4	29.4	88.2
150	161.2	0.0	353.8





#73
Isopropylbenzene
Concen: 10.120 ug/l
RT: 12.255 min Scan# 1728
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion:105 Resp: 144180
Ion Ratio Lower Upper
105 100
120 25.9 13.2 39.6

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#74
N-ethyl acetate
Concen: 9.901 ug/l
RT: 12.072 min Scan# 1698
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16



Tgt Ion: 43 Resp: 40345
Ion Ratio Lower Upper
43 100
70 38.2 31.4 47.0
55 20.5 17.0 25.6
61 23.1 18.6 28.0





#75

1,1,2,2-Tetrachloroethane

Concen: 10.152 ug/l

RT: 12.505 min Scan# 1769

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC010

Tgt Ion: 83	Resp: 2224
Ion Ratio	Lower Upper
83 100	
131 11.1	4.9 14.7
85 60.8	32.0 96.0

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#76

1,2,3-Trichloropropane

Concen: 5.297 ug/l m




RT: 12.560 min Scan# 1778


Delta R.T. 0.006 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16


Tgt Ion: 75	Resp: 16056
Ion Ratio	Lower Upper
75 100	
77 0.0	0.0 0.0





#77
Bromobenzene
Concen: 10.005 ug/l
RT: 12.536 min Scan# 1774
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion:156 Resp: 3044
Ion Ratio Lower Upper
156 100
77 189.7 94.5 283.6
158 98.1 48.4 145.3

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#78
n-propylbenzene
Concen: 10.101 ug/l
RT: 12.597 min Scan# 1784
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16



Tgt Ion: 91 Resp: 173575
Ion Ratio Lower Upper
91 100
120 22.1 11.1 33.3









#79
2-Chlorotoluene
Concen: 10.146 ug/l
RT: 12.682 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16

Instrument :
MSVOA_Y
ClientSampleId :
VSTDICC010

Tgt Ion: 91 Resp: 9581
Ion Ratio Lower Upper
91 100
126 33.6 16.8 50.4




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#80
1,3,5-Trimethylbenzene
Concen: 10.126 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16


Tgt Ion:105 Resp: 116461
Ion Ratio Lower Upper
105 100
120 47.4 24.2 72.6





#81
trans-1,4-Dichloro-2-butene
Concen: 10.388 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010



Tgt Ion: 75 Resp: 8934
Ion Ratio Lower Upper
75 100
53 87.6 75.8 113.8
89 44.0 36.2 54.2




Manual Integrations
APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#82
4-Chlorotoluene
Concen: 10.018 ug/l
RT: 12.779 min Scan# 1814
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16


Tgt Ion: 91 Resp: 97309
Ion Ratio Lower Upper
91 100
126 33.9 16.8 50.4





#83
 tert-Butylbenzene
 Concen: 10.005 ug/l
 RT: 12.999 min Scan# 1850
 Delta R.T. -0.000 min
 Lab File: VY022254.D
 Acq: 15 May 2025 10:16




Instrument : MSVOA_Y
 ClientSampleId : VSTDIC010



Tgt Ion:	119	Resp:	102274
Ion Ratio	Lower	Upper	
119	100		
91	66.9	32.8	98.3
134	25.7	12.9	38.6




Manual Integrations
 APPROVED


Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025



#84
 1,2,4-Trimethylbenzene
 Concen: 10.088 ug/l
 RT: 13.042 min Scan# 1857
 Delta R.T. -0.000 min
 Lab File: VY022254.D
 Acq: 15 May 2025 10:16


Tgt Ion:	105	Resp:	115175
Ion Ratio	Lower	Upper	
105	100		
120	44.0	22.3	66.9





#85
 sec-Butylbenzene
 Concen: 10.112 ug/l
 RT: 13.176 min Scan# 1879
 Delta R.T. -0.000 min
 Lab File: VY022254.D
 Acq: 15 May 2025 10:16




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDIC010




Tgt Ion:105 Resp: 15344
 Ion Ratio Lower Upper
 105 100
 134 19.9 9.8 29.5

Manual Integrations
 APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
 Supervised By :Semsettin Yesilyurt 05/16/2025




#86
 p-Isopropyltoluene
 Concen: 9.728 ug/l
 RT: 13.291 min Scan# 1898
 Delta R.T. -0.000 min
 Lab File: VY022254.D
 Acq: 15 May 2025 10:16




Tgt Ion:119 Resp: 124147
 Ion Ratio Lower Upper
 119 100
 134 25.4 12.9 38.6
 91 24.1 12.1 36.3





#87
1,3-Dichlorobenzene
Concen: 9.822 ug/l
RT: 13.285 min Scan# 1897
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument : MSVOA_Y
ClientSampleId : VSTDIC010




Tgt Ion:146 Resp: 6101
Ion Ratio Lower Upper
146 100
111 39.3 19.9 59.7
148 62.1 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#88
1,4-Dichlorobenzene
Concen: 10.163 ug/l
RT: 13.365 min Scan# 1910
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16



Tgt Ion:146 Resp: 60723
Ion Ratio Lower Upper
146 100
111 42.0 20.0 59.9
148 62.5 32.1 96.5





#89

n-Butylbenzene

Concen: 9.970 ug/l

RT: 13.621 min Scan# 1951

Delta R.T. 0.006 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :

VSTDIC010





Tgt Ion: 91 Resp: 12037

Ion	Ratio	Lower	Upper
91	100		
92	53.6	26.3	78.9
134	25.5	12.4	37.4

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#90

Hexachloroethane

Concen: 10.127 ug/l

RT: 13.883 min Scan# 1995




Delta R.T. 0.006 min


Lab File: VY022254.D

Acq: 15 May 2025 10:16

Tgt Ion:117 Resp: 26089


Ion	Ratio	Lower	Upper
117	100		
201	68.2	35.2	105.6





#91
1,2-Dichlorobenzene
Concen: 10.020 ug/l
RT: 13.663 min Scan# 1958
Delta R.T. 0.006 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16




Instrument :
MSVOA_Y
ClientSampleId :
VSTDIC010




Tgt Ion:146 Resp: 5228
Ion Ratio Lower Upper
146 100
111 39.9 20.5 61.5
148 63.5 32.0 96.0

Manual Integrations
APPROVED



Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025




#92
1,2-Dibromo-3-Chloropropane
Concen: 10.155 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY022254.D
Acq: 15 May 2025 10:16



Tgt Ion: 75 Resp: 3588
Ion Ratio Lower Upper
75 100
155 88.0 43.4 130.1
157 109.8 54.6 163.9





#93

1,2,4-Trichlorobenzene

Concen: 9.646 ug/l

RT: 14.925 min Scan# 2166

Delta R.T. 0.006 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :





VSTDIC010

Tgt Ion:180	Resp:	2907
Ion Ratio	Lower	Upper
180	100	
182	95.4	48.0 144.0
145	31.4	15.0 45.1

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#94

Hexachlorobutadiene

Concen: 9.927 ug/l




RT: 15.023 min Scan# 2182


Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Tgt Ion:225	Resp:	16686
Ion Ratio	Lower	Upper
225	100	
223	62.2	31.5 94.5
227	65.8	32.1 96.3





#95

Naphthalene

Concen: 9.495 ug/l

RT: 15.145 min Scan# 21

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Instrument :

MSVOA_Y

ClientSampleId :





VSTDICC010

Tgt Ion:128	Resp:	5430
Ion Ratio	Lower	Upper
128	100	
127	12.6	10.2 15.2
129	11.5	8.8 13.2

Manual Integrations

APPROVED

Reviewed By :Mahesh Dadoda 05/16/2025
Supervised By :Semsettin Yesilyurt 05/16/2025



#96

1,2,3-Trichlorobenzene

Concen: 9.696 ug/l

RT: 15.328 min Scan# 2232

Delta R.T. -0.000 min

Lab File: VY022254.D

Acq: 15 May 2025 10:16

Tgt Ion:180	Resp:	24790
Ion Ratio	Lower	Upper
180	100	
182	95.5	47.1 141.3
145	33.4	15.9 47.6

