

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: SW8260-Low

Datafile : VX046287.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0521WBS01	Benzene	20	20.3	ug/L	102			82	109	
	Toluene	20	20.5	ug/L	103			82	110	
	Ethyl Benzene	20	20.1	ug/L	101			83	109	
	m/p-Xylenes	40	40.3	ug/L	101			82	110	
	o-Xylene	20	20.9	ug/L	104			83	109	
	Isopropylbenzene	20	20.4	ug/L	102			83	112	
	N-propylbenzene	20	20.4	ug/L	102			83	112	
	1,3,5-Trimethylbenzene	20	21.1	ug/L	106			85	112	
	tert-Butylbenzene	20	20.2	ug/L	101			83	112	
	1,2,4-Trimethylbenzene	20	20.7	ug/L	104			85	111	
	Sec-butylbenzene	20	20.3	ug/L	102			81	114	
	p-Isopropyltoluene	20	20.5	ug/L	103			78	116	
	n-Butylbenzene	20	19.7	ug/L	99			75	115	