

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q2075

Client: CDM Smith

Analytical Method: SW8260-Low

Datafile : VX046288.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0521WBSD01	Benzene	20	20.4	ug/L	102	0		82	109	20
	Toluene	20	21.1	ug/L	106	3		82	110	20
	Ethyl Benzene	20	20.4	ug/L	102	1		83	109	20
	m/p-Xylenes	40	41.1	ug/L	103	2		82	110	20
	o-Xylene	20	21.0	ug/L	105	1		83	109	20
	Isopropylbenzene	20	20.8	ug/L	104	2		83	112	20
	N-propylbenzene	20	20.7	ug/L	104	2		83	112	20
	1,3,5-Trimethylbenzene	20	21.0	ug/L	105	1		85	112	20
	tert-Butylbenzene	20	20.5	ug/L	103	2		83	112	20
	1,2,4-Trimethylbenzene	20	20.9	ug/L	104	0		85	111	20
	Sec-butylbenzene	20	21.1	ug/L	106	4		81	114	20
	p-Isopropyltoluene	20	20.7	ug/L	104	1		78	116	20
	n-Butylbenzene	20	20.2	ug/L	101	2		75	115	20