

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075
 Lab File ID: VX046284.D Date Analyzed: 05/21/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:55
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	95080	5.54	161730	6.75	140445	10.05
UPPER LIMIT	190160	6.037	323460	7.251	280890	10.549
LOWER LIMIT	47540	5.037	80865	6.251	70222.5	9.549
EPA SAMPLE NO.						
FB-05152025	63566	5.55	126684	6.76	120601	10.05
FB-05162025	67975	5.55	135035	6.76	128288	10.05
VX0521WBL01	66690	5.54	132814	6.76	124685	10.05
VX0521WBS01	83102	5.54	146697	6.76	130479	10.05
VX0521WBSD01	82656	5.54	142441	6.76	127622	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.