

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075
Lab File ID: VX046331.D Date Analyzed: 05/23/2025
Instrument ID: MSVOA_X Time Analyzed: 09:24
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	89943	5.54	154938	6.76	137123	10.05
UPPER LIMIT	179886	6.044	309876	7.257	274246	10.549
LOWER LIMIT	44971.5	5.044	77469	6.257	68561.5	9.549
EPA SAMPLE NO.						
SS-MW1-11.5ME	61392	5.54	119098	6.76	110366	10.05
VX0523MBL01	59074	5.54	117281	6.76	111918	10.05
VX0523MBS01	83340	5.54	146324	6.76	129426	10.05

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.