

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02
 Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075
 Lab File ID: VY022329.D Date Analyzed: 05/20/2025
 Instrument ID: MSVOA_Y Time Analyzed: 08:40
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	180374	7.71	298369	8.62	252391	11.41
UPPER LIMIT	360748	8.207	596738	9.116	504782	11.914
LOWER LIMIT	90187	7.207	149185	8.116	126196	10.914
EPA SAMPLE NO.						
SS-10	237248	7.71	425729	8.62	352723	11.41
SS-910	241551	7.71	430984	8.62	344445	11.41
SS-11	239791	7.71	432569	8.62	345893	11.41
SS-11MS	134739	7.71	248613	8.62	210937	11.41
SS-11MSD	117744	7.71	220285	8.62	197296	11.41
VY0520SBL01	253729	7.71	453295	8.62	355808	11.42
VY0520SBS01	162404	7.71	280923	8.62	236555	11.42
VY0520SBSD01	169930	7.71	290999	8.62	242816	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.