

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CAMP02  
 Lab Code: CHEM Case No.: Q2075 SAS No.: Q2075 SDG NO.: Q2075  
 Lab File ID: VX046331.D Date Analyzed: 05/23/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 09:24  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 66629         | 12.018 |  |  |  |  |
| UPPER LIMIT    | 133258        | 12.518 |  |  |  |  |
| LOWER LIMIT    | 33314.5       | 11.518 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| SS-MW1-11.5ME  | 48680         | 12.02  |  |  |  |  |
| VX0523MBL01    | 48375         | 12.02  |  |  |  |  |
| VX0523MBS01    | 63342         | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.