

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QC Batch ID # VY052025

Review By	Mahesh Dadoda	Review On	5/21/2025 3:43:30 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/23/2025 2:22:55 AM		
SubDirectory	VY052025	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133969			
CCC		VP133970,VP133971			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VY022328.D	20 May 2025 08:09	SY/MD	Ok
2	VSTDCCC050	VY022329.D	20 May 2025 08:40	SY/MD	Ok,M
3	VY0520SBL01	VY022330.D	20 May 2025 09:48	SY/MD	Ok
4	VY0520SBS01	VY022331.D	20 May 2025 10:21	SY/MD	Ok,M
5	VY0520SBSD01	VY022332.D	20 May 2025 10:43	SY/MD	Ok,M
6	Q2062-19	VY022333.D	20 May 2025 11:28	SY/MD	Ok
7	Q2062-23	VY022334.D	20 May 2025 11:51	SY/MD	Ok
8	Q2071-11	VY022335.D	20 May 2025 12:15	SY/MD	Ok
9	Q2056-02	VY022336.D	20 May 2025 12:38	SY/MD	Ok
10	Q2056-03	VY022337.D	20 May 2025 13:01	SY/MD	Ok
11	Q2056-04	VY022338.D	20 May 2025 13:25	SY/MD	Ok
12	Q2056-05	VY022339.D	20 May 2025 13:48	SY/MD	Not Ok
13	Q2080-02	VY022340.D	20 May 2025 14:12	SY/MD	Ok
14	Q2080-08	VY022341.D	20 May 2025 14:35	SY/MD	Ok
15	Q2032-02	VY022342.D	20 May 2025 14:59	SY/MD	Ok
16	Q2075-01	VY022343.D	20 May 2025 15:22	SY/MD	Ok
17	Q2075-02	VY022344.D	20 May 2025 15:45	SY/MD	Ok
18	Q2075-03	VY022345.D	20 May 2025 16:09	SY/MD	Ok
19	Q2084-02	VY022346.D	20 May 2025 16:32	SY/MD	Ok
20	Q2032-01	VY022347.D	20 May 2025 16:56	SY/MD	Dilution
21	Q2032-05MS	VY022348.D	20 May 2025 17:19	SY/MD	Ok,M

Instrument ID: **MSVOA_Y**

Daily Analysis Runlog For Sequence/QCBatch ID # VY052025

Review By	Mahesh Dadoda	Review On	5/21/2025 3:43:30 PM		
Supervise By	Semsettin Yesilyurt	Supervise On	5/23/2025 2:22:55 AM		
SubDirectory	VY052025	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP133969			
CCC		VP133970,VP133971			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	Q2032-06MSD	VY022349.D	20 May 2025 17:41	SY/MD	Ok,M
23	Q2075-04MS	VY022350.D	20 May 2025 18:04	SY/MD	Ok,M
24	Q2075-05MSD	VY022351.D	20 May 2025 18:26	SY/MD	Ok,M
25	VSTDCCC050	VY022352.D	20 May 2025 19:34	SY/MD	Ok,M

M : Manual Integration