

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052225

Review By	John Carlone	Review On	5/23/2025 11:05:57 AM
Supervise By	Maresh Dadoda	Supervise On	5/23/2025 4:07:06 PM
SubDirectory	VY052225	HP Acquire Method	MSVOA_Y HP Processing Method 82y051525s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP134001
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP134002,VP134003 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY022379.D	22 May 2025 08:05		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY022380.D	22 May 2025 08:35		SY/MD	Ok,M
3	VY0522SBL01	VY0522SBL01	VY022381.D	22 May 2025 09:07		SY/MD	Ok
4	VY0522SBS01	VY0522SBS01	VY022382.D	22 May 2025 09:36		SY/MD	Ok,M
5	VY0522SBSD01	VY0522SBSD01	VY022383.D	22 May 2025 09:59		SY/MD	Ok,M
6	Q2087-03RE	NB-08-05202025RE	VY022384.D	22 May 2025 10:38	vial-B ,Internal Standard Fail	SY/MD	Confirms
7	Q2101-03	TP-1-MHE-VOC	VY022385.D	22 May 2025 11:01	vial-B	SY/MD	Ok
8	Q2097-11RE	RT3419RE	VY022386.D	22 May 2025 11:25	vial-B ,Internal Standard Fail	SY/MD	Confirms
9	Q2102-01	LAW-25-0077	VY022387.D	22 May 2025 11:48	vial-B	SY/MD	Ok
10	Q2104-01RE	TR-06-052125RE	VY022388.D	22 May 2025 12:11	vial-B ,Internal Standard Fail; Surrogate Fail	SY/MD	Confirms
11	Q2074-01	TP-12	VY022389.D	22 May 2025 12:35	vial-A	SY/MD	Ok
12	Q2074-02	TP-7	VY022390.D	22 May 2025 12:58	vial-A	SY/MD	Ok
13	Q2074-03	TP-15	VY022391.D	22 May 2025 13:22	vial-A	SY/MD	Ok
14	Q2074-04	TP-20	VY022392.D	22 May 2025 13:45	vial-A	SY/MD	Ok
15	Q2074-05	TP-38	VY022393.D	22 May 2025 14:09	vial-A	SY/MD	Ok
16	Q2074-06	TP-19	VY022394.D	22 May 2025 14:32	vial-A	SY/MD	Ok
17	Q2074-07	TP-40	VY022395.D	22 May 2025 14:55	vial-A	SY/MD	Ok

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY052225

Review By	John Carlone	Review On	5/23/2025 11:05:57 AM		
Supervise By	Mahesh Dadoda	Supervise On	5/23/2025 4:07:06 PM		
SubDirectory	VY052225	HP Acquire Method	MSVOA_Y	HP Processing Method	82y051525s.m
STD. NAME		STD REF.#			
Tune/Reschk Initial Calibration Stds		VP134001			
CCC		VP134002,VP134003			
Internal Standard/PEM		VP133934			
ICV/I.BLK					
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	Q2074-08	TP-18	VY022396.D	22 May 2025 15:19	vial-A	SY/MD	Ok
19	Q2100-01	TP-16	VY022397.D	22 May 2025 15:42	vial-A ,Internal Standard Fail	SY/MD	ReRun
20	Q2100-02	TP-22	VY022398.D	22 May 2025 16:06	vial-A	SY/MD	Ok
21	Q2100-03	TP-21	VY022399.D	22 May 2025 16:29	vial-A ,Not purge	SY/MD	Not Ok
22	Q2100-04	TP-45	VY022400.D	22 May 2025 16:53	vial-A	SY/MD	Ok
23	Q2112-01	EO-03-05222025	VY022401.D	22 May 2025 17:16	vial-A ,Internal Standard Fail	SY/MD	ReRun
24	Q2109-03	TP-02-MHF-VOC	VY022402.D	22 May 2025 17:40	vial-A	SY/MD	Ok
25	Q2075-06	SS-MW1-11.5	VY022403.D	22 May 2025 18:03	vial-A ,Need MeOH	SY/MD	Dilution
26	VIBLK	VIBLK	VY022404.D	22 May 2025 18:26		SY/MD	Ok
27	VSTDCCC050	VSTDCCC050EC	VY022405.D	22 May 2025 18:49		SY/MD	Ok,M

M : Manual Integration